

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG121217\
 Data File : BG031518.D
 Acq On : 13 Dec 2017 8:03
 Operator : SJ/JU
 Sample : I6763-10
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 120517-TIDO-F-U-S

Integration Parameters: rteint.p
 Integrator: RTE
 Smoothing : ON Filtering: 5
 Sampling : 1 Min Area: 3 % of largest Peak
 Start Thrs: 0.2 Max Peaks: 100
 Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
 Peak separation: 5

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG121117.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	5.184	315	322	332	rBV	47167	77051	1.24%	0.393%
2	5.671	399	405	424	rBV	281671	516496	8.32%	2.634%
3	7.281	671	679	690	rBV2	171880	358528	5.78%	1.828%
4	7.645	733	741	751	rBV	532429	990956	15.97%	5.053%
5	8.110	812	820	830	rBV	171974	304809	4.91%	1.554%
6	8.433	866	875	888	rBV	735629	1335676	21.52%	6.811%
7	9.279	1011	1019	1030	rBV	488024	982480	15.83%	5.010%
8	10.924	1292	1299	1317	rBV	198570	442008	7.12%	2.254%
9	13.350	1705	1712	1734	rBV	1789260	3000860	48.36%	15.302%
10	14.731	1940	1947	1958	rBV2	403913	734419	11.83%	3.745%
11	16.224	2195	2201	2220	rBV	382089	739871	11.92%	3.773%
12	17.475	2408	2414	2426	rBV2	642273	1027534	16.56%	5.240%
13	20.066	2849	2855	2870	rBV	4461380	6205691	100.00%	31.644%
14	21.747	3134	3141	3155	rBV	835258	1487059	23.96%	7.583%
15	25.031	3689	3700	3718	rBV2	461984	1407530	22.68%	7.177%


Sum of corrected areas: 19610968

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG121217\
Data File : BG031518.D
Acq On : 13 Dec 2017 8:03
Operator : SJ/JU
Sample : I6763-10
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
120517-TIDO-F-U-S

Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG121117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P



Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG121217\
Data File : BG031518.D
Acq On : 13 Dec 2017 8:03
Operator : SJ/JU
Sample : I6763-10
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
120517-TIDO-F-U-S










Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG121117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 2-Pentanone, 4-hydroxy-4-me... Concentration Rank 3

R.T.	EstConc	Area	Relative to ISTD	R.T.
5.18	5.06 ng	77051	1,4-Dichlorobenzene-d4	8.11

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	64
2			Acetic acid, 1,1-dimethylethyl e...	116	C6H12O2	000540-88-5	38
3			Morpholine, 4-methyl-	101	C5H11NO	000109-02-4	9
4			Butane, 1-ethoxy-	102	C6H14O	000628-81-9	9
5			2,3-Butanedione, monooxime	101	C4H7NO2	000057-71-6	9



Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA G\DATA\BG121217\
Data File : BG031518.D
Acq On : 13 Dec 2017 8:03
Operator : SJ/JU
Sample : I6763-10
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
120517-TIDO-F-U-S










Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG121117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 2 unknown7.65 Concentration Rank 2

R.T.	EstConc	Area	Relative to ISTD	R.T.
7.65	65.02 ng	990956	1,4-Dichlorobenzene-d4	8.11

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			3-Chloro-6-fluoro-pyrazine	132	C4H2ClFN2	1000146-10-7	35
2			(E)-3-Chloro-2-methyl-2-pentenal	132	C6H9ClO	031357-76-3	35
3			5-Aminoindole	132	C8H8N2	005192-03-0	22
4			Tranylcypromine-propionyl	189	C12H15NO	1000123-86-3	12
5			4-Chloro-6-fluoro-pyrimidine	132	C4H2ClFN2	051422-01-6	9



Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG121217\
Data File : BG031518.D
Acq On : 13 Dec 2017 8:03
Operator : SJ/JU
Sample : I6763-10
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
120517-TIDO-F-U-S

Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG121117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
2-Pentanone, 4-hy...	5.18	5.1	ng	77051	1	8.11	304809	20.0
unknown7.65	7.65	65.0	ng	990956	1	8.11	304809	20.0