




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:16:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(3) 1,4-Dioxane-d8 (S)

3.284min (+ 0.008) 3.21 ng/uL

response 18306

Ion	Exp%	Act%
96.00	100.00	100.00
64.00	82.80	114.47#
34.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:16:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(3) 1,4-Dioxane-d8 (S)

3.284min (+ 0.008) 3.38 ng/uL m

response 19288

Ion	Exp%	Act%
96.00	100.00	100.00
64.00	82.80	114.47#
34.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:16:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(54) 4-Nitrophenol-d4 (S)

14.683min (+ 0.008) 0.11 ng/u1

response 643

Ion	Exp%	Act%
143.00	100.00	100.00
113.00	89.70	126.58#
41.00	48.20	1887.56#
42.00	26.60	586.27#




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:16:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(54) 4-Nitrophenol-d4 (S)

14.712min (+ 0.038) 5.62 ng/ul m

response 31685

Ion	Exp%	Act%
143.00	100.00	100.00
113.00	89.70	93.11
41.00	48.20	237.17#
42.00	26.60	83.65#




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:16:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(73) Anthracene-d10 (S)

17.209min (-0.097) 23.26 ng/u1

response 1088670

Ion	Exp%	Act%
188.00	100.00	100.00
94.00	10.70	12.41
80.00	10.80	12.47
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
 Data File : BG063675.D
 Acq On : 13 Dec 2024 12:52
 Operator : RC/JU
 Sample : P5153-16
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:16:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 23:59:23 2024
 Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
 Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(73) Anthracene-d10 (S)

17.309min (+ 0.002) 19.79 ng/ul m

response 926423

Ion	Exp%	Act%
188.00	100.00	100.00
94.00	10.70	14.81#
80.00	10.80	13.41#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:16:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(87) Chrysene

21.516min (+ 0.014) 2.82 ng/u1

response 153410

Ion	Exp%	Act%
228.00	100.00	100.00
226.00	29.20	37.22#
229.00	18.50	55.99#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
 Data File : BG063675.D
 Acq On : 13 Dec 2024 12:52
 Operator : RC/JU
 Sample : P5153-16
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:16:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 23:59:23 2024
 Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
 Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(87) Chrysene

21.516min (+ 0.014) 3.28 ng/ul m

response 178069

Ion	Exp%	Act%
228.00	100.00	100.00
226.00	29.20	37.22#
229.00	18.50	55.99#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:16:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(88) Perylene-d12 (I)

24.465min (-0.015) 20.00 ng/u1

response 933

Ion	Exp%	Act%
264.00	100.00	100.00
260.00	23.00	92.53#
265.00	21.20	72.13#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:16:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(88) Perylene-d12 (I)

24.530min (+ 0.049) 20.00 ng/ul m

response 510859

Ion	Exp%	Act%
264.00	100.00	100.00
260.00	23.00	32.05#
265.00	21.20	28.92#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:20:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(90) Benzo(b)fluoranthene

23.502min (-0.021) 0.04 ng/u1

response 1109

Ion	Exp%	Act%
252.00	100.00	100.00
253.00	22.10	166.52#
125.00	8.70	715.18#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:20:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(91) Benzo(k)fluoranthene

23.566min (-0.021) 2.50 ng/u1

response 72336

Ion	Exp%	Act%
252.00	100.00	100.00
253.00	22.10	54.45#
125.00	8.10	159.70#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:20:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(91) Benzo(k)fluoranthene

23.625min (+ 0.038) 1.34 ng/ul m

response 38930

Ion	Exp%	Act%
252.00	100.00	100.00
253.00	22.10	59.61#
125.00	8.10	250.23#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:20:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(92) Benzo(a)pyrene-d12 (S)

24.242min (-0.033) 0.03 ng/u1

response 729

Ion	Exp%	Act%
264.00	100.00	100.00
132.00	11.70	218.23#
118.00	6.90	199.04#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:20:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(92) Benzo(a)pyrene-d12 (S)

24.324min (+ 0.049) 19.97 ng/ul m

response 479112

Ion	Exp%	Act%
264.00	100.00	100.00
132.00	11.70	20.42#
118.00	6.90	18.61#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
 Data File : BG063675.D
 Acq On : 13 Dec 2024 12:52
 Operator : RC/JU
 Sample : P5153-16
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:24:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 23:59:23 2024
 Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
 Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(93) Benzo(a)pyrene (C)

24.318min (-0.021) 0.15 ng/u1

response 4197

Ion	Exp%	Act%
252.00	100.00	100.00
253.00	20.70	94.09#
125.00	10.00	412.45#
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:24:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024



TIC: BG063675.D\data.ms

(93) Benzo(a)pyrene (C)

24.389min (+ 0.049) 2.86 ng/ul m

response 77934

Ion	Exp%	Act%
252.00	100.00	100.00
253.00	20.70	66.41#
125.00	10.00	229.58#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
 Data File : BG063675.D
 Acq On : 13 Dec 2024 12:52
 Operator : RC/JU
 Sample : P5153-16
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:29:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 23:59:23 2024
 Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
 Supervised By :mohammad ahmed 12/14/2024

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.820	152		223341	20.000	ng/ul	0.00
20) Naphthalene-d8	10.605	136		878458	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.454	164		584817	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.209	188		1088670	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.475	240		909904	20.000	ng/ul	# 0.01
88) Perylene-d12	24.530	264		510859m	20.000	ng/ul	0.05
System Monitoring Compounds							
3) 1,4-Dioxane-d8	3.284	96		19288m	3.384	ng/uL	0.00
4) Pyridine-d5	3.713	84		265013	14.740	ng/ul	0.03
7) Phenol-d5	7.015	99		404470	19.664	ng/ul	0.02
9) Bis-(2-Chloroethyl)eth...	7.150	67		271925	19.141	ng/ul	0.00
11) 2-Chlorophenol-d4	7.362	132		291929	21.922	ng/ul	0.01
15) 4-Methylphenol-d8	8.543	113		253784	15.891	ng/ul	0.01
21) Nitrobenzene-d5	8.972	128		91623	14.608	ng/ul	0.00
24) 2-Nitrophenol-d4	9.688	143		19291	2.744	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.247	165		296838	21.806	ng/ul	0.01
31) 4-Chloroaniline-d4	10.752	131		38614	2.234	ng/ul	0.00
46) Dimethylphthalate-d6	13.860	166		901048	22.643	ng/ul	0.00
49) Acenaphthylene-d8	14.148	160		672720	14.830	ng/ul	0.00
54) 4-Nitrophenol-d4	14.712	143		31685m	5.620	ng/ul	0.04
60) Fluorene-d10	15.452	176		718191	20.412	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.576	200		769	0.136	ng/ul	0.00
73) Anthracene-d10	17.309	188		926423m	19.793	ng/ul	0.00
81) Pyrene-d10	19.612	212		1023241	21.611	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.324	264		479112m	19.971	ng/ul	0.05
Target Compounds							
						Qvalue	
6) Benzaldehyde	6.962	77		14707	1.166	ng/ul#	89
8) Phenol	7.045	94		29989	1.368	ng/ul	96
30) Naphthalene	10.658	128		85001	1.935	ng/ul	98
36) 2-Methylnaphthalene	12.268	142		129034	4.088	ng/ul	93
37) 1-Methylnaphthalene	12.491	142		39571	1.231	ng/ul	96
72) Phenanthrene	17.250	178		223578	4.262	ng/ul#	87
74) Anthracene	17.344	178		54965	1.035	ng/ul#	46
80) Fluoranthene	19.277	202		179919	3.343	ng/ul#	95
82) Pyrene	19.642	202		175968	3.073	ng/ul#	89
85) Benzo(a)anthracene	21.457	228		93710	1.643	ng/ul#	29
86) Bis(2-ethylhexyl)phtha...	21.357	149		60233	2.356	ng/ul#	89
87) Chrysene	21.516	228		178069m	3.276	ng/ul	
90) Benzo(b)fluoranthene	23.566	252		151013m	5.201	ng/ul	
91) Benzo(k)fluoranthene	23.625	252		38930m	1.344	ng/ul	
93) Benzo(a)pyrene	24.389	252		77934m	2.860	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121324\
Data File : BG063675.D
Acq On : 13 Dec 2024 12:52
Operator : RC/JU
Sample : P5153-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
E28Q3

Manual IntegrationsAPPROVED

Quant Time: Dec 13 23:29:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024

