

Data Path : Z:\HPCHEM1\BNA G\DATA\BG121416\
 Data File : BG025208.D
 Acq On : 15 Dec 2016 12:01
 Operator : UM/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02008

Quant Time: Dec 15 12:58:58 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG113016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 04:58:54 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	137	0.00
2	1,4-Dioxane	0.504	0.438	13.1	122	0.00
3 S	1,4-Dioxane-d8	0.387	0.363	6.2	130	0.00
4	Benzaldehyde	1.299	1.387	-6.8	140	0.00
5 S	Phenol-d5	1.725	1.860	-7.8	148	0.00
6	Phenol	1.771	1.935	-9.3	145	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.067	1.152	-8.0	146	0.00
8	Bis(2-Chloroethyl)ether	1.296	1.332	-2.8	144	0.00
9 S	2-Chlorophenol-d4	1.207	1.254	-3.9	140	0.00
10	2-Chlorophenol	1.213	1.242	-2.4	139	0.00
11	2-Methylphenol	1.304	1.365	-4.7	144	0.00
12	2,2'-oxybis(1-Chloropropane	2.204	2.542	-15.3	154#	0.00
13 S	4-Methylphenol-d8	1.442	1.604	-11.2	155#	0.00
14	Acetophenone	2.194	2.349	-7.1	148	0.00
15 P	N-Nitroso-di-n-propylamine	1.267	1.367	-7.9	144	0.00
16	4-Methylphenol	1.451	1.553	-7.0	155#	0.00
17	Hexachloroethane	0.537	0.545	-1.5	136	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	149	0.00
19 S	Nitrobenzene-d5	0.142	0.144	-1.4	151#	0.00
20	Nitrobenzene	0.435	0.447	-2.8	158#	0.00
21	Isophorone	0.824	0.821	0.4	147	0.00
22 S	2-Nitrophenol-d4	0.177	0.185	-4.5	160#	0.00
23 C	2-Nitrophenol	0.179	0.179	0.0	150#	0.00
24	2,4-Dimethylphenol	0.410	0.427	-4.1	151#	0.00
25	Bis(2-Chloroethoxy)methane	0.420	0.425	-1.2	147	0.00
26 S	2,4-Dichlorophenol-d3	0.340	0.366	-7.6	160#	0.00
27 C	2,4-Dichlorophenol	0.330	0.348	-5.5	155#	0.00
28	Naphthalene	0.930	0.919	1.2	148	0.00
29 S	4-Chloroaniline-d4	0.334	0.406	-21.6#	181#	0.00
30	4-Chloroaniline	0.341	0.407	-19.4	174#	0.00
31 C	Hexachlorobutadiene	0.283	0.269	4.9	140	0.00
32	Caprolactam	0.137	0.131	4.4	145	0.00
33 C	4-Chloro-3-methylphenol	0.406	0.416	-2.5	151#	0.00
34	2-Methylnaphthalene	0.734	0.736	-0.3	145	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	146	0.00
36	1,2,4,5-Tetrachlorobenzene	0.603	0.610	-1.2	152#	0.00
37	Hexachlorocyclopentadiene	0.354	0.295	16.7	132	0.00
38 C	2,4,6-Trichlorophenol	0.379	0.401	-5.8	157#	0.00
39	2,4,5-Trichlorophenol	0.415	0.434	-4.6	150#	0.00
40	1,1'-Biphenyl	1.213	1.253	-3.3	149	0.00
41	2-Chloronaphthalene	0.965	1.005	-4.1	157#	0.00
42	2-Nitroaniline	0.378	0.428	-13.2	159#	0.00
43 S	Dimethylphthalate-d6	1.376	1.328	3.5	136	0.00
44	Dimethylphthalate	1.352	1.331	1.6	142	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG121416\
 Data File : BG025208.D
 Acq On : 15 Dec 2016 12:01
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02008

Quant Time: Dec 15 12:58:58 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG113016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 04:58:54 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.291	0.292	-0.3	143	0.00
46 S	Acenaphthylene-d8	1.507	1.560	-3.5	150	0.00
47	Acenaphthylene	1.465	1.477	-0.8	144	0.00
48	3-Nitroaniline	0.252	0.279	-10.7	155#	0.00
49 C	Acenaphthene	1.004	1.046	-4.2	152#	0.00
50	2,4-Dinitrophenol	0.189	0.191	-1.1	166#	0.00
51 S	4-Nitrophenol-d4	0.236	0.215	8.9	131	0.00
52	4-Nitrophenol	0.284	0.268	5.6	140	0.00
53	Dibenzofuran	1.587	1.583	0.3	144	0.00
54	2,4-Dinitrotoluene	0.438	0.419	4.3	138	0.00
55	2,3,4,6-Tetrachlorophenol	0.439	0.423	3.6	140	0.00
56	Diethylphthalate	1.432	1.355	5.4	135	0.00
57 S	Fluorene-d10	1.295	1.284	0.8	142	0.00
58	Fluorene	1.292	1.298	-0.5	143	0.00
59	4-Chlorophenyl-phenylether	0.728	0.713	2.1	142	0.00
60	4-Nitroaniline	0.303	0.268	11.6	129	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	133	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.124	0.128	-3.2	144	0.00
63	4,6-Dinitro-2-methylphenol	0.128	0.136	-6.3	146	0.00
64	N-Nitrosodiphenylamine	0.515	0.549	-6.6	139	0.00
65	4-Bromophenyl-phenylether	0.226	0.231	-2.2	136	0.00
66	Hexachlorobenzene	0.254	0.270	-6.3	137	0.00
67	Atrazine	0.245	0.232	5.3	122	0.00
68 C	Pentachlorophenol	0.153	0.152	0.7	137	0.00
69	Phenanthrene	0.955	0.971	-1.7	128	0.00
70 S	Anthracene-d10	0.844	0.860	-1.9	131	0.00
71	Anthracene	0.981	0.978	0.3	125	0.00
72	Carbazole	0.819	0.854	-4.3	128	0.00
73	Di-n-butylphthalate	1.034	1.020	1.4	125	0.00
74 C	Fluoranthene	1.134	1.145	-1.0	118	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
76 S	Pyrene-d10	0.824	0.842	-2.2	118	0.00
77	Pyrene	0.961	0.981	-2.1	115	0.00
78	Butylbenzylphthalate	0.376	0.402	-6.9	121	0.00
79	3,3'-Dichlorobenzidine	0.349	0.397	-13.8	129	0.00
80	Benzo(a)anthracene	1.039	1.075	-3.5	119	0.00
81	Bis(2-ethylhexyl)phthalate	0.518	0.539	-4.1	120	0.00
82	Chrysene	0.972	0.995	-2.4	118	0.00
83 I	Perylene-d12	1.000	1.000	0.0	114	0.00
84	Di-n-octyl phthalate	0.812	0.948	-16.7	123	0.00
85	Benzo(b)fluoranthene	0.992	1.025	-3.3	114	0.00
86	Benzo(k)fluoranthene	0.943	1.000	-6.0	118	0.00
87 S	Benzo(a)pyrene-d12	0.810	0.843	-4.1	115	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG121416\
Data File : BG025208.D
Acq On : 15 Dec 2016 12:01
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02008

Quant Time: Dec 15 12:58:58 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG113016.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 15 04:58:54 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	0.954	0.993	-4.1	115	0.00
89	Indeno(1,2,3-cd)pyrene	1.185	1.149	3.0	109	0.00
90	Dibenzo(a,h)anthracene	0.998	0.943	5.5	107	0.00
91	Benzo(a,h,i)perylene	0.989	0.875	11.5	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0