


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG121417\  
 Data File : BG031578.D  
 Acq On : 15 Dec 2017 3:18  
 Operator : SJ/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG121117.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Dec 11 17:25:55 2017



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2476

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 28.0      | 21186   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 27.9      | 21106   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.3       | 55      | PASS             |
| 127         | 198          | 10           | 80           | 38.7      | 29280   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 75736   | PASS             |
| 199         | 198          | 5            | 9            | 7.1       | 5368    | PASS             |
| 275         | 198          | 10           | 60           | 27.0      | 20479   | PASS             |
| 365         | 198          | 1            | 100          | 3.3       | 2506    | PASS             |
| 441         | 442          | 0.01         | 24           | 15.5      | 6981    | PASS             |
| 442         | 198          | 50           | 100          | 59.6      | 45162   | PASS             |
| 443         | 442          | 15           | 24           | 18.9      | 8554    | PASS             |