

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121421\
 Data File : BG051556.D
 Acq On : 16 Dec 2021 4:33
 Operator : CG/JU
 Sample : PB141371BL
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK371

Quant Time: Dec 16 05:35:33 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 14 14:19:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.278	152	28693	20.000	ng/u1	0.00
20) Naphthalene-d8	11.115	136	117085	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.910	164	89471	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.659	188	219802	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.977	240	220125	20.000	ng/u1	# 0.00
88) Perylene-d12	25.472	264	215924	20.000	ng/u1	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.602	96	4202	6.082	ng/uL	0.00
4) Pyridine-d5	4.031	84	66841	32.020	ng/u1	-0.01
7) Phenol-d5	7.403	99	83791	32.706	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.585	67	49655	32.604	ng/u1	0.00
11) 2-Chlorophenol-d4	7.802	132	58990	33.107	ng/u1	0.00
15) 4-Methylphenol-d8	8.971	113	64603	32.623	ng/u1	0.00
21) Nitrobenzene-d5	9.441	128	31148	36.285	ng/u1	-0.01
24) 2-Nitrophenol-d4	10.175	143	34078	36.315	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.722	165	67069	32.958	ng/u1	0.00
31) 4-Chloroaniline-d4	11.239	131	87135	32.522	ng/u1	0.00
46) Dimethylphthalate-d6	14.305	166	239103	33.389	ng/u1	0.00
49) Acenaphthylene-d8	14.605	160	283332	31.857	ng/u1	0.00
54) 4-Nitrophenol-d4	15.069	143	35757	30.705	ng/u1	0.00
60) Fluorene-d10	15.897	176	208688	32.526	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.997	200	34391	30.031	ng/u1	-0.01
73) Anthracene-d10	17.759	188	354114	33.773	ng/u1	0.00
81) Pyrene-d10	20.033	212	434615	32.806	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.231	264	395006	34.768	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121421\
Data File : BG051556.D
Acq On : 16 Dec 2021 4:33
Operator : CG/JU
Sample : PB141371BL
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK371

Quant Time: Dec 16 05:35:33 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 14 14:19:57 2021
Response via : Initial Calibration

