

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121421\
 Data File : BG051560.D
 Acq On : 16 Dec 2021 7:16
 Operator : CG/JU
 Sample : M4971-15
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BG3D2

Quant Time: Dec 16 08:10:14 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 14 14:19:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.278	152	32085	20.000	ng/u1	0.00
20) Naphthalene-d8	11.115	136	129613	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.910	164	96965	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.659	188	234775	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.977	240	237176	20.000	ng/u1	# 0.00
88) Perylene-d12	25.472	264	233937	20.000	ng/u1	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.608	96	3990	5.165	ng/uL	0.00
4) Pyridine-d5	4.037	84	26472	11.341	ng/u1	0.00
7) Phenol-d5	7.403	99	23603	8.239	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.579	67	56035	32.903	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.802	132	52734	26.467	ng/u1	0.00
15) 4-Methylphenol-d8	8.971	113	41084	18.553	ng/u1	0.00
21) Nitrobenzene-d5	9.447	128	33771	35.538	ng/u1	0.00
24) 2-Nitrophenol-d4	10.181	143	35821	34.483	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.722	165	67305	29.877	ng/u1	0.00
31) 4-Chloroaniline-d4	11.233	131	82037	27.659	ng/u1	0.00
46) Dimethylphthalate-d6	14.305	166	276194	35.587	ng/u1	0.00
49) Acenaphthylene-d8	14.605	160	319177	33.114	ng/u1	0.00
54) 4-Nitrophenol-d4	15.063	143	10730	8.502	ng/u1	-0.01
60) Fluorene-d10	15.897	176	242839	34.924	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.997	200	45258	37.000	ng/u1	-0.01
73) Anthracene-d10	17.753	188	421218	37.611	ng/u1	0.00
81) Pyrene-d10	20.033	212	515651	36.124	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.231	264	478439	38.869	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121421\
Data File : BG051560.D
Acq On : 16 Dec 2021 7:16
Operator : CG/JU
Sample : M4971-15
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BG3D2

Quant Time: Dec 16 08:10:14 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 14 14:19:57 2021
Response via : Initial Calibration

