

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121423\
 Data File : BG060016.D
 Acq On : 14 Dec 2023 17:42
 Operator : MA/JU
 Sample : 05718-04
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 WC-LIQUID-20231206

Quant Time: Dec 15 02:35:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 01:47:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	73536	20.000	ng	0.00
21) Naphthalene-d8	10.760	136	253627	20.000	ng	# 0.00
39) Acenaphthene-d10	14.591	164	237388	20.000	ng	0.00
64) Phenanthrene-d10	17.341	188	829548	20.000	ng	0.00
76) Chrysene-d12	21.606	240	812970	20.000	ng	0.00
86) Perylene-d12	24.797	264	959192	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.525	112	436697	101.659	ng	0.00
7) Phenol-d6	7.106	99	638328	105.638	ng	0.00
23) Nitrobenzene-d5	9.115	82	415417	74.374	ng	0.00
42) 2,4,6-Tribromophenol	16.078	330	738088	133.416	ng	0.00
45) 2-Fluorobiphenyl	13.216	172	1339332	69.052	ng	0.00
79) Terphenyl-d14	19.961	244	3873528	86.989	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121423\
Data File : BG060016.D
Acq On : 14 Dec 2023 17:42
Operator : MA/JU
Sample : 05718-04
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
WC-LIQUID-20231206

Quant Time: Dec 15 02:35:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 14 01:47:12 2023
Response via : Initial Calibration

