

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG121517\
 Data File : BG031623.D
 Acq On : 16 Dec 2017 13:06
 Operator : SJ/JU
 Sample : I6880-14
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 120817-JETT-F-U-S

Quant Time: Dec 18 03:28:40 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG121117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 11 17:25:55 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.10	152	70129	20.00	ng	0.00
21) Naphthalene-d8	10.91	136	305638	20.00	ng	0.00
38) Acenaphthene-d10	14.72	164	213417	20.00	ng	-0.01
63) Phenanthrene-d10	17.46	188	563321	20.00	ng	-0.01
75) Chrysene-d12	21.74	240	627809	20.00	ng	-0.01
86) Perylene-d12	25.02	264	527041	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.66	112	192422	48.63	ng	-0.01
7) Phenol-d6	7.26	99	154260	28.09	ng	0.00
23) Nitrobenzene-d5	9.27	82	480497	96.90	ng	0.00
41) 2,4,6-Tribromophenol	16.21	330	265465	106.17	ng	0.00
44) 2-Fluorobiphenyl	13.35	172	1449732	99.71	ng	0.00
78) Terphenyl-d14	20.06	244	2443711	88.56	ng	0.00


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG121517\
Data File : BG031623.D
Acq On : 16 Dec 2017 13:06
Operator : SJ/JU
Sample : I6880-14
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
120817-JETT-F-U-S

Quant Time: Dec 18 03:28:40 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG121117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 11 17:25:55 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.000 ng

RT: 8.10 min Scan# 819

Delta R.T. -0.01 min

Lab File: BG031623.D

Acq: 16 Dec 2017 13:06

Instrument :

BNA_G

ClientSampleId :

120817-JETT-F-U-S




Tgt Ion:152 Resp: 70129

Ion Ratio Lower Upper


152 100

150 154.9 126.6 189.8

115 59.5 53.0 79.4



Time-->



#5

2-Fluorophenol

Concen: 48.635 ng

RT: 5.66 min Scan# 403

Delta R.T. -0.01 min

Lab File: BG031623.D

Acq: 16 Dec 2017 13:06




Tgt Ion:112 Resp: 192422


Ion Ratio Lower Upper

112 100

64 53.0 56.3 84.5#

63 28.3 30.1 45.1#





#7

Phenol-d6

Concen: 28.086 ng

RT: 7.26 min Scan# 677

Delta R.T. -0.01 min

Lab File: BG031623.D

Acq: 16 Dec 2017 13:06

Instrument :

BNA_G

ClientSampleId :

120817-JETT-F-U-S





Tgt Ion: 99 Resp: 154260

Ion Ratio Lower Upper

99 100

42 17.0 14.1 21.1

71 32.7 26.1 39.1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.91 min Scan# 1298

Delta R.T. -0.01 min

Lab File: BG031623.D

Acq: 16 Dec 2017 13:06

Tgt Ion: 136 Resp: 305638




Ion Ratio Lower Upper


136 100

137 10.3 9.2 13.8

54 6.6 10.3 15.5#

68 5.1 4.5 6.7









#23
 Nitrobenzene-d5
 Concen: 96.900 ng
 RT: 9.27 min Scan# 1018
 Delta R.T. -0.01 min
 Lab File: BG031623.D
 Acq: 16 Dec 2017 13:06




Instrument :
 BNA_G
 ClientSampleId :
 120817-JETT-F-U-S


Tgt Ion: 82 Resp: 480497
 Ion Ratio Lower Upper
 82 100
 128 48.1 29.7 44.5#
 54 49.9 43.1 64.7



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.72 min Scan# 1946
 Delta R.T. -0.01 min
 Lab File: BG031623.D
 Acq: 16 Dec 2017 13:06

Tgt Ion: 164 Resp: 213417
 Ion Ratio Lower Upper
 164 100
 162 103.3 78.8 118.2
 160 45.1 35.4 53.0









#41
 2,4,6-Tribromophenol
 Concen: 106.175 ng
 RT: 16.21 min Scan# 2200
 Delta R.T. -0.01 min
 Lab File: BG031623.D
 Acq: 16 Dec 2017 13:06

Instrument :
 BNA_G
 ClientSampleId :
 120817-JETT-F-U-S

Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.1	77.0	115.6
141	34.4	25.2	37.8




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 99.712 ng
 RT: 13.35 min Scan# 1712
 Delta R.T. -0.01 min
 Lab File: BG031623.D
 Acq: 16 Dec 2017 13:06

Tgt Ion	Ratio	Lower	Upper
172	100		
171	38.4	30.0	45.0
170	24.8	19.5	29.3



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.46 min Scan# 2413
Delta R.T. -0.01 min
Lab File: BG031623.D
Acq: 16 Dec 2017 13:06




Instrument :
BNA_G
ClientSampleId :
120817-JETT-F-U-S


Tgt Ion:188 Resp: 563321
Ion Ratio Lower Upper
188 100
94 6.4 6.9 10.3#
80 7.4 7.7 11.5#



#75
Chrysene-d12
Concen: 20.000 ng
RT: 21.74 min Scan# 3140
Delta R.T. -0.01 min
Lab File: BG031623.D
Acq: 16 Dec 2017 13:06

Tgt Ion:240 Resp: 627809
Ion Ratio Lower Upper
240 100
120 7.7 6.8 10.2
236 26.2 20.9 31.3





#78

Terphenyl-d14

Concen: 88.559 ng

RT: 20.06 min Scan# 2855

Delta R.T. -0.01 min

Lab File: BG031623.D

Acq: 16 Dec 2017 13:06

Instrument :

BNA_G

ClientSampleId :

120817-JETT-F-U-S


Tgt Ion:244 Resp: 2443711

Ion Ratio Lower Upper

244 100

212 8.9 5.8 8.8#




122 8.8 7.0 10.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 25.02 min Scan# 3699

Delta R.T. -0.01 min

Lab File: BG031623.D

Acq: 16 Dec 2017 13:06


Tgt Ion:264 Resp: 527041

Ion Ratio Lower Upper

264 100

260 24.4 17.4 26.0

265 22.3 17.7 26.5



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

