


Data Path : Z:\HPCHEM1\BNA\_G\Data\BG121517\  
 Data File : BG031608.D  
 Acq On : 16 Dec 2017 3:43  
 Operator : SJ/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG121117.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Dec 11 17:25:55 2017



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2475

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 25.4      | 18853   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.1       | 217     | PASS             |
| 69          | 198          | 0.00         | 100          | 25.5      | 18962   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.7       | 136     | PASS             |
| 127         | 198          | 10           | 80           | 37.6      | 27933   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 74306   | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 5045    | PASS             |
| 275         | 198          | 10           | 60           | 28.4      | 21069   | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 2556    | PASS             |
| 441         | 442          | 0.01         | 24           | 15.6      | 7510    | PASS             |
| 442         | 198          | 50           | 100          | 64.6      | 48002   | PASS             |
| 443         | 442          | 15           | 24           | 19.4      | 9290    | PASS             |