

Data Path : Z:\HPCHEM1\BNA G\DATA\BG121615\
 Data File : BG020214.D
 Acq On : 16 Dec 2015 7:10
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02038

Quant Time: Dec 16 07:49:27 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG121415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 16 03:45:51 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	0.00
2	1,4-Dioxane	0.535	0.486	9.2	121	0.00
3 S	1,4-Dioxane-d8	0.362	0.343	5.2	112	0.00
4	Benzaldehyde	1.142	1.227	-7.4	128	0.00
5 S	Phenol-d5	1.764	1.783	-1.1	133	0.00
6	Phenol	1.889	1.938	-2.6	132	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.052	1.058	-0.6	127	0.00
8	Bis(2-Chloroethyl)ether	1.296	1.287	0.7	124	0.00
9 S	2-Chlorophenol-d4	1.283	1.325	-3.3	130	0.00
10	2-Chlorophenol	1.343	1.409	-4.9	131	0.00
11	2-Methylphenol	1.441	1.485	-3.1	132	0.00
12	2,2'-oxybis(1-Chloropropane	1.866	1.838	1.5	124	0.00
13 S	4-Methylphenol-d8	1.508	1.538	-2.0	129	0.00
14	Acetophenone	2.362	2.410	-2.0	128	0.00
15 P	N-Nitroso-di-n-propylamine	1.250	1.276	-2.1	126	0.00
16	4-Methylphenol	1.595	1.639	-2.8	129	0.00
17	Hexachloroethane	0.565	0.553	2.1	123	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	130	0.00
19 S	Nitrobenzene-d5	0.156	0.164	-5.1	134	0.00
20	Nitrobenzene	0.413	0.423	-2.4	131	0.00
21	Isophorone	0.789	0.789	0.0	125	0.00
22 S	2-Nitrophenol-d4	0.173	0.189	-9.2	137	0.00
23 C	2-Nitrophenol	0.186	0.205	-10.2	139	0.00
24	2,4-Dimethylphenol	0.422	0.433	-2.6	130	0.00
25	Bis(2-Chloroethoxy)methane	0.427	0.427	0.0	125	0.00
26 S	2,4-Dichlorophenol-d3	0.351	0.375	-6.8	135	0.00
27 C	2,4-Dichlorophenol	0.361	0.381	-5.5	133	0.00
28	Naphthalene	1.044	1.063	-1.8	128	0.00
29 S	4-Chloroaniline-d4	0.395	0.449	-13.7	138	0.00
30	4-Chloroaniline	0.403	0.447	-10.9	133	0.00
31 C	Hexachlorobutadiene	0.269	0.270	-0.4	128	0.00
32	Caprolactam	0.127	0.136	-7.1	131	0.00
33 C	4-Chloro-3-methylphenol	0.429	0.454	-5.8	133	0.00
34	2-Methylnaphthalene	0.798	0.819	-2.6	130	0.00
35	1-Methylnaphthalene	0.767	0.785	-2.3	131	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
37	1,2,4,5-Tetrachlorobenzene	0.751	0.779	-3.7	134	0.00
38	Hexachlorocyclopentadiene	0.464	0.440	5.2	133	0.00
39 C	2,4,6-Trichlorophenol	0.481	0.499	-3.7	136	0.00
40	2,4,5-Trichlorophenol	0.535	0.577	-7.9	137	0.00
41	1,1'-Biphenyl	1.509	1.537	-1.9	133	0.00
42	2-Chloronaphthalene	1.208	1.225	-1.4	132	0.00
43	2-Nitroaniline	0.411	0.429	-4.4	133	0.00
44 S	Dimethylphthalate-d6	1.618	1.683	-4.0	133	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG121615\
 Data File : BG020214.D
 Acq On : 16 Dec 2015 7:10
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02038

Quant Time: Dec 16 07:49:27 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG121415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 16 03:45:51 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.652	1.714	-3.8	134	0.00
46	2,6-Dinitrotoluene	0.342	0.357	-4.4	132	0.00
47 S	Acenaphthylene-d8	1.882	1.933	-2.7	131	0.00
48	Acenaphthylene	2.000	2.064	-3.2	133	0.00
49	3-Nitroaniline	0.332	0.370	-11.4	136	0.00
50 C	Acenaphthene	1.345	1.391	-3.4	133	0.00
51	2,4-Dinitrophenol	0.195	0.223	-14.4	165#	0.00
52 S	4-Nitrophenol-d4	0.290	0.297	-2.4	127	0.00
53	4-Nitrophenol	0.279	0.278	0.4	124	0.00
54	Dibenzofuran	2.001	2.036	-1.7	131	0.00
55	2,4-Dinitrotoluene	0.502	0.548	-9.2	135	0.00
56	2,3,4,6-Tetrachlorophenol	0.515	0.560	-8.7	137	0.00
57	Diethylphthalate	1.715	1.772	-3.3	131	0.00
58 S	Fluorene-d10	1.467	1.521	-3.7	135	0.00
59	Fluorene	1.607	1.663	-3.5	134	0.00
60	4-Chlorophenyl-phenylether	0.906	0.939	-3.6	134	0.00
61	4-Nitroaniline	0.359	0.391	-8.9	133	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.119	0.128	-7.6	148	0.00
64	4,6-Dinitro-2-methylphenol	0.127	0.133	-4.7	141	0.00
65	N-Nitrosodiphenylamine	0.556	0.578	-4.0	134	0.00
66	4-Bromophenyl-phenylether	0.235	0.245	-4.3	136	0.00
67	Hexachlorobenzene	0.252	0.266	-5.6	138	0.00
68	Atrazine	0.237	0.250	-5.5	133	0.00
69 C	Pentachlorophenol	0.152	0.160	-5.3	140	0.00
70	Phenanthrene	1.094	1.129	-3.2	133	0.00
71 S	Anthracene-d10	0.922	0.943	-2.3	131	0.00
72	Anthracene	1.110	1.143	-3.0	131	0.00
73	1,2,3,4-Tetrachlorobenzene	0.313	0.316	-1.0	132	0.00
74	Pentachlorobenzene	0.291	0.299	-2.7	133	0.00
75	Carbazole	0.932	0.966	-3.6	126	0.00
76	Di-n-butylphthalate	1.083	1.131	-4.4	130	0.00
77 C	Fluoranthene	1.279	1.355	-5.9	127	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	132	0.00
79 S	Pyrene-d10	0.932	0.935	-0.3	130	0.00
80	Pyrene	1.210	1.207	0.2	127	0.00
81	Butylbenzylphthalate	0.421	0.436	-3.6	134	0.00
82	3,3'-Dichlorobenzidine	0.365	0.383	-4.9	133	0.00
83	Benzo(a)anthracene	1.167	1.198	-2.7	131	0.00
84	Bis(2-ethylhexyl)phthalate	0.601	0.620	-3.2	133	0.00
85	Chrysene	1.099	1.115	-1.5	130	0.00
86 I	Perylene-d12	1.000	1.000	0.0	132	0.00
87	Di-n-octyl phthalate	1.099	1.117	-1.6	129	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG121615\
Data File : BG020214.D
Acq On : 16 Dec 2015 7:10
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02038

Quant Time: Dec 16 07:49:27 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG121415.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 16 03:45:51 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.204	1.219	-1.2	129	0.00
89	Benzo(k)fluoranthene	1.155	1.178	-2.0	133	0.00
90 S	Benzo(a)pyrene-d12	0.998	1.025	-2.7	133	0.00
91 C	Benzo(a)pyrene	1.141	1.173	-2.8	132	0.00
92	Indeno(1,2,3-cd)pyrene	1.359	1.388	-2.1	132	0.00
93	Dibenzo(a,h)anthracene	1.128	1.160	-2.8	133	0.00
94	Benzo(a,h,i)perylene	1.114	1.118	-0.4	132	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0