


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG121715\
Data File : BG020215.D
Acq On : 16 Dec 2015 7:49
Operator : UM/SJ
Sample : DFTPP08
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP08

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG121415.M
Title : SVOA CALIBRATION
Last Update : Thu Dec 17 06:39:33 2015



AutoFind: Scans 2561, 2562, 2563; Background Corrected with Scan 2555

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.1	14799	PASS
68	69	0.00	2	1.5	256	PASS
69	198	0.00	100	32.7	16644	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.2	20412	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	50837	PASS
199	198	5	9	6.9	3500	PASS
275	198	10	60	30.4	15445	PASS
365	198	1	100	3.9	1970	PASS
441	442	0.01	24	15.8	6937	PASS
442	198	50	100	86.2	43824	PASS
443	442	15	24	19.1	8358	PASS

m/z	Abundance
37.90	154.0
39.00	1281.0
39.90	265.0
43.90	312.0
49.00	338.0
50.00	3892.0
51.00	17480.0
52.00	813.0
55.90	639.0
57.00	1367.0
61.00	315.0
62.00	245.0
63.00	697.0
65.00	327.0
68.00	308.0
68.90	18696.0
73.00	289.0
74.00	1977.0
75.00	3198.0
76.00	965.0
77.00	20360.0
78.00	1349.0
79.00	1491.0
80.00	1212.0
81.00	1502.0
82.00	444.0
82.90	398.0
83.90	201.0
84.90	327.0
85.90	525.0
86.90	233.0
90.90	371.0
91.90	314.0
92.90	2422.0
93.90	202.0
96.00	184.0
98.00	1885.0
98.90	1662.0
100.90	860.0
102.90	357.0
103.90	523.0
104.90	595.0
106.00	280.0
107.00	6329.0
107.90	983.0
109.90	11458.0
110.90	1840.0
112.00	260.0
115.90	361.0
116.90	4855.0
117.90	380.0
121.90	471.0
122.90	765.0
123.90	424.0
124.80	372.0
126.90	22008.0
127.90	1504.0
128.90	9113.0
129.90	792.0
130.90	228.0
133.90	309.0
134.90	695.0
136.00	413.0
136.90	480.0
140.90	1281.0
141.90	512.0
142.90	351.0
145.90	382.0
146.90	847.0
147.90	1437.0
149.00	350.0
151.10	153.0
151.40	154.0
152.90	342.0
154.00	370.0
154.90	798.0
155.90	1299.0
156.90	246.0
157.90	199.0
158.90	237.0
159.90	413.0
160.90	678.0
161.90	205.0
164.90	542.0
165.90	343.0
167.00	2217.0
167.90	1241.0
169.00	218.0
171.90	270.0
172.90	489.0
173.90	517.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

174.90	907.0
175.90	223.0
176.80	472.0
177.90	238.0
178.90	2299.0
179.90	1315.0
181.00	644.0
183.90	218.0
184.90	1039.0
185.90	7520.0
186.90	2297.0
188.00	274.0
188.90	506.0
190.90	248.0
191.90	695.0
192.90	550.0
195.90	1243.0
197.90	50208.0
198.90	3485.0
199.90	266.0
201.40	265.0
202.90	342.0
203.90	2042.0
205.00	3709.0
206.00	14988.0
206.90	2167.0
207.90	478.0
208.90	162.0
209.90	329.0
210.90	413.0
214.90	226.0
216.00	366.0
216.90	4258.0
217.90	428.0
220.90	2450.0
221.90	581.0
222.90	1174.0
224.00	8304.0
224.90	2305.0
226.00	301.0
226.90	4054.0
227.90	648.0
228.90	729.0
230.90	295.0
233.90	178.0
234.90	200.0
235.90	255.0
237.00	226.0
238.80	165.0
240.80	222.0
241.90	396.0
242.90	582.0
244.00	5768.0
244.90	903.0
245.90	1578.0
246.90	266.0
248.90	211.0
252.90	269.0
254.90	31192.0
255.90	4952.0
256.90	499.0
257.90	2253.0
258.90	276.0
264.90	778.0
270.90	282.0
271.80	189.0
272.90	1174.0
273.90	2623.0
274.90	14419.0
275.90	1908.0
276.90	1405.0
277.90	224.0
292.90	230.0
295.90	4100.0
296.90	652.0
302.90	437.0
304.00	157.0
314.00	173.0
314.80	356.0
315.90	225.0
323.00	1291.0
323.90	235.0
326.80	214.0
328.00	188.0
331.90	191.0
334.00	829.0
340.90	159.0
345.90	238.0
351.90	322.0
353.00	239.0
353.90	500.0
364.90	1788.0
365.90	308.0
372.00	662.0
383.00	194.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

402.00	226.0
402.90	199.0
420.90	261.0
422.00	334.0
422.90	1995.0
423.90	385.0
441.00	5216.0
442.00	33560.0
443.00	6284.0
444.00	782.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
38.00	244.0
39.00	1398.0
39.90	268.0
43.90	252.0
49.00	266.0
50.00	4742.0
51.00	17056.0
52.00	1192.0
55.90	541.0
56.90	1173.0
60.90	327.0
61.90	318.0
62.90	727.0
65.00	400.0
68.10	462.0
68.90	19752.0
72.90	236.0
74.00	2104.0
74.90	3093.0
76.00	1109.0
77.00	22112.0
78.00	1316.0
79.00	1501.0
79.90	1216.0
80.90	1423.0
81.90	457.0
83.00	304.0
83.90	166.0
84.90	215.0
85.90	477.0
86.90	238.0
90.90	450.0
91.90	322.0
92.90	2389.0
93.80	162.0
98.00	1806.0
98.90	1462.0
99.90	159.0
101.00	869.0
102.90	257.0
104.00	594.0
104.90	544.0
106.00	233.0
107.00	7287.0
107.90	1203.0
109.90	12303.0
110.90	2256.0
112.00	253.0
115.90	475.0
116.90	5753.0
117.90	344.0
121.90	338.0
122.90	793.0
123.90	330.0
125.00	352.0
127.00	24192.0
128.00	1927.0
128.90	9514.0
129.90	1036.0
131.00	189.0
133.90	370.0
134.90	850.0
135.90	392.0
136.90	391.0
137.80	151.0
139.90	232.0
140.90	1668.0
142.00	428.0
142.90	238.0
146.00	221.0
146.90	717.0
147.90	1728.0
148.90	277.0
150.90	218.0
152.90	496.0
153.90	303.0
155.00	921.0
155.90	1202.0
156.80	256.0
157.80	260.0
158.90	158.0
160.00	396.0
160.90	580.0
161.90	187.0
164.90	567.0
166.00	615.0
166.90	2584.0
168.00	1155.0
168.90	213.0
171.90	182.0
172.90	312.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

174.00	846.0
174.90	1196.0
175.90	366.0
176.90	465.0
178.00	201.0
178.90	2073.0
180.00	1702.0
181.00	681.0
183.90	272.0
184.90	1233.0
186.00	8701.0
187.00	2462.0
187.90	293.0
188.90	558.0
189.90	198.0
190.90	352.0
191.90	710.0
192.90	904.0
193.90	210.0
195.90	1667.0
197.90	59976.0
198.90	4049.0
199.90	381.0
201.30	288.0
202.90	620.0
203.90	2653.0
204.90	4076.0
206.00	17624.0
206.90	2377.0
207.90	624.0
208.90	322.0
209.80	207.0
210.90	746.0
214.80	212.0
215.90	445.0
216.90	4980.0
217.90	901.0
220.90	2283.0
221.80	571.0
222.90	1220.0
223.90	10086.0
224.90	2631.0
225.90	408.0
226.90	4852.0
228.00	489.0
228.90	862.0
230.00	203.0
230.90	336.0
233.80	313.0
234.90	254.0
235.90	200.0
237.00	462.0
238.90	169.0
240.00	171.0
240.90	350.0
241.90	577.0
243.00	615.0
244.00	7146.0
245.00	871.0
245.90	1687.0
246.90	426.0
248.80	382.0
252.70	231.0
254.90	41472.0
255.90	5822.0
256.90	500.0
257.90	2842.0
258.90	403.0
264.90	1163.0
265.90	166.0
271.90	225.0
272.90	1297.0
273.90	3253.0
274.90	18848.0
275.90	2233.0
276.90	1338.0
277.90	363.0
284.00	190.0
284.90	411.0
292.90	365.0
293.80	172.0
295.00	183.0
295.90	5760.0
297.00	830.0
302.90	613.0
303.90	275.0
313.90	252.0
314.90	608.0
315.90	387.0
321.00	193.0
323.00	1456.0
324.00	261.0
326.90	258.0
328.00	216.0
332.00	161.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

332.90	166.0
334.00	867.0
334.90	223.0
341.00	245.0
345.90	251.0
351.90	580.0
352.90	343.0
354.00	360.0
364.90	1988.0
365.90	352.0
371.90	808.0
372.90	179.0
382.90	318.0
401.90	343.0
402.90	378.0
404.00	212.0
420.90	461.0
421.90	376.0
423.00	2934.0
424.00	515.0
441.00	8332.0
442.00	53424.0
443.00	10408.0
444.00	982.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
39.00	741.0
40.00	178.0
44.00	321.0
49.00	248.0
50.00	2849.0
51.00	9862.0
52.00	462.0
56.00	346.0
56.90	787.0
61.00	167.0
62.00	204.0
62.90	526.0
64.90	318.0
68.90	11486.0
74.00	1377.0
74.90	1964.0
75.90	658.0
77.00	13556.0
78.00	629.0
78.90	979.0
79.90	894.0
80.90	1006.0
81.90	382.0
82.90	248.0
83.90	165.0
84.90	239.0
86.00	354.0
90.90	181.0
91.90	304.0
92.90	1697.0
97.90	1280.0
99.00	958.0
101.00	626.0
102.90	293.0
103.90	452.0
104.90	379.0
107.00	4438.0
107.90	728.0
109.90	7939.0
110.90	1434.0
115.90	353.0
116.90	3337.0
117.90	391.0
122.00	314.0
122.90	460.0
123.90	247.0
124.90	204.0
127.00	15038.0
128.00	1356.0
128.90	6706.0
130.00	556.0
133.90	269.0
134.90	446.0
135.90	228.0
136.90	362.0
140.90	960.0
141.80	165.0
142.90	239.0
146.80	397.0
147.90	1199.0
148.80	224.0
152.90	272.0
153.80	197.0
154.90	560.0
156.00	757.0
157.90	153.0
160.00	349.0
161.00	343.0
164.90	361.0
165.90	357.0
166.90	1932.0
167.90	746.0
168.80	163.0
170.90	165.0
171.90	161.0
172.90	186.0
173.90	495.0
174.90	814.0
176.00	277.0
176.90	443.0
178.10	205.0
178.90	1496.0
179.90	954.0
180.90	440.0
184.90	895.0
185.90	5257.0
186.90	1678.0
188.00	204.0
188.90	353.0
190.90	221.0
191.90	445.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

192.90	585.0
195.90	1355.0
197.90	42328.0
198.90	2967.0
199.90	258.0
200.90	162.0
201.30	181.0
202.90	340.0
204.00	1772.0
204.90	2577.0
206.00	11410.0
206.90	1714.0
207.90	507.0
209.00	170.0
210.20	180.0
210.80	473.0
215.90	381.0
216.90	3639.0
217.90	441.0
220.90	1547.0
221.70	445.0
222.90	799.0
223.90	6908.0
224.90	1834.0
225.90	180.0
226.90	3076.0
227.90	360.0
228.90	410.0
230.90	221.0
233.90	206.0
234.90	186.0
237.00	230.0
240.90	205.0
242.00	274.0
242.90	561.0
244.00	4856.0
244.90	736.0
245.90	1164.0
246.90	327.0
248.80	320.0
254.90	28056.0
255.90	4380.0
256.90	369.0
257.90	1298.0
258.80	272.0
264.90	754.0
272.90	931.0
273.90	2173.0
274.90	13068.0
275.90	1614.0
276.90	1121.0
277.80	202.0
282.90	155.0
285.00	185.0
292.90	272.0
295.90	3855.0
296.90	536.0
303.00	438.0
314.90	511.0
315.90	233.0
320.90	195.0
323.00	1335.0
326.90	331.0
327.90	153.0
332.90	216.0
333.90	700.0
335.00	235.0
340.90	213.0
345.90	260.0
352.00	413.0
352.90	208.0
353.90	432.0
364.90	2136.0
365.90	279.0
372.00	727.0
372.90	202.0
382.90	226.0
389.80	187.0
401.90	329.0
402.90	364.0
403.90	167.0
420.90	455.0
421.90	403.0
423.00	2241.0
424.00	521.0
441.00	7265.0
442.00	44488.0
443.00	8383.0
443.90	916.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
39.90	210.0
207.00	161.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08