

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121724\
 Data File : BG063705.D
 Acq On : 17 Dec 2024 19:17
 Operator : RC/JU
 Sample : PB165588BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS588

Quant Time: Dec 18 02:22:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 23:59:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.808	152	135403	20.000	ng/u1	0.00
20) Naphthalene-d8	10.593	136	600946	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.447	164	491814	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.197	188	1180842	20.000	ng/u1	0.00
79) Chrysene-d12	21.445	240	1106248	20.000	ng/u1	-0.02
88) Perylene-d12	24.453	264	1149991	20.000	ng/u1	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.266	96	19163	5.546	ng/uL	0.00
4) Pyridine-d5	3.671	84	279175	25.613	ng/u1	-0.02
7) Phenol-d5	6.991	99	389469	31.231	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.138	67	240640	27.939	ng/u1	0.00
11) 2-Chlorophenol-d4	7.344	132	249935	30.958	ng/u1	0.00
15) 4-Methylphenol-d8	8.525	113	293685	30.333	ng/u1	0.00
21) Nitrobenzene-d5	8.959	128	133089	31.017	ng/u1	0.00
24) 2-Nitrophenol-d4	9.676	143	161813	33.643	ng/u1	-0.02
28) 2,4-Dichlorophenol-d3	10.228	165	305673	32.825	ng/u1	0.00
31) 4-Chloroaniline-d4	10.734	131	351102	29.694	ng/u1	0.00
46) Dimethylphthalate-d6	13.854	166	1060318	31.684	ng/u1	0.00
49) Acenaphthylene-d8	14.136	160	1160784	30.428	ng/u1	0.00
54) 4-Nitrophenol-d4	14.682	143	170932	36.049	ng/u1	0.00
60) Fluorene-d10	15.440	176	972905	32.881	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	175636	28.647	ng/u1	0.00
73) Anthracene-d10	17.297	188	1646309	32.428	ng/u1	0.00
81) Pyrene-d10	19.594	212	1891457	32.857	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.253	264	1764728	32.678	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.295	88	39923	9.763	ng/uL#	95
5) Pyridine	3.689	79	280464	24.023	ng/u1	94
6) Benzaldehyde	6.944	77	166216	21.729	ng/u1	89
8) Phenol	7.020	94	403296	30.345	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.232	93	280135	27.549	ng/u1	98
12) 2-Chlorophenol	7.379	128	251752m	30.806	ng/u1	
13) 2-Methylphenol	8.260	108	291389	30.335	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.325	45	421427	27.917	ng/u1	97
16) Acetophenone	8.624	105	480734	29.274	ng/u1	92
17) N-Nitroso-di-n-propyla...	8.607	70	273444	27.906	ng/u1	94
18) 4-Methylphenol	8.589	108	315941	30.052	ng/u1	100
19) Hexachloroethane	8.877	117	105406	26.250	ng/u1	97
22) Nitrobenzene	9.000	77	392399	27.386	ng/u1	95
23) Isophorone	9.523	82	761441	28.126	ng/u1	98
25) 2-Nitrophenol	9.711	139	160316	32.433	ng/u1#	87
26) 2,4-Dimethylphenol	9.782	107	246869	22.009	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.005	93	419828	28.671	ng/u1#	95
29) 2,4-Dichlorophenol	10.252	162	294446	31.822	ng/u1	98
30) Naphthalene	10.646	128	887110	29.526	ng/u1	99
32) 4-Chloroaniline	10.757	127	285981	25.421	ng/u1	99
33) Hexachlorobutadiene	10.928	225	189904	25.263	ng/u1	99
34) Caprolactam	11.515	113	119045m	37.219	ng/u1	
35) 4-Chloro-3-methylphenol	11.903	107	373734	34.404	ng/u1	98
36) 2-Methylnaphthalene	12.255	142	662446	30.678	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121724\
 Data File : BG063705.D
 Acq On : 17 Dec 2024 19:17
 Operator : RC/JU
 Sample : PB165588BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS588

Quant Time: Dec 18 02:22:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 11 23:59:23 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.479	142	667183	30.332	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.632	216	406958	25.848	ng/ul	97
40) Hexachlorocyclopentadiene	12.608	237	59367	9.538	ng/ul	99
41) 2,4,6-Trichlorophenol	12.878	196	252203	28.342	ng/ul	95
42) 2,4,5-Trichlorophenol	12.961	196	281018	30.260	ng/ul	94
43) 1,1'-Biphenyl	13.272	154	944387	29.117	ng/ul	96
44) 2-Chloronaphthalene	13.313	162	757731	29.093	ng/ul	98
45) 2-Nitroaniline	13.525	65	299394	34.318	ng/ul	97
47) Dimethylphthalate	13.901	163	1038116	30.952	ng/ul	100
48) 2,6-Dinitrotoluene	14.024	165	206348	33.127	ng/ul#	92
50) Acenaphthylene	14.165	152	1236775	30.235	ng/ul	96
51) 3-Nitroaniline	14.353	138	223748	38.851	ng/ul#	91
52) Acenaphthene	14.512	153	873969	30.144	ng/ul	95
53) 2,4-Dinitrophenol	14.576	184	77040m	24.201	ng/ul	
55) 4-Nitrophenol	14.694	109	145828	29.618	ng/ul	90
56) Dibenzofuran	14.847	168	1273988	31.168	ng/ul	98
57) 2,4-Dinitrotoluene	14.823	165	334080	36.094	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	15.082	232	223633	27.936	ng/ul	96
59) Diethylphthalate	15.270	149	1051706	32.436	ng/ul	98
61) Fluorene	15.499	166	1037491	31.787	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.487	204	543469	29.888	ng/ul	98
63) 4-Nitroaniline	15.522	138	239328	41.510	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.593	198	178281	28.226	ng/ul#	89
67) N-Nitrosodiphenylamine	15.704	169	931739	32.584	ng/ul	97
68) 4-Bromophenyl-phenylether	16.386	248	365275	30.353	ng/ul	94
69) Hexachlorobenzene	16.509	284	417337	30.952	ng/ul	96
70) Atrazine	16.656	200	154782	12.805	ng/ul	96
71) Pentachlorophenol	16.862	266	128071m	22.528	ng/ul	
72) Phenanthrene	17.238	178	1857836	32.648	ng/ul	98
74) Anthracene	17.332	178	1835025	31.861	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.243	216	435416	27.637	ng/ul	97
76) Pentachlorobenzene	14.770	250	447246	28.655	ng/ul	100
77) Carbazole	17.602	167	1650044	35.392	ng/ul	99
78) Di-n-butylphthalate	18.160	149	1996664	35.462	ng/ul	99
80) Fluoranthene	19.259	202	2159600	33.006	ng/ul	96
82) Pyrene	19.623	202	2275718	32.691	ng/ul	96
83) Butylbenzylphthalate	20.516	149	834026	39.030	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.345	252	658366	32.968	ng/ul	96
85) Benzo(a)anthracene	21.427	228	2138318	30.843	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.339	149	1193467	38.394	ng/ul	98
87) Chrysene	21.486	228	2000503	30.275	ng/ul	96
89) Di-n-octyl phthalate	22.467	149	2051257	41.326	ng/ul	100
90) Benzo(b)fluoranthene	23.507	252	2113875	32.342	ng/ul	100
91) Benzo(k)fluoranthene	23.566	252	2087593	32.020	ng/ul	99
93) Benzo(a)pyrene	24.318	252	1898615	30.948	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	27.861	276	2460762	32.844	ng/ul	96
95) Dibenzo(a,h)anthracene	27.908	278	2012687	33.573	ng/ul	98
96) Benzo(g,h,i)perylene	28.918	276	1929601	32.456	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121724\
Data File : BG063705.D
Acq On : 17 Dec 2024 19:17
Operator : RC/JU
Sample : PB165588BS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SLCS588

Quant Time: Dec 18 02:22:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 11 23:59:23 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.763 ng/uL
RT: 3.295 min Scan# 31
Delta R.T. -0.016 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion: 88 Resp: 39923
Ion Ratio Lower Upper
88 100
43 31.4 20.9 31.3#
58 74.2 57.7 86.5



#5
Pyridine
Concen: 24.023 ng/uL
RT: 3.689 min Scan# 102
Delta R.T. -0.016 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion: 79 Resp: 280464
Ion Ratio Lower Upper
79 100
52 69.6 60.9 91.3
51 34.5 29.5 44.3









#6
Benzaldehyde
Concen: 21.729 ng/ul
RT: 6.944 min Scan# 61
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion: 77 Resp: 166216
Ion Ratio Lower Upper
77 100
105 89.6 65.2 97.8
106 88.1 61.0 91.4



#8
Phenol
Concen: 30.345 ng/ul
RT: 7.020 min Scan# 669
Delta R.T. 0.002 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion: 94 Resp: 403296
Ion Ratio Lower Upper
94 100
65 28.1 25.6 38.4
66 40.8 34.2 51.2









#10
Bis(2-Chloroethyl)ether
Concen: 27.549 ng/ul
RT: 7.232 min Scan# 704
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion: 93 Resp: 280135
Ion Ratio Lower Upper
93 100
63 79.3 63.8 95.6
95 30.4 27.6 41.4



#12
2-Chlorophenol
Concen: 30.806 ng/ul m
RT: 7.379 min Scan# 730
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:128 Resp: 251752
Ion Ratio Lower Upper
128 100
64 62.9 48.9 73.3
130 36.5 24.2 36.2#









#13
2-Methylphenol
Concen: 30.335 ng/ul
RT: 8.260 min Scan# 81
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:108 Resp: 291389
Ion Ratio Lower Upper
108 100
107 87.2 72.6 108.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 27.917 ng/ul
RT: 8.325 min Scan# 891
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion: 45 Resp: 421427
Ion Ratio Lower Upper
45 100
77 15.4 11.5 17.3
79 9.9 9.2 13.8









#16
Acetophenone
Concen: 29.274 ng/ul
RT: 8.624 min Scan# 941
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:105 Resp: 480734
Ion Ratio Lower Upper
105 100
77 84.6 73.6 110.4
51 27.0 25.4 38.2



#17
N-Nitroso-di-n-propylamine
Concen: 27.906 ng/ul
RT: 8.607 min Scan# 939
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion: 70 Resp: 273444
Ion Ratio Lower Upper
70 100
42 54.1 47.4 71.0
101 7.2 6.1 9.1
130 16.3 11.4 17.2









#18
4-Methylphenol
Concen: 30.052 ng/ul
RT: 8.589 min Scan# 91
Delta R.T. 0.002 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:108 Resp: 315941
Ion Ratio Lower Upper
108 100
107 112.0 89.2 133.8



#19
Hexachloroethane
Concen: 26.250 ng/ul
RT: 8.877 min Scan# 985
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:117 Resp: 105406
Ion Ratio Lower Upper
117 100
201 110.7 86.8 130.2
199 70.6 59.4 89.2









#22
Nitrobenzene
Concen: 27.386 ng/ul
RT: 9.000 min Scan# 1006
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion: 77 Resp: 392399
Ion Ratio Lower Upper
77 100
123 36.8 27.0 40.4
65 12.0 10.8 16.2



#23
Isophorone
Concen: 28.126 ng/ul
RT: 9.523 min Scan# 1095
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion: 82 Resp: 761441
Ion Ratio Lower Upper
82 100
95 6.5 5.8 8.6
138 13.6 10.3 15.5









#25
2-Nitrophenol
Concen: 32.433 ng/ul
RT: 9.711 min Scan# 1127
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:139 Resp: 160316
Ion Ratio Lower Upper
139 100
65 61.5 62.2 93.2#
109 34.0 26.9 40.3



#26
2,4-Dimethylphenol
Concen: 22.009 ng/ul
RT: 9.782 min Scan# 1139
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:107 Resp: 246869
Ion Ratio Lower Upper
107 100
121 49.7 38.4 57.6
122 77.8 61.8 92.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 28.671 ng/ul
 RT: 10.005 min Scan# 1177
 Delta R.T. -0.010 min
 Lab File: BG063705.D
 Acq: 17 Dec 2024 19:17




Instrument :
 BNA_G
 ClientSampleId :
 SLCS588


Tgt Ion: 93 Resp: 419828
 Ion Ratio Lower Upper
 93 100
 95 32.4 23.5 35.3
 123 9.6 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 31.822 ng/ul
 RT: 10.252 min Scan# 1219
 Delta R.T. -0.004 min
 Lab File: BG063705.D
 Acq: 17 Dec 2024 19:17

Tgt Ion:162 Resp: 294446
 Ion Ratio Lower Upper
 162 100
 164 64.2 53.0 79.4
 98 40.9 33.0 49.6









#30
Naphthalene
Concen: 29.526 ng/ul
RT: 10.646 min Scan# 1286
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:128 Resp: 887110
Ion Ratio Lower Upper
128 100
129 10.8 8.2 12.2
127 12.8 9.8 14.8



#32
4-Chloroaniline
Concen: 25.421 ng/ul
RT: 10.757 min Scan# 1305
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:127 Resp: 285981
Ion Ratio Lower Upper
127 100
129 30.2 24.5 36.7









#33
Hexachlorobutadiene
Concen: 25.263 ng/ul
RT: 10.928 min Scan# 11
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:225 Resp: 189904
Ion Ratio Lower Upper
225 100
223 64.1 51.2 76.8
227 65.9 51.1 76.7



#34
Caprolactam
Concen: 37.219 ng/ul m
RT: 11.515 min Scan# 1434
Delta R.T. 0.008 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:113 Resp: 119045
Ion Ratio Lower Upper
113 100
55 193.5 162.3 243.5
56 141.7 107.1 160.7









#35
4-Chloro-3-methylphenol
Concen: 34.404 ng/ul
RT: 11.903 min Scan# 11900
Delta R.T. 0.002 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:107 Resp: 373734
Ion Ratio Lower Upper
107 100
144 23.0 16.3 24.5
142 69.0 55.0 82.4



#36
2-Methylnaphthalene
Concen: 30.678 ng/ul
RT: 12.255 min Scan# 1560
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:142 Resp: 662446
Ion Ratio Lower Upper
142 100
141 87.5 69.5 104.3









#37
1-Methylnaphthalene
Concen: 30.332 ng/ul
RT: 12.479 min Scan# 1598
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:142 Resp: 667183
Ion Ratio Lower Upper
142 100
141 91.3 73.8 110.6
116 3.7 3.5 5.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 25.848 ng/ul
RT: 12.632 min Scan# 1624
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:216 Resp: 406958
Ion Ratio Lower Upper
216 100
214 74.9 62.6 93.8
179 20.2 16.5 24.7
108 14.4 11.2 16.8









#40
Hexachlorocyclopentadiene
Concen: 9.538 ng/ul
RT: 12.608 min Scan# 1
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:237 Resp: 59367
Ion Ratio Lower Upper
237 100
235 63.6 51.0 76.6
272 13.3 9.7 14.5



#41
2,4,6-Trichlorophenol
Concen: 28.342 ng/ul
RT: 12.878 min Scan# 1666
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:196 Resp: 252203
Ion Ratio Lower Upper
196 100
198 94.7 72.4 108.6
200 32.4 22.5 33.7





#42

2,4,5-Trichlorophenol

Concen: 30.260 ng/ul

RT: 12.961 min Scan# 1

Delta R.T. -0.004 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17

Instrument :

BNA_G

ClientSampleId :

SLCS588





Tgt Ion:196 Resp: 281018

Ion Ratio Lower Upper

196 100

198 98.5 74.2 111.4

200 32.2 23.5 35.3



#43

1,1'-Biphenyl

Concen: 29.117 ng/ul

RT: 13.272 min Scan# 1733

Delta R.T. -0.010 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17




Tgt Ion:154 Resp: 944387


Ion Ratio Lower Upper

154 100

153 40.5 35.1 52.7

76 16.3 13.4 20.2





#44

2-Chloronaphthalene

Concen: 29.093 ng/ul

RT: 13.313 min Scan# 1

Delta R.T. -0.010 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17

Instrument :

BNA_G

ClientSampleId :

SLCS588





Tgt Ion:162 Resp: 757731

Ion Ratio Lower Upper

162 100

164 33.0 25.2 37.8

127 35.3 29.2 43.8



#45

2-Nitroaniline

Concen: 34.318 ng/ul

RT: 13.525 min Scan# 1776

Delta R.T. -0.004 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17




Tgt Ion: 65 Resp: 299394


Ion Ratio Lower Upper

65 100

92 63.2 49.5 74.3

138 81.1 61.8 92.8









#47
Dimethylphthalate
Concen: 30.952 ng/ul
RT: 13.901 min Scan# 1840
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:163 Resp: 1038116
Ion Ratio Lower Upper
163 100
194 3.4 3.0 4.4
164 11.0 8.8 13.2



#48
2,6-Dinitrotoluene
Concen: 33.127 ng/ul
RT: 14.024 min Scan# 1861
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:165 Resp: 206348
Ion Ratio Lower Upper
165 100
89 68.6 59.4 89.2
121 20.6 20.8 31.2#





#50

Acenaphthylene

Concen: 30.235 ng/ul

RT: 14.165 min Scan# 1885

Delta R.T. -0.010 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17

Instrument :

BNA_G

ClientSampleId :

SLCS588





Tgt Ion:152 Resp: 1236775

Ion Ratio Lower Upper

152 100

151 20.6 14.6 22.0

153 13.0 10.0 15.0



#51

3-Nitroaniline

Concen: 38.851 ng/ul

RT: 14.353 min Scan# 1917

Delta R.T. -0.010 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17




Tgt Ion:138 Resp: 223748


Ion Ratio Lower Upper

138 100

108 12.7 8.4 12.6#


92 139.2 120.4 180.6








#52
Acenaphthene
Concen: 30.144 ng/ul
RT: 14.512 min Scan# 1944
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Instrument :
BNA_G
ClientSampleId :
SLCS588







Tgt Ion:153 Resp: 873969
Ion Ratio Lower Upper
153 100
152 45.9 40.2 60.4
154 85.0 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 24.201 ng/ul m
RT: 14.576 min Scan# 1955
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:184 Resp: 77040
Ion Ratio Lower Upper
184 100
63 79.1 52.1 78.1#
154 61.6 47.8 71.6









#55
4-Nitrophenol
Concen: 29.618 ng/ul
RT: 14.694 min Scan# 1975
Delta R.T. 0.008 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:109 Resp: 145828
Ion Ratio Lower Upper
109 100
139 116.8 86.4 129.6
65 143.5 127.0 190.4



#56
Dibenzofuran
Concen: 31.168 ng/ul
RT: 14.847 min Scan# 2001
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:168 Resp: 1273988
Ion Ratio Lower Upper
168 100
139 33.7 27.8 41.8









#57
2,4-Dinitrotoluene
Concen: 36.094 ng/ul
RT: 14.823 min Scan# 1997
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:165 Resp: 334080
Ion Ratio Lower Upper
165 100
63 49.3 45.0 67.4
182 2.2 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 27.936 ng/ul
RT: 15.082 min Scan# 2041
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:232 Resp: 223633
Ion Ratio Lower Upper
232 100
131 40.5 30.3 45.5
130 2.5 1.8 2.6
166 29.0 24.9 37.3









#59
Diethylphthalate
Concen: 32.436 ng/ul
RT: 15.270 min Scan# 2073
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:149 Resp: 1051706
Ion Ratio Lower Upper
149 100
177 22.8 17.6 26.4
150 12.3 9.0 13.4



#61
Fluorene
Concen: 31.787 ng/ul
RT: 15.499 min Scan# 2112
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:166 Resp: 1037491
Ion Ratio Lower Upper
166 100
165 100.4 81.9 122.9
167 14.1 11.4 17.0









#62
4-Chlorophenyl-phenylether
Concen: 29.888 ng/ul
RT: 15.487 min Scan# 2110
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:204 Resp: 543469
Ion Ratio Lower Upper
204 100
206 31.8 25.6 38.4
141 50.9 42.3 63.5



#63
4-Nitroaniline
Concen: 41.510 ng/ul
RT: 15.522 min Scan# 2116
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:138 Resp: 239328
Ion Ratio Lower Upper
138 100
92 65.8 53.2 79.8
108 99.0 79.4 119.0









#66
4,6-Dinitro-2-methylphenol
Concen: 28.226 ng/ul
RT: 15.593 min Scan# 2128
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:198 Resp: 178281
Ion Ratio Lower Upper
198 100
182 4.0 3.4 5.2
77 24.4 25.2 37.8#



#67
N-Nitrosodiphenylamine
Concen: 32.584 ng/ul
RT: 15.704 min Scan# 2147
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:169 Resp: 931739
Ion Ratio Lower Upper
169 100
168 65.2 54.0 81.0
167 36.0 29.8 44.8









#68
4-Bromophenyl-phenylether
Concen: 30.353 ng/ul
RT: 16.386 min Scan# 21
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:248 Resp: 365275
Ion Ratio Lower Upper
248 100
250 100.5 76.1 114.1
141 62.9 46.0 69.0



#69
Hexachlorobenzene
Concen: 30.952 ng/ul
RT: 16.509 min Scan# 2284
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:284 Resp: 417337
Ion Ratio Lower Upper
284 100
142 28.1 23.9 35.9
249 32.9 24.6 36.8









#70
Atrazine
Concen: 12.805 ng/ul
RT: 16.656 min Scan# 21
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:200 Resp: 154782
Ion Ratio Lower Upper
200 100
173 24.4 21.3 31.9
215 44.9 37.8 56.8



#71
Pentachlorophenol
Concen: 22.528 ng/ul m
RT: 16.862 min Scan# 2344
Delta R.T. -0.004 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:266 Resp: 128071
Ion Ratio Lower Upper
266 100
264 64.1 49.3 73.9
268 64.0 50.5 75.7





#72

Phenanthrene

Concen: 32.648 ng/ul

RT: 17.238 min Scan# 2408

Delta R.T. -0.010 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17

Instrument :

BNA_G

ClientSampleId :

SLCS588





Tgt Ion:178 Resp: 1857836

Ion Ratio Lower Upper

178 100

179 16.7 12.2 18.2

176 19.0 15.8 23.6



#74

Anthracene

Concen: 31.861 ng/ul

RT: 17.332 min Scan# 2424

Delta R.T. -0.004 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17




Tgt Ion:178 Resp: 1835025


Ion Ratio Lower Upper

178 100

179 15.7 12.4 18.6

176 19.2 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 27.637 ng/ul

RT: 13.243 min Scan# 1

Delta R.T. -0.004 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17

Instrument :

BNA_G

ClientSampleId :

SLCS588





Tgt Ion:216 Resp: 435416

Ion Ratio Lower Upper

216 100

214 78.3 60.2 90.4

218 48.4 37.6 56.4



#76

Pentachlorobenzene

Concen: 28.655 ng/ul

RT: 14.770 min Scan# 1988

Delta R.T. -0.010 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17




Tgt Ion:250 Resp: 447246


Ion Ratio Lower Upper

250 100

248 61.8 49.8 74.8

252 63.3 50.4 75.6





#77

Carbazole

Concen: 35.392 ng/ul

RT: 17.602 min Scan# 2470

Delta R.T. -0.010 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17

Instrument :

BNA_G

ClientSampleId :

SLCS588





Tgt Ion:167 Resp: 1650044

Ion Ratio Lower Upper

167 100

166 23.1 18.2 27.4

139 11.8 10.1 15.1



#78

Di-n-butylphthalate

Concen: 35.462 ng/ul

RT: 18.160 min Scan# 2565

Delta R.T. -0.010 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17




Tgt Ion:149 Resp: 1996664


Ion Ratio Lower Upper

149 100

150 9.1 7.2 10.8

104 5.9 5.4 8.0









#80
Fluoranthene
Concen: 33.006 ng/u1
RT: 19.259 min Scan# 21
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:202 Resp: 2159600
Ion Ratio Lower Upper
202 100
101 11.7 8.1 12.1
100 9.2 6.5 9.7



#82
Pyrene
Concen: 32.691 ng/u1
RT: 19.623 min Scan# 2814
Delta R.T. -0.010 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:202 Resp: 2275718
Ion Ratio Lower Upper
202 100
101 12.8 8.7 13.1
100 11.0 7.8 11.6





#83

Butylbenzylphthalate

Concen: 39.030 ng/ul

RT: 20.516 min Scan# 2966

Delta R.T. -0.010 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17

Instrument :

BNA_G

ClientSampleId :

SLCS588





Tgt Ion:149 Resp: 834026

Ion Ratio Lower Upper

149 100

91 82.3 68.7 103.1

206 23.1 17.3 25.9



#84

3,3'-Dichlorobenzidine

Concen: 32.968 ng/ul

RT: 21.345 min Scan# 3107

Delta R.T. -0.016 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17




Tgt Ion:252 Resp: 658366


Ion Ratio Lower Upper

252 100

254 65.9 55.8 83.6

126 9.1 6.4 9.6









#85
Benzo(a)anthracene
Concen: 30.843 ng/ul
RT: 21.427 min Scan# 3121
Delta R.T. -0.016 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:228 Resp: 2138318
Ion Ratio Lower Upper
228 100
229 20.0 15.7 23.5
226 28.2 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 38.394 ng/ul
RT: 21.339 min Scan# 3106
Delta R.T. -0.016 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:149 Resp: 1193467
Ion Ratio Lower Upper
149 100
167 28.7 22.1 33.1
279 4.2 2.9 4.3





#87

Chrysene

Concen: 30.275 ng/ul

RT: 21.486 min Scan# 3132

Delta R.T. -0.016 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17

Instrument :

BNA_G

ClientSampleId :

SLCS588





Tgt Ion:228 Resp: 2000503

Ion Ratio Lower Upper

228 100

226 31.4 23.4 35.0

229 20.4 14.8 22.2



#89

Di-n-octyl phthalate

Concen: 41.326 ng/ul




RT: 22.467 min Scan# 3298


Delta R.T. -0.022 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17

Tgt Ion:149 Resp: 2051257









#90
Benzo(b)fluoranthene
Concen: 32.342 ng/ul
RT: 23.507 min Scan# 3475
Delta R.T. -0.016 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17




Instrument :
BNA_G
ClientSampleId :
SLCS588


Tgt Ion:252 Resp: 2113875
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 9.2 7.0 10.4



#91
Benzo(k)fluoranthene
Concen: 32.020 ng/ul
RT: 23.566 min Scan# 3485
Delta R.T. -0.022 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:252 Resp: 2087593
Ion Ratio Lower Upper
252 100
253 22.4 17.7 26.5
125 9.0 6.5 9.7





#93

Benzo(a)pyrene

Concen: 30.948 ng/ul

RT: 24.318 min Scan# 3

Delta R.T. -0.022 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17

Instrument :

BNA_G

ClientSampleId :

SLCS588





Tgt Ion:252 Resp: 1898615

Ion Ratio Lower Upper

252 100

253 22.1 16.6 24.8

125 10.3 8.0 12.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 32.844 ng/ul

RT: 27.861 min Scan# 4216

Delta R.T. -0.028 min

Lab File: BG063705.D

Acq: 17 Dec 2024 19:17




Tgt Ion:276 Resp: 2460762


Ion Ratio Lower Upper

276 100

138 19.8 13.7 20.5

277 24.2 18.4 27.6









#95
Dibenzo(a,h)anthracene
Concen: 33.573 ng/ul
RT: 27.908 min Scan# 41
Delta R.T. -0.028 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Instrument :
BNA_G
ClientSampleId :
SLCS588

Tgt Ion:278 Resp: 2012687

Ion	Ratio	Lower	Upper
278	100		
139	13.2	9.7	14.5
279	22.6	17.3	25.9



#96
Benzo(g,h,i)perylene
Concen: 32.456 ng/ul
RT: 28.918 min Scan# 4396
Delta R.T. -0.022 min
Lab File: BG063705.D
Acq: 17 Dec 2024 19:17

Tgt Ion:276 Resp: 1929601

Ion	Ratio	Lower	Upper
276	100		
138	19.4	13.8	20.8
277	24.2	17.8	26.8

