

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG121916\
 Data File : BG025350.D
 Acq On : 20 Dec 2016 12:10
 Operator : UM/SJ
 Sample : H5905-11
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DA811

Quant Time: Dec 20 12:51:13 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG113016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 20 03:02:46 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.23	152	69991	20.00	ng/ul	0.00
18) Naphthalene-d8	11.05	136	299327	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.85	164	229197	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.59	188	486564	20.00	ng/ul	0.00
75) Chrysene-d12	21.89	240	570115	20.00	ng/ul	0.00
83) Perylene-d12	25.31	264	593516	20.00	ng/ul	0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.57	96	4965	3.66	ng/uL	0.00
5) Phenol-d5	7.38	99	141134	23.37	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.54	67	92276	24.70	ng/ul	0.00
9) 2-Chlorophenol-d4	7.75	132	101655	24.07	ng/ul	0.00
13) 4-Methylphenol-d8	8.93	113	119754	23.73	ng/ul	0.00
19) Nitrobenzene-d5	9.41	128	53892	25.42	ng/ul	0.00
22) 2-Nitrophenol-d4	10.13	143	66243	25.07	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.68	165	130412	25.66	ng/ul	0.00
29) 4-Chloroaniline-d4	11.20	131	55306	11.07	ng/ul	0.01
43) Dimethylphthalate-d6	14.24	166	428320	27.17	ng/ul	0.00
46) Acenaphthylene-d8	14.55	160	484985	28.08	ng/ul	0.00
51) 4-Nitrophenol-d4	15.09	143	46727	17.25	ng/ul	0.02
57) Fluorene-d10	15.83	176	393074	26.49	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.97	200	64591	21.47	ng/ul	0.00
70) Anthracene-d10	17.69	188	587409	28.60	ng/ul	0.00
76) Pyrene-d10	19.96	212	699283	29.78	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.08	264	717109	29.82	ng/ul	0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Phenol	7.41	94	20214	3.26	ng/ul	93
44) Dimethylphthalate	14.29	163	104700	6.76	ng/ul	98
77) Pyrene	19.99	202	28270	1.03	ng/ul	97
81) Bis(2-ethylhexyl)phthalate	21.72	149	4945010	334.82	ng/ul#	50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG121916\
Data File : BG025350.D
Acq On : 20 Dec 2016 12:10
Operator : UM/SJ
Sample : H5905-11
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DA811

Quant Time: Dec 20 12:51:13 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG113016.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 20 03:02:46 2016
Response via : Initial Calibration

