

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG121916\
 Data File : BG025325.D
 Acq On : 19 Dec 2016 19:47
 Operator : UM/SJ
 Sample : PB95443BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK43

Quant Time: Dec 20 02:46:06 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG113016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 20 02:21:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.22	152	52065	20.00	ng/ul	0.00
18) Naphthalene-d8	11.05	136	217351	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.84	164	181289	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.59	188	429582	20.00	ng/ul	0.00
75) Chrysene-d12	21.88	240	594348	20.00	ng/ul	0.00
83) Perylene-d12	25.30	264	603040	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.57	96	7494	7.43	ng/uL	0.00
5) Phenol-d5	7.38	99	146681	32.66	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.53	67	104360	37.56	ng/ul	0.00
9) 2-Chlorophenol-d4	7.75	132	114950	36.60	ng/ul	0.00
13) 4-Methylphenol-d8	8.93	113	129595	34.52	ng/ul	0.00
19) Nitrobenzene-d5	9.40	128	61442	39.92	ng/ul	0.00
22) 2-Nitrophenol-d4	10.13	143	69422	36.18	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.68	165	128536	34.83	ng/ul	0.00
29) 4-Chloroaniline-d4	11.20	131	165133	45.53	ng/ul	0.00
43) Dimethylphthalate-d6	14.24	166	507888	40.73	ng/ul	0.00
46) Acenaphthylene-d8	14.54	160	529527	38.77	ng/ul	0.00
51) 4-Nitrophenol-d4	15.07	143	59147	27.60	ng/ul	0.00
57) Fluorene-d10	15.83	176	429329	36.58	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.97	200	75059	28.26	ng/ul	0.00
70) Anthracene-d10	17.69	188	719342	39.67	ng/ul	0.00
76) Pyrene-d10	19.97	212	956237	39.06	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.06	264	999233	40.89	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG121916\
Data File : BG025325.D
Acq On : 19 Dec 2016 19:47
Operator : UM/SJ
Sample : PB95443BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK43

Quant Time: Dec 20 02:46:06 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG113016.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 20 02:21:52 2016
Response via : Initial Calibration

