

Data Path : Z:\HPCHEM1\BNA G\DATA\BG122015\
 Data File : BG020338.D
 Acq On : 20 Dec 2015 12:16
 Operator : UM/SJ
 Sample : G4803-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C04B8

Quant Time: Dec 21 00:39:22 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG121415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 21 00:36:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.08	152	19341	20.00	ng/ul	0.00
18) Naphthalene-d8	10.91	136	83364	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.72	164	57313	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.47	188	132908	20.00	ng/ul	0.00
78) Chrysene-d12	21.74	240	162188	20.00	ng/ul	0.00
86) Perylene-d12	25.01	264	155169	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.46	96	597	1.70	ng/uL	0.00
5) Phenol-d5	7.24	99	14784	8.67	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.41	67	34600	34.00	ng/ul	0.00
9) 2-Chlorophenol-d4	7.61	132	38770	31.25	ng/ul	0.00
13) 4-Methylphenol-d8	8.79	113	28331	19.42	ng/ul	0.00
19) Nitrobenzene-d5	9.25	128	23950	36.86	ng/ul	0.00
22) 2-Nitrophenol-d4	9.98	143	28863	39.91	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.52	165	50908	34.76	ng/ul	0.00
29) 4-Chloroaniline-d4	11.04	131	4162	2.53	ng/ul	0.00
44) Dimethylphthalate-d6	14.12	166	172059	37.11	ng/ul	0.00
47) Acenaphthylene-d8	14.42	160	201294	37.32	ng/ul	0.00
52) 4-Nitrophenol-d4	14.91	143	6180	7.43	ng/ul	0.00
58) Fluorene-d10	15.71	176	154067	36.64	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.82	200	28514	36.19	ng/ul	0.00
71) Anthracene-d10	17.56	188	234021	38.21	ng/ul	0.00
79) Pyrene-d10	19.85	212	277431	36.70	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.79	264	302485	39.08	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG122015\
Data File : BG020338.D
Acq On : 20 Dec 2015 12:16
Operator : UM/SJ
Sample : G4803-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C04B8

Quant Time: Dec 21 00:39:22 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG121415.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 21 00:36:38 2015
Response via : Initial Calibration

