

Data Path : Z:\HPCHEM1\BNA G\DATA\BG122015\
 Data File : BG020339.D
 Acq On : 20 Dec 2015 12:55
 Operator : UM/SJ
 Sample : G4803-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C04C2

Quant Time: Dec 21 00:39:32 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG121415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 21 00:36:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.09	152	17316	20.00	ng/ul	0.00
18) Naphthalene-d8	10.91	136	76454	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.72	164	52555	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.46	188	128303	20.00	ng/ul	0.00
78) Chrysene-d12	21.74	240	158163	20.00	ng/ul	0.00
86) Perylene-d12	25.01	264	151017	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	7.24	99	1728	1.13	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.41	67	2220	2.44	ng/ul	0.00
9) 2-Chlorophenol-d4	7.62	132	3042	2.74	ng/ul	0.00
13) 4-Methylphenol-d8	8.80	113	2486	1.90	ng/ul	0.00
19) Nitrobenzene-d5	9.25	128	1243	2.09	ng/ul	0.00
22) 2-Nitrophenol-d4	9.98	143	1567	2.36	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.53	165	3636	2.71	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul	
44) Dimethylphthalate-d6	14.12	166	9597	2.26	ng/ul	0.00
47) Acenaphthylene-d8	14.42	160	10801	2.18	ng/ul	0.00
52) 4-Nitrophenol-d4	14.92	143	464	0.61	ng/ul	0.01
58) Fluorene-d10	15.71	176	8545	2.22	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.83	200	949	1.25	ng/ul	0.00
71) Anthracene-d10	17.56	188	14385	2.43	ng/ul	0.00
79) Pyrene-d10	19.85	212	16961	2.30	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.78	264	15817	2.10	ng/ul	-0.01

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG122015\
Data File : BG020339.D
Acq On : 20 Dec 2015 12:55
Operator : UM/SJ
Sample : G4803-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C04C2

Quant Time: Dec 21 00:39:32 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG121415.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 21 00:36:38 2015
Response via : Initial Calibration

