


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122021\
Data File : BG051674.D
Acq On : 20 Dec 2021 9:10
Operator : CG/JU
Sample : DFTPP444
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP444

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
Title : SVOA CALIBRATION
Last Update : Thu Dec 23 13:37:07 2021



AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2478

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	43.8	28608	PASS
68	69	0.00	2	0.8	239	PASS
69	198	0.00	100	47.1	30755	PASS
70	69	0.00	2	0.7	200	PASS
127	198	10	80	48.5	31661	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	65331	PASS
199	198	5	9	6.8	4457	PASS
275	198	10	60	31.7	20726	PASS
365	198	1	100	4.1	2678	PASS
441	198	0.01	100	11.2	7317	PASS
442	198	50	100	76.7	50096	PASS
443	442	15	24	18.9	9489	PASS

Scan 2478 (18.044 min): BG051674.D\data.ms
DFTPP444

m/z	Abundance
39.95	305.0
44.05	448.0
134.85	183.0
166.95	190.0
207.05	692.0
208.05	185.0

Instrument :
BNA_G
ClientSampleId :
DFTPP444

m/z	Abundance
37.25	195.0
38.05	336.0
39.15	1749.0
40.15	222.0
44.25	317.0
49.05	198.0
50.15	5696.0
51.15	23256.0
52.15	1204.0
55.05	169.0
56.05	772.0
57.15	1604.0
61.15	269.0
62.05	556.0
63.05	803.0
65.25	503.0
69.05	23136.0
70.05	212.0
73.15	341.0
74.05	2326.0
75.15	3752.0
76.15	1281.0
77.15	23720.0
78.15	2082.0
79.15	2074.0
80.05	1451.0
81.15	2060.0
82.05	458.0
83.15	306.0
85.15	402.0
86.15	607.0
87.15	205.0
91.05	592.0
92.05	774.0
93.05	2832.0
96.05	272.0
98.15	2480.0
99.15	1797.0
99.95	226.0
101.05	978.0
103.15	378.0
104.05	613.0
104.95	750.0
105.95	240.0
107.15	7872.0
108.15	1411.0
110.05	11527.0
111.15	1661.0
112.15	477.0
115.95	401.0
116.25	426.0
117.15	6601.0
117.95	387.0
121.95	531.0
123.05	801.0
124.05	570.0
124.75	236.0
125.15	262.0
127.05	23280.0
128.05	1804.0
129.05	10006.0
130.15	1611.0
133.95	191.0
135.15	860.0
136.15	324.0
137.05	510.0
139.95	152.0
141.05	1072.0
142.05	714.0
143.05	290.0
146.05	368.0
147.15	875.0
148.05	1635.0
149.15	358.0
150.95	224.0
152.95	496.0
154.05	204.0
155.15	1032.0
156.15	1218.0
157.15	405.0
158.15	371.0
159.05	205.0
159.85	354.0
161.05	710.0
162.15	233.0
165.05	613.0
165.95	393.0
167.05	2597.0
168.15	1277.0
169.05	260.0
169.95	250.0

Instrument :
BNA_G
ClientSampleId :
DFTPP444

171.15	157.0
171.95	297.0
172.95	321.0
173.95	377.0
175.05	925.0
176.05	378.0
176.95	332.0
178.95	1589.0
180.05	1151.0
181.05	697.0
184.05	280.0
185.05	863.0
186.05	6907.0
187.15	2144.0
188.15	221.0
189.05	504.0
190.95	365.0
192.15	505.0
193.05	667.0
196.05	1148.0
198.05	42888.0
198.95	2767.0
200.05	404.0
201.75	257.0
203.05	454.0
203.95	1627.0
205.05	3594.0
206.05	12350.0
207.05	2258.0
208.05	649.0
208.95	312.0
210.25	224.0
211.05	572.0
216.95	3280.0
217.95	635.0
221.15	1878.0
221.95	333.0
223.05	555.0
224.05	6696.0
224.95	2015.0
225.85	210.0
227.05	3564.0
228.05	660.0
229.05	545.0
230.95	178.0
233.75	151.0
235.05	433.0
235.95	166.0
237.05	270.0
238.75	157.0
240.25	155.0
241.05	217.0
242.15	400.0
243.05	641.0
244.05	5549.0
244.95	748.0
245.95	1009.0
247.05	184.0
249.05	246.0
255.05	29656.0
256.05	4292.0
256.95	457.0
258.05	1443.0
259.05	270.0
265.05	743.0
271.05	172.0
273.05	1015.0
273.95	2033.0
275.05	13411.0
276.05	1860.0
277.05	1158.0
277.85	204.0
284.85	171.0
293.05	240.0
296.05	4023.0
296.95	303.0
303.05	419.0
314.95	400.0
316.15	339.0
321.15	152.0
323.05	1198.0
323.95	217.0
326.95	406.0
334.05	488.0
341.05	173.0
345.95	304.0
352.05	316.0
352.95	215.0
353.95	354.0
364.95	1513.0
365.85	311.0
372.05	674.0
373.15	185.0
401.75	248.0
402.95	449.0

Instrument :
BNA_G
ClientSampleId :
DFTPP444

403.95	181.0
421.05	200.0
422.05	316.0
423.05	2165.0
423.95	282.0
440.95	3291.0
442.05	26728.0
442.95	4606.0
444.05	643.0

Instrument :
BNA_G
ClientSampleId :
DFTPP444

m/z	Abundance
38.05	556.0
39.15	2635.0
40.05	302.0
44.05	281.0
50.15	9221.0
51.15	33272.0
52.15	2020.0
55.25	171.0
56.15	1162.0
57.15	2293.0
61.05	350.0
62.15	491.0
63.15	1537.0
65.15	781.0
68.05	717.0
69.05	36264.0
70.05	238.0
73.15	432.0
74.05	3546.0
75.15	6029.0
76.15	1857.0
77.15	37920.0
78.15	2671.0
79.05	2353.0
80.05	2237.0
81.05	2687.0
82.05	575.0
83.05	611.0
85.05	491.0
86.15	1028.0
87.15	379.0
91.25	728.0
92.05	733.0
93.05	5756.0
94.15	248.0
95.05	193.0
95.95	319.0
98.15	3179.0
99.05	2496.0
99.95	248.0
101.05	1657.0
103.05	541.0
104.05	774.0
105.15	1103.0
107.05	11952.0
108.05	1951.0
110.05	20368.0
111.05	2833.0
112.05	417.0
112.85	163.0
116.05	870.0
117.05	10549.0
118.15	850.0
120.05	237.0
122.15	904.0
123.05	1266.0
124.05	490.0
125.05	540.0
127.05	37664.0
128.15	3338.0
129.05	16992.0
130.05	1342.0
131.05	351.0
133.95	462.0
135.05	1623.0
136.05	540.0
137.25	516.0
138.05	160.0
139.05	233.0
139.85	150.0
140.25	250.0
141.05	2523.0
142.05	989.0
143.05	341.0
146.05	497.0
146.95	1104.0
148.05	2346.0
149.05	622.0
151.15	464.0
152.05	173.0
153.05	752.0
153.95	554.0
155.05	1315.0
156.15	1778.0
157.05	362.0
158.05	558.0
159.05	265.0
160.15	814.0
160.95	1031.0
162.15	352.0
164.95	553.0

Instrument :

BNA_G

ClientSampleId :

DFTPP444

166.15	700.0
167.05	3909.0
168.05	1368.0
169.05	357.0
172.05	333.0
173.05	677.0
174.15	630.0
175.05	1706.0
176.15	406.0
177.05	719.0
178.05	388.0
179.05	3268.0
180.05	2069.0
181.05	1338.0
181.95	262.0
184.15	230.0
185.05	1532.0
186.05	12996.0
187.05	3066.0
188.15	389.0
189.05	673.0
191.05	468.0
191.95	1191.0
193.05	1279.0
194.05	237.0
196.05	1910.0
197.95	77408.0
199.05	5617.0
200.05	325.0
201.35	227.0
202.15	224.0
203.05	631.0
204.05	2938.0
205.05	5006.0
206.05	22872.0
207.05	3344.0
208.05	965.0
209.05	520.0
210.25	537.0
211.05	782.0
212.15	288.0
214.95	299.0
216.05	752.0
217.05	6822.0
218.05	809.0
221.05	2905.0
222.05	784.0
223.05	1489.0
224.05	12237.0
225.05	3205.0
227.05	5940.0
228.05	789.0
229.05	1050.0
230.05	204.0
230.95	584.0
233.95	380.0
234.95	490.0
235.85	162.0
236.25	175.0
237.05	560.0
239.05	449.0
240.05	234.0
241.05	398.0
242.05	738.0
243.05	844.0
244.05	10434.0
245.05	1355.0
246.05	2558.0
247.05	454.0
250.75	181.0
253.05	259.0
253.95	636.0
255.05	49544.0
256.05	8237.0
257.05	681.0
257.95	3090.0
258.95	500.0
263.85	220.0
265.05	1228.0
266.15	175.0
272.05	205.0
272.95	1750.0
274.05	4274.0
275.05	23616.0
276.05	2763.0
276.95	2497.0
277.95	301.0
283.15	197.0
283.95	172.0
284.95	377.0
288.85	285.0
293.15	449.0
296.05	6754.0
297.05	1073.0
303.15	662.0

Instrument :
BNA_G
ClientSampleId :
DFTPP444

303.95	172.0
313.95	441.0
314.95	758.0
316.05	623.0
320.95	261.0
321.95	181.0
322.95	1720.0
324.05	396.0
327.05	286.0
327.85	269.0
332.05	251.0
333.95	1257.0
335.05	373.0
340.85	217.0
345.95	424.0
352.05	552.0
353.05	504.0
354.05	235.0
364.95	2942.0
365.85	421.0
371.95	1110.0
372.95	207.0
382.85	434.0
401.85	269.0
402.85	628.0
403.85	217.0
420.95	338.0
421.95	616.0
422.95	4019.0
423.85	611.0
440.95	8091.0
442.05	53816.0
442.95	10294.0
444.15	811.0

Instrument :
BNA_G
ClientSampleId :
DFTPP444

m/z	Abundance
38.05	309.0
39.15	2598.0
40.05	311.0
41.15	230.0
44.15	395.0
49.25	340.0
50.15	7460.0
51.15	29296.0
52.15	1655.0
56.25	906.0
57.05	2209.0
61.05	344.0
62.25	367.0
63.15	1419.0
64.15	293.0
65.25	394.0
69.05	32864.0
70.05	150.0
73.05	410.0
74.15	4167.0
75.05	5240.0
76.25	1969.0
77.15	34024.0
78.15	2619.0
79.05	2600.0
80.05	2393.0
81.05	3303.0
82.05	760.0
83.15	353.0
85.05	422.0
86.15	799.0
86.95	431.0
88.15	163.0
91.05	728.0
92.05	563.0
93.05	4434.0
94.15	197.0
96.15	165.0
98.05	3535.0
99.05	2438.0
100.05	165.0
101.05	1255.0
103.05	676.0
104.05	1111.0
104.95	514.0
105.95	275.0
107.05	12556.0
108.05	1935.0
110.05	18840.0
111.05	2916.0
112.15	393.0
115.95	563.0
117.05	9445.0
118.15	804.0
122.15	653.0
123.05	1439.0
124.15	558.0
125.05	456.0
127.05	34040.0
128.15	2702.0
129.05	16156.0
130.05	1557.0
131.05	226.0
132.15	168.0
134.05	477.0
135.05	1084.0
135.95	420.0
137.05	769.0
138.05	160.0
139.85	152.0
141.05	2494.0
142.05	682.0
143.15	458.0
146.15	321.0
147.15	1343.0
148.05	2161.0
149.05	594.0
150.05	195.0
151.15	297.0
153.15	888.0
153.95	379.0
155.05	1180.0
156.15	1797.0
157.15	307.0
158.05	381.0
159.05	292.0
160.05	662.0
161.05	1024.0
162.05	167.0
164.05	167.0
165.05	847.0

Instrument :
BNA_G
ClientSampleId :
DFTPP444

166.05	744.0
167.05	3787.0
168.05	1585.0
168.95	380.0
170.05	191.0
170.95	211.0
171.95	449.0
172.95	392.0
174.05	807.0
175.05	1585.0
175.95	364.0
176.95	745.0
178.05	257.0
178.95	2764.0
180.05	2690.0
180.95	1238.0
182.25	179.0
184.25	220.0
185.05	1520.0
186.05	11189.0
187.05	4019.0
188.05	525.0
188.95	697.0
189.95	190.0
191.05	582.0
192.05	922.0
193.05	1355.0
194.05	325.0
194.95	206.0
196.05	2536.0
198.05	75696.0
199.05	4988.0
199.95	445.0
201.55	515.0
203.15	551.0
204.05	3718.0
205.05	5074.0
206.05	22904.0
207.05	3846.0
208.05	814.0
209.15	355.0
209.95	293.0
211.05	1149.0
211.75	329.0
214.95	473.0
216.05	456.0
217.05	6530.0
217.95	854.0
220.15	162.0
221.05	3902.0
221.85	874.0
222.95	1355.0
224.05	12446.0
225.05	3430.0
226.05	345.0
227.05	6155.0
228.05	693.0
229.05	1557.0
229.85	222.0
231.05	362.0
234.15	576.0
234.85	375.0
237.15	453.0
239.05	288.0
240.15	188.0
240.95	307.0
242.15	757.0
243.05	1042.0
244.05	10450.0
245.05	1448.0
246.05	2201.0
247.05	314.0
247.85	220.0
249.15	294.0
250.05	184.0
251.05	160.0
252.95	173.0
255.05	56456.0
256.05	7929.0
257.05	527.0
258.05	3366.0
259.05	359.0
263.95	152.0
265.05	1442.0
272.95	1617.0
273.95	4104.0
275.05	25152.0
275.95	3476.0
277.05	2006.0
278.15	396.0
281.05	168.0
282.95	252.0
285.05	574.0
286.15	232.0
292.95	428.0

Instrument :
BNA_G
ClientSampleId :
DFTPP444

294.05	254.0
296.05	7014.0
297.05	1037.0
303.05	1037.0
304.05	210.0
314.05	371.0
314.95	763.0
315.95	360.0
321.05	428.0
321.95	235.0
323.05	2271.0
323.95	361.0
326.95	535.0
328.05	351.0
331.95	299.0
333.05	179.0
334.05	1358.0
335.05	435.0
341.05	337.0
345.95	776.0
352.05	742.0
352.95	364.0
354.05	1174.0
364.95	3580.0
365.95	299.0
371.95	1221.0
372.95	285.0
383.05	216.0
390.05	370.0
391.05	247.0
401.85	265.0
402.95	512.0
403.85	382.0
421.05	655.0
421.95	553.0
422.95	4217.0
423.95	1012.0
440.95	10569.0
442.05	69744.0
443.05	13568.0
444.05	1062.0

Instrument :
BNA_G
ClientSampleId :
DFTPP444