




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122022\
Data File : BG056024.D
Acq On : 20 Dec 2022 10:33
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 21 03:55:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 21 01:11:10 2022
Response via : Initial Calibration



TIC: BG056024.D\data.ms

(77) Benzidine

19.904min (-0.000) 0.00 ng m

response 197424

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.60	13.93
183.00	11.50	12.29
0.00	0.00	0.00