



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122023\
Data File : BG060103.D
Acq On : 20 Dec 2023 17:04
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121323.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Dec 19 04:37:23 2023






AutoFind: Scans 2498, 2499, 2500; Background Corrected with Scan 2491

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.8	59051	PASS
68	69	0.00	2	0.8	430	PASS
69	198	0.00	100	27.7	56707	PASS
70	69	0.00	2	0.3	166	PASS
127	198	10	80	25.8	52869	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	205013	PASS
199	198	5	9	7.8	16056	PASS
275	198	10	60	41.2	84379	PASS
365	198	1	100	9.4	19202	PASS
441	198	0.01	100	24.9	50976	PASS
442	442	50	100	100.0	315029	PASS
443	442	15	24	19.4	61013	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122023\
Data File : BG060103.D
Acq On : 20 Dec 2023 17:04
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Dec 20 17:51:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 19 04:37:23 2023
Response via : Initial Calibration



TIC: BG060103.D\data.ms

(70) Pentachlorophenol (C)

16.984min (-0.005) 429063.81 ng




response 302587

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	59.90	69.33
264.00	61.00	61.94
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122023\
Data File : BG060103.D
Acq On : 20 Dec 2023 17:04
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Dec 20 17:51:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 19 04:37:23 2023
Response via : Initial Calibration



TIC: BG060103.D\data.ms

(77) Benzidine

19.581min (-0.005) 495881.43 ng

response 1464975

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	14.08
183.00	11.20	11.35
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/20/2023	BNA_G	<u>BG060103.D</u>
Compound Name	Response	Retention Time
DDT	821171	20.809
DDD	21.063	20.362
DDE	954	19.863
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
975.063	822146.063	0.12

Instrument :
BNA_G
ClientSampleId :
DFTPP