

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG122216\
 Data File : BG025391.D
 Acq On : 22 Dec 2016 16:22
 Operator : UM/SJ
 Sample : H6103-06
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017

Quant Time: Dec 23 00:16:30 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 21 18:42:01 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.22	152	51603	20.00	ng	0.00
21) Naphthalene-d8	11.04	136	206127	20.00	ng	0.00
38) Acenaphthene-d10	14.84	164	164637	20.00	ng	0.00
63) Phenanthrene-d10	17.59	188	390376	20.00	ng	0.00
75) Chrysene-d12	21.88	240	571352	20.00	ng	0.00
86) Perylene-d12	25.29	264	589723	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.75	112	477624	168.24	ng	0.00
7) Phenol-d6	7.38	99	649990	162.57	ng	0.00
23) Nitrobenzene-d5	9.40	82	529213	113.42	ng	0.00
41) 2,4,6-Tribromophenol	16.33	330	408464	154.10	ng	0.00
44) 2-Fluorobiphenyl	13.46	172	1127459	110.87	ng	0.00
78) Terphenyl-d14	20.17	244	2212934	102.69	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.59	88	19400	16.02	ng	# 92
3) Pyridine	4.00	79	56517	16.10	ng	94
4) n-Nitrosodimethylamine	3.92	42	33733	16.18	ng	# 97
6) Aniline	7.53	93	57567	10.63	ng	96
8) 2-Chlorophenol	7.78	128	57125	17.84	ng	93
9) Benzaldehyde	7.36	77	52227	17.85	ng	86
10) Phenol	7.41	94	82027	18.51	ng	94
11) bis(2-Chloroethyl)ether	7.63	93	56436	16.52	ng	97
12) 1,3-Dichlorobenzene	8.10	146	64733	16.70	ng	95
13) 1,4-Dichlorobenzene	8.25	146	68102	16.96	ng	94
14) 1,2-Dichlorobenzene	8.57	146	62556	16.32	ng	95
15) Benzyl Alcohol	8.46	79	69049	18.09	ng	95
16) 2,2'-oxybis(1-Chloropropan	8.73	45	108011	17.08	ng	93
17) 2-Methylphenol	8.67	107	53865	18.51	ng	97
18) Hexachloroethane	9.30	117	25965	16.83	ng	85
19) n-Nitroso-di-n-propylamine	9.01	70	54645	16.18	ng	98
20) 3+4-Methylphenols	9.00	107	70229	17.43	ng	89
22) Acetophenone	9.04	105	98230	17.15	ng	# 93
24) Nitrobenzene	9.44	77	91434	17.87	ng	97
25) Isophorone	9.96	82	142253	16.84	ng	97
26) 2-Nitrophenol	10.16	139	35983	18.38	ng	97
27) 2,4-Dimethylphenol	10.20	122	61605	19.14	ng	94
28) bis(2-Chloroethoxy)methane	10.43	93	76362	17.41	ng	94
29) 2,4-Dichlorophenol	10.70	162	62767	18.23	ng	93
30) 1,2,4-Trichlorobenzene	10.90	180	73273	17.61	ng	96
31) Naphthalene	11.09	128	179182	17.41	ng	96
32) Benzoic acid	10.32	122	44683	17.27	ng	92
33) 4-Chloroaniline	11.22	127	38247	8.83	ng	# 85
34) Hexachlorobutadiene	11.35	225	57914	17.05	ng	96
35) Caprolactam	11.98	113	22211	17.17	ng	# 82
36) 4-Chloro-3-methylphenol	12.32	107	76725	18.05	ng	95
37) 2-Methylnaphthalene	12.68	142	137760	18.06	ng	91
39) 1,2,4,5-Tetrachlorobenzene	13.05	216	96383	17.73	ng	96
40) Hexachlorocyclopentadiene	13.00	237	17633	16.45	ng	89


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG122216\
 Data File : BG025391.D
 Acq On : 22 Dec 2016 16:22
 Operator : UM/SJ
 Sample : H6103-06
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017

Quant Time: Dec 23 00:16:30 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 21 18:42:01 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.29	196	61500	17.97	ng	94
43) 2,4,5-Trichlorophenol	13.37	196	65441	18.27	ng #	89
45) 1,1'-Biphenyl	13.67	154	195624	17.67	ng	96
46) 2-Chloronaphthalene	13.73	162	152023	17.57	ng	97
47) 2-Nitroaniline	13.94	65	65229	17.86	ng	95
48) Acenaphthylene	14.57	152	238286	17.77	ng	94
49) Dimethylphthalate	14.28	163	212165	17.68	ng #	95
50) 2,6-Dinitrotoluene	14.42	165	45374	17.30	ng	98
51) Acenaphthene	14.90	154	150775	17.19	ng	97
52) 3-Nitroaniline	14.76	138	34387	13.64	ng	89
53) 2,4-Dinitrophenol	15.00	184	11408	11.29	ng #	90
54) Dibenzofuran	15.24	168	247047	17.82	ng	98
55) 4-Nitrophenol	15.10	139	31629	16.80	ng #	78
56) 2,4-Dinitrotoluene	15.22	165	67933	17.79	ng #	91
57) Fluorene	15.88	166	206582	17.80	ng	95
58) 2,3,4,6-Tetrachlorophenol	15.47	232	71817	18.71	ng	93
59) Diethylphthalate	15.63	149	215297	17.11	ng	99
60) 4-Chlorophenyl-phenylether	15.87	204	115104	17.50	ng	91
61) 4-Nitroaniline	15.93	138	45133	17.79	ng	95
62) Azobenzene	16.16	77	216431	17.83	ng	97
64) 4,6-Dinitro-2-methylphenol	15.98	198	41766	15.45	ng	86
65) n-Nitrosodiphenylamine	16.08	169	184534	17.49	ng	95
66) 4-Bromophenyl-phenylether	16.76	248	81598	16.94	ng	98
67) Hexachlorobenzene	16.88	284	92586	16.75	ng	94
68) Atrazine	17.02	200	63063	13.64	ng	96
69) Pentachlorophenol	17.25	266	39789	13.89	ng	98
70) Phenanthrene	17.72	178	357951	17.79	ng	94
71) Anthracene	17.72	178	357951	17.51	ng	97
72) Carbazole	18.00	167	332379	18.18	ng #	94
73) Di-n-butylphthalate	18.50	149	405979	18.05	ng	99
74) Fluoranthene	19.63	202	494256	18.60	ng	90
76) Benzidine	19.80	184	135052	9.48	ng	97
77) Pyrene	19.99	202	521537	17.47	ng	93
79) Butylbenzylphthalate	20.84	149	207445	17.30	ng	94
80) Benzo(a)anthracene	21.86	228	571506	17.93	ng	91
81) 3,3'-Dichlorobenzidine	21.76	252	137602	11.12	ng	95
82) Chrysene	21.93	228	545756	17.86	ng	95
83) Bis(2-ethylhexyl)phthalate	21.70	149	290851	17.43	ng	98
84) Di-n-octyl phthalate	22.96	149	497056	17.94	ng #	93
85) Indeno(1,2,3-cd)pyrene	29.21	276	670829	17.27	ng #	96
87) Benzo(b)fluoranthene	24.19	252	585352	18.19	ng #	95
88) Benzo(k)fluoranthene	24.19	252	585352	18.84	ng	92
89) Benzo(a)pyrene	25.13	252	556931	17.92	ng #	97
90) Dibenzo(a,h)anthracene	29.26	278	584904	17.76	ng	98
91) Benzo(g,h,i)perylene	30.45	276	556127	16.99	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.22 min Scan# 915

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:152 Resp: 51603

Ion Ratio Lower Upper

152 100

150 140.9 114.0 171.0




115 54.8 48.1 72.1



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 16.02 ng

RT: 3.59 min Scan# 127

Delta R.T. -0.01 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion: 88 Resp: 19400

Ion Ratio Lower Upper

88 100

58 98.3 79.4 119.2



43 58.4 33.8 50.6#




Abundance Ion 88.00 (87.70 to 88.70): BGC

Ion 58.00 (57.70 to 58.70): BGC

Ion 43.00 (42.70 to 43.70): BGC





#3

Pyridine

Concen: 16.10 ng

RT: 4.00 min Scan# 198

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion: 79 Resp: 56517

Ion Ratio Lower Upper




79 100

52 92.5 77.8 116.6

51 60.0 43.2 64.8



Abundance Ion 79.00 (78.70 to 79.70): BGC
Ion 52.00 (51.70 to 52.70): BGC
Ion 51.00 (50.70 to 51.70): BGC



#4

n-Nitrosodimethylamine

Concen: 16.18 ng

RT: 3.92 min Scan# 184

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion: 42 Resp: 33733

Ion Ratio Lower Upper



42 100


74 98.9 76.7 115.1

44 9.4 6.1 9.1#



Abundance Ion 42.00 (41.70 to 42.70): BGC
Ion 74.00 (73.70 to 74.70): BGC
Ion 44.00 (43.70 to 44.70): BGC





#5

2-Fluorophenol

Concen: 168.24 ng

RT: 5.75 min Scan# 496

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion: 112 Resp: 477624

Ion Ratio Lower Upper

112 100

64 73.1 61.8 92.6




63 44.8 37.2 55.8



Abundance Ion 112.00 (111.70 to 112.70): BGC

Ion 64.00 (63.70 to 64.70): BGC

Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 10.63 ng

RT: 7.53 min Scan# 799

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion: 93 Resp: 57567

Ion Ratio Lower Upper

93 100

66 44.5 35.4 53.2



65 21.4 21.2 31.8




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 66.00 (65.70 to 66.70): BGC

Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 162.57 ng

RT: 7.38 min Scan# 772

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion: 99 Resp: 649990

Ion Ratio Lower Upper




99 100

42 26.7 20.5 30.7

71 51.4 37.6 56.4



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#8

2-Chlorophenol

Concen: 17.84 ng

RT: 7.78 min Scan# 841

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion: 128 Resp: 57125

Ion Ratio Lower Upper



128 100

130 36.4 11.4 51.4

64 71.4 46.6 86.6



Abundance Ion 128.00 (127.70 to 128.70): BGC
Ion 130.00 (129.70 to 130.70): BGC
Ion 64.00 (63.70 to 64.70): BGC





#9

Benzaldehyde

Concen: 17.85 ng

RT: 7.36 min Scan# 769

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion: 77 Resp: 52227

Ion Ratio Lower Upper

77 100

105 68.4 60.9 100.9




106 63.4 55.1 95.1



Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC



#10

Phenol

Concen: 18.51 ng

RT: 7.41 min Scan# 777

Delta R.T. 0.01 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion: 94 Resp: 82027

Ion Ratio Lower Upper

94 100

65 35.6 20.6 60.6



66 52.0 35.5 75.5




Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC









#11
bis(2-Chloroethyl)ether
Concen: 16.52 ng
RT: 7.63 min Scan# 815
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion: 93 Resp: 56436
Ion Ratio Lower Upper
93 100
63 100.9 78.5 118.5
95 32.7 9.8 49.8



#12
1,3-Dichlorobenzene
Concen: 16.70 ng
RT: 8.10 min Scan# 895
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion: 146 Resp: 64733
Ion Ratio Lower Upper
146 100
148 63.1 49.2 73.8
75 47.7 33.4 50.2





#13

1,4-Dichlorobenzene

Concen: 16.96 ng

RT: 8.25 min Scan# 921

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:146 Resp: 68102

Ion Ratio Lower Upper

146 100

148 70.7 52.2 78.2




111 49.5 37.0 55.6



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 16.32 ng

RT: 8.57 min Scan# 976

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:146 Resp: 62556

Ion Ratio Lower Upper

146 100

148 63.5 54.6 81.8



111 46.7 39.7 59.5




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E









#15
Benzyl Alcohol
Concen: 18.09 ng
RT: 8.46 min Scan# 957
Delta R.T. 0.01 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion: 79 Resp: 69049
Ion Ratio Lower Upper
79 100
108 57.7 45.5 68.3
77 61.2 54.3 81.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 17.08 ng
RT: 8.73 min Scan# 1003
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion: 45 Resp: 108011
Ion Ratio Lower Upper
45 100
77 10.7 0.0 30.6
79 14.4 0.0 28.5





#17

2-Methylphenol

Concen: 18.51 ng

RT: 8.67 min Scan# 992

Delta R.T. 0.01 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017

Tgt Ion:107 Resp: 53865


Ion Ratio Lower Upper

107 100

108 105.8 85.6 128.4

77 60.0 51.4 77.2

79 64.1 49.2 73.8






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BG

Ion 79.00 (78.70 to 79.70): BG



#18

Hexachloroethane

Concen: 16.83 ng

RT: 9.30 min Scan# 1099

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:117 Resp: 25965

Ion Ratio Lower Upper

117 100

119 103.0 69.8 104.6



201 134.9 95.4 143.0




Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E






#19
n-Nitroso-di-n-propylamine
Concen: 16.18 ng
RT: 9.01 min Scan# 1051
Delta R.T. -0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017

Tgt Ion:	70	Resp:	54645
Ion	Ratio	Lower	Upper
70	100		
42	69.6	56.1	84.1
101	9.8	7.5	11.3
130	15.9	14.6	21.8






Abundance

Ion 70.00 (69.70 to 70.70): BGC

Ion 42.00 (41.70 to 42.70): BGC


Ion 101.00 (100.70 to 101.70): BGC

Ion 130.00 (129.70 to 130.70): BGC



#20
3+4-Methylphenols
Concen: 17.43 ng
RT: 9.00 min Scan# 1049
Delta R.T. 0.01 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:	107	Resp:	70229
Ion	Ratio	Lower	Upper
107	100		
108	105.4	70.8	110.8
77	45.2	25.8	65.8
79	46.7	20.1	60.1





Abundance


Ion 107.00 (106.70 to 107.70): BGC

Ion 108.00 (107.70 to 108.70): BGC

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 11.04 min Scan# 1396
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017

Tgt Ion:	136	Resp:	206127
Ion	Ratio	Lower	Upper
136	100		
137	12.5	8.9	13.3
54	12.6	9.3	13.9
68	7.0	5.1	7.7




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC






#22
Acetophenone
Concen: 17.15 ng
RT: 9.04 min Scan# 1056
Delta R.T. -0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:	105	Resp:	98230
Ion	Ratio	Lower	Upper
105	100		
71	0.0	0.9	1.3#
51	37.4	34.0	51.0
120	24.0	17.3	25.9



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BGC
Ion 51.00 (50.70 to 51.70): BGC
Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 113.42 ng
 RT: 9.40 min Scan# 1116
 Delta R.T. -0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22




Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017


Tgt Ion	Ratio	Lower	Upper
82	100		
128	33.0	26.4	39.6
54	58.9	49.4	74.2



#24
 Nitrobenzene
 Concen: 17.87 ng
 RT: 9.44 min Scan# 1124
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22

Tgt Ion	Ratio	Lower	Upper
77	100		
123	33.3	26.1	39.1
65	12.6	12.4	18.6





#25

Isophorone

Concen: 16.84 ng

RT: 9.96 min Scan# 1211

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion: 82 Resp: 142253

Ion Ratio Lower Upper




82 100

95 8.2 7.5 11.3

138 14.5 12.3 18.5



Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 95.00 (94.70 to 95.70): BGC
Ion 138.00 (137.70 to 138.70): BGC



#26

2-Nitrophenol

Concen: 18.38 ng

RT: 10.16 min Scan# 1246

Delta R.T. 0.01 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion: 139 Resp: 35983

Ion Ratio Lower Upper



139 100


109 48.8 38.6 58.0

65 75.8 57.1 85.7



Abundance Ion 139.00 (138.70 to 139.70): BGC
Ion 109.00 (108.70 to 109.70): BGC
Ion 65.00 (64.70 to 65.70): BGC









#27
 2,4-Dimethylphenol
 Concen: 19.14 ng
 RT: 10.20 min Scan# 1253
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22




Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017


Tgt Ion: 122 Resp: 61605
 Ion Ratio Lower Upper
 122 100
 107 123.7 104.2 156.4
 121 52.0 46.0 69.0



#28
 bis(2-Chloroethoxy)methane
 Concen: 17.41 ng
 RT: 10.43 min Scan# 1292
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22

Tgt Ion: 93 Resp: 76362
 Ion Ratio Lower Upper
 93 100
 95 28.0 25.7 38.5
 123 9.8 8.3 12.5





#29

2,4-Dichlorophenol

Concen: 18.23 ng

RT: 10.70 min Scan# 1338

Delta R.T. 0.01 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:162 Resp: 62767

Ion Ratio Lower Upper

162 100

164 64.2 42.4 82.4




98 31.3 20.3 60.3



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): B



#30

1,2,4-Trichlorobenzene

Concen: 17.61 ng

RT: 10.90 min Scan# 1372

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:180 Resp: 73273

Ion Ratio Lower Upper

180 100

182 91.8 77.0 115.4



145 27.0 23.4 35.0




Abundance Ion 180.00 (179.70 to 180.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 145.00 (144.70 to 145.70): E









#31
Naphthalene
Concen: 17.41 ng
RT: 11.09 min Scan# 1405
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017

Tgt Ion:128 Resp: 179182
Ion Ratio Lower Upper
128 100
129 11.4 9.3 13.9
127 11.8 11.4 17.0




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#32
Benzoic acid
Concen: 17.27 ng
RT: 10.32 min Scan# 1273
Delta R.T. -0.04 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:122 Resp: 44683
Ion Ratio Lower Upper
122 100
105 123.4 120.1 160.1
77 123.1 104.6 144.6



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BG









#33
4-Chloroaniline
Concen: 8.83 ng
RT: 11.22 min Scan# 1426
Delta R.T. 0.01 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion:	127	Resp:	38247
Ion Ratio		Lower	Upper
127	100		
129	45.1	23.6	35.4#
65	41.0	37.7	56.5
92	26.3	17.6	26.4



#34
Hexachlorobutadiene
Concen: 17.05 ng
RT: 11.35 min Scan# 1449
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:	225	Resp:	57914
Ion Ratio		Lower	Upper
225	100		
223	62.0	50.7	76.1
227	55.8	49.0	73.6





#35

Caprolactam

Concen: 17.17 ng

RT: 11.98 min Scan# 1556

Delta R.T. -0.01 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:113 Resp: 22211

Ion Ratio Lower Upper

113 100

55 191.5 198.6 238.6#




56 133.3 140.4 180.4#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BG

Ion 56.00 (55.70 to 56.70): BG



#36

4-Chloro-3-methylphenol

Concen: 18.05 ng

RT: 12.32 min Scan# 1614

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:107 Resp: 76725

Ion Ratio Lower Upper

107 100

144 22.8 17.4 26.0



142 72.8 54.1 81.1




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 18.06 ng
RT: 12.68 min Scan# 1675
Delta R.T. -0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion:142 Resp: 137760
Ion Ratio Lower Upper
142 100
141 97.1 70.4 105.6
115 49.7 35.5 53.3



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.84 min Scan# 2042
Delta R.T. -0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:164 Resp: 164637
Ion Ratio Lower Upper
164 100
162 96.7 85.1 127.7
160 44.5 34.7 52.1









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 17.73 ng
 RT: 13.05 min Scan# 1737
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22




Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017


Tgt Ion:	216	Resp:	96383
Ion	Ratio	Lower	Upper
216	100		
214	81.9	64.4	96.6
179	18.5	17.4	26.0
108	15.8	15.6	23.4



#40
 Hexachlorocyclopentadiene
 Concen: 16.45 ng
 RT: 13.00 min Scan# 1730
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22

Tgt Ion:	237	Resp:	17633
Ion	Ratio	Lower	Upper
237	100		
235	67.4	39.4	79.4
272	16.5	0.0	32.0









#41
 2,4,6-Tribromophenol
 Concen: 154.10 ng
 RT: 16.33 min Scan# 2296
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22




Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017


Tgt Ion:330 Resp: 408464
 Ion Ratio Lower Upper
 330 100
 332 97.8 77.3 115.9
 141 32.8 32.1 48.1



#42
 2,4,6-Trichlorophenol
 Concen: 17.97 ng
 RT: 13.29 min Scan# 1778
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22

Tgt Ion:196 Resp: 61500
 Ion Ratio Lower Upper
 196 100
 198 94.2 81.4 122.2
 200 35.6 27.0 40.6









#43
 2,4,5-Trichlorophenol
 Concen: 18.27 ng
 RT: 13.37 min Scan# 1793
 Delta R.T. 0.01 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22




Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017


Tgt Ion:196 Resp: 65441
 Ion Ratio Lower Upper
 196 100
 198 91.3 79.9 119.9
 97 43.9 45.4 68.0#
 132 22.1 20.7 31.1



#44
 2-Fluorobiphenyl
 Concen: 110.87 ng
 RT: 13.46 min Scan# 1808
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22

Tgt Ion:172 Resp: 1127459
 Ion Ratio Lower Upper
 172 100
 171 38.4 32.2 48.2
 170 25.1 21.8 32.6









#45
 1,1'-Biphenyl
 Concen: 17.67 ng
 RT: 13.67 min Scan# 1844
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22




Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017


Tgt Ion	Ratio	Lower	Upper
154	100		
153	40.3	23.0	63.0
76	12.1	0.0	33.5



#46
 2-Chloronaphthalene
 Concen: 17.57 ng
 RT: 13.73 min Scan# 1853
 Delta R.T. -0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22

Tgt Ion	Ratio	Lower	Upper
162	100		
127	38.4	32.2	48.4
164	32.3	27.3	40.9









#47
2-Nitroaniline
Concen: 17.86 ng
RT: 13.94 min Scan# 1890
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion: 65 Resp: 65229
Ion Ratio Lower Upper
65 100
92 55.2 49.0 73.4
138 70.7 58.6 87.8



#48
Acenaphthylene
Concen: 17.77 ng
RT: 14.57 min Scan# 1996
Delta R.T. -0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion: 152 Resp: 238286
Ion Ratio Lower Upper
152 100
151 19.2 18.2 27.2
153 12.4 11.3 16.9





#49

Dimethylphthalate

Concen: 17.68 ng

RT: 14.28 min Scan# 1947

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:163 Resp: 212165

Ion Ratio Lower Upper

163 100

194 5.1 3.8 5.6




164 9.3 9.3 13.9#



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 17.30 ng

RT: 14.42 min Scan# 1971

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:165 Resp: 45374

Ion Ratio Lower Upper

165 100

63 80.6 63.9 95.9



89 71.0 58.6 88.0




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BGC

Ion 89.00 (88.70 to 89.70): BGC





#51

Acenaphthene

Concen: 17.19 ng

RT: 14.90 min Scan# 2053

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:154 Resp: 150775

Ion Ratio Lower Upper

154 100

153 113.0 87.8 131.6




152 51.9 42.2 63.4



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 13.64 ng

RT: 14.76 min Scan# 2029

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:138 Resp: 34387

Ion Ratio Lower Upper

138 100

108 14.7 10.9 16.3



92 130.0 115.3 172.9




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BGC









#53
2,4-Dinitrophenol
Concen: 11.29 ng
RT: 15.00 min Scan# 2070
Delta R.T. 0.01 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017

Tgt Ion:184 Resp: 11408
Ion Ratio Lower Upper
184 100
63 50.7 52.7 79.1#
154 69.7 57.3 85.9




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BG
Ion 154.00 (153.70 to 154.70): E






#54
Dibenzofuran
Concen: 17.82 ng
RT: 15.24 min Scan# 2110
Delta R.T. -0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:168 Resp: 247047
Ion Ratio Lower Upper
168 100
139 44.6 35.3 52.9
169 16.1 11.5 17.3



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 16.80 ng
RT: 15.10 min Scan# 2086
Delta R.T. 0.01 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion:139 Resp: 31629
Ion Ratio Lower Upper
139 100
109 107.3 101.2 141.2
65 117.8 135.3 175.3#



#56
2,4-Dinitrotoluene
Concen: 17.79 ng
RT: 15.22 min Scan# 2107
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:165 Resp: 67933
Ion Ratio Lower Upper
165 100
63 60.6 44.0 66.0
89 74.8 67.3 100.9
182 3.2 3.8 5.6#





#57

Fluorene

Concen: 17.80 ng

RT: 15.88 min Scan# 2220

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:166 Resp: 206582

Ion Ratio Lower Upper

166 100

165 93.4 78.6 117.8




167 11.8 11.6 17.4



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 18.71 ng

RT: 15.47 min Scan# 2149

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Tgt Ion:232 Resp: 71817


Ion Ratio Lower Upper

232 100

131 39.5 34.9 52.3

130 3.0 2.3 3.5

166 25.9 24.2 36.4






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E





#59

Diethylphthalate

Concen: 17.11 ng

RT: 15.63 min Scan# 2177

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:149 Resp: 215297

Ion Ratio Lower Upper

149 100

177 25.7 20.1 30.1




150 13.2 10.2 15.2



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 17.50 ng

RT: 15.87 min Scan# 2217

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:204 Resp: 115104

Ion Ratio Lower Upper

204 100

206 35.6 30.1 45.1



141 54.1 50.6 76.0




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 17.79 ng
RT: 15.93 min Scan# 2228
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion:138 Resp: 45133
Ion Ratio Lower Upper
138 100
92 69.4 44.2 84.2
108 119.6 104.8 144.8



#62
Azobenzene
Concen: 17.83 ng
RT: 16.16 min Scan# 2267
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion: 77 Resp: 216431
Ion Ratio Lower Upper
77 100
182 24.1 4.7 44.7
105 16.2 0.0 36.3
51 33.1 16.4 56.4









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.59 min Scan# 2510
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion:188 Resp: 390376
Ion Ratio Lower Upper
188 100
94 5.1 5.8 8.8#
80 7.4 7.5 11.3#



#64
4,6-Dinitro-2-methylphenol
Concen: 15.45 ng
RT: 15.98 min Scan# 2237
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:198 Resp: 41766
Ion Ratio Lower Upper
198 100
51 54.9 22.6 62.6
105 38.3 14.4 54.4









#65
n-Nitrosodiphenylamine
Concen: 17.49 ng
RT: 16.08 min Scan# 2254
Delta R.T. -0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion:169 Resp: 184534
Ion Ratio Lower Upper
169 100
168 65.4 55.7 83.5
167 34.4 29.8 44.6



#66
4-Bromophenyl-phenylether
Concen: 16.94 ng
RT: 16.76 min Scan# 2369
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:248 Resp: 81598
Ion Ratio Lower Upper
248 100
250 95.4 75.0 112.6
141 71.8 55.2 82.8





#67

Hexachlorobenzene

Concen: 16.75 ng

RT: 16.88 min Scan# 2390

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:284 Resp: 92586

Ion Ratio Lower Upper

284 100

142 35.1 29.9 44.9




249 32.0 29.2 43.8



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 13.64 ng

RT: 17.02 min Scan# 2413

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:200 Resp: 63063

Ion Ratio Lower Upper

200 100

173 20.8 3.0 43.0



215 46.0 28.4 68.4




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E





#69

Pentachlorophenol

Concen: 13.89 ng

RT: 17.25 min Scan# 2452

Delta R.T. 0.01 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:266 Resp: 39789

Ion Ratio Lower Upper

266 100

268 62.6 48.6 72.8




264 60.8 50.1 75.1



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#70

Phenanthrene

Concen: 17.79 ng

RT: 17.72 min Scan# 2533

Delta R.T. 0.10 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:178 Resp: 357951

Ion Ratio Lower Upper

178 100

176 18.7 17.7 26.5



179 16.3 15.0 22.6




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 17.51 ng

RT: 17.72 min Scan# 2533

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:178 Resp: 357951

Ion Ratio Lower Upper

178 100

176 18.7 16.4 24.6




179 16.3 13.8 20.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 18.18 ng

RT: 18.00 min Scan# 2580

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:167 Resp: 332379

Ion Ratio Lower Upper

167 100

166 23.4 20.2 30.2



139 12.4 12.8 19.2#




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E









#73
Di-n-butylphthalate
Concen: 18.05 ng
RT: 18.50 min Scan# 2666
Delta R.T. -0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion:149 Resp: 405979
Ion Ratio Lower Upper
149 100
150 10.5 9.1 13.7
104 8.4 6.7 10.1



#74
Fluoranthene
Concen: 18.60 ng
RT: 19.63 min Scan# 2857
Delta R.T. 0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:202 Resp: 494256
Ion Ratio Lower Upper
202 100
101 5.7 0.0 30.1
203 16.9 1.3 41.3





#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.88 min Scan# 3240

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017





Tgt Ion:240 Resp: 571352

Ion Ratio Lower Upper

240 100

120 4.8 5.7 8.5#

236 25.0 22.2 33.4



#76

Benzidine

Concen: 9.48 ng

RT: 19.80 min Scan# 2887

Delta R.T. 0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22




Tgt Ion:184 Resp: 135052


Ion Ratio Lower Upper

184 100

185 14.8 13.0 19.6

183 12.6 11.0 16.6









#77
 Pyrene
 Concen: 17.47 ng
 RT: 19.99 min Scan# 2919
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22




Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017


Tgt Ion:	202	Resp:	521537
Ion	Ratio	Lower	Upper
202	100		
200	20.8	19.6	29.4
203	17.2	15.8	23.8



#78
 Terphenyl-d14
 Concen: 102.69 ng
 RT: 20.17 min Scan# 2949
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22

Tgt Ion:	244	Resp:	2212934
Ion	Ratio	Lower	Upper
244	100		
212	10.1	10.0	15.0
122	8.3	8.6	12.8









#79
Butylbenzylphthalate
Concen: 17.30 ng
RT: 20.84 min Scan# 3063
Delta R.T. -0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22




Instrument :
BNA_G
ClientSampleId :
LOQ-WATER-06-QT1-2017


Tgt Ion:149 Resp: 207445
Ion Ratio Lower Upper
149 100
91 86.5 63.5 95.3
206 25.9 21.0 31.6



#80
Benzo(a)anthracene
Concen: 17.93 ng
RT: 21.86 min Scan# 3237
Delta R.T. -0.00 min
Lab File: BG025391.D
Acq: 22 Dec 2016 16:22

Tgt Ion:228 Resp: 571506
Ion Ratio Lower Upper
228 100
226 25.9 24.9 37.3
229 19.2 18.2 27.2









#81
 3,3'-Dichlorobenzidine
 Concen: 11.12 ng
 RT: 21.76 min Scan# 3221
 Delta R.T. 0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22




Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017


Tgt Ion: 252 Resp: 137602
 Ion Ratio Lower Upper
 252 100
 254 62.0 53.1 79.7
 126 7.6 7.0 10.4



#82
 Chrysene
 Concen: 17.86 ng
 RT: 21.93 min Scan# 3249
 Delta R.T. -0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22

Tgt Ion: 228 Resp: 545756
 Ion Ratio Lower Upper
 228 100
 226 31.4 27.0 40.4
 229 19.7 17.8 26.6









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 17.43 ng
 RT: 21.70 min Scan# 3210
 Delta R.T. -0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22




Instrument :
 BNA_G
 ClientSampleId :
 LOQ-WATER-06-QT1-2017


Tgt Ion	Ratio	Lower	Upper
149	100		
167	28.7	23.7	35.5
279	4.9	4.6	6.8



#84
 Di-n-octyl phthalate
 Concen: 17.94 ng
 RT: 22.96 min Scan# 3424
 Delta R.T. -0.00 min
 Lab File: BG025391.D
 Acq: 22 Dec 2016 16:22

Tgt Ion	Ratio	Lower	Upper
149	100		
167	1.9	1.4	2.2
43	11.7	11.8	17.6#





#85

Indeno(1,2,3-cd)pyrene

Concen: 17.27 ng

RT: 29.21 min Scan# 4489

Delta R.T. -0.02 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017


Tgt Ion:276 Resp: 670829

Ion Ratio Lower Upper

276 100

138 0.9 4.7 7.1#




277 26.3 20.6 31.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#86

Perylene-d12

Concen: 20.00 ng

RT: 25.29 min Scan# 3821

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22


Tgt Ion:264 Resp: 589723

Ion Ratio Lower Upper

264 100

260 21.6 21.8 32.8#



265 22.7 18.2 27.4




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E





#87

Benzo(b)fluoranthene

Concen: 18.19 ng

RT: 24.19 min Scan# 3634

Delta R.T. -0.00 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22





Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017

Tgt Ion:	252	Resp:	585352
Ion Ratio	Lower	Upper	
252	100		
253	20.7	18.7	28.1
125	5.2	5.3	7.9#



#88

Benzo(k)fluoranthene

Concen: 18.84 ng




RT: 24.19 min Scan# 3634


Delta R.T. -0.08 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Tgt Ion:	252	Resp:	585352
Ion Ratio	Lower	Upper	
252	100		
253	20.7	20.2	30.2
125	5.2	5.1	7.7





#89

Benzo(a)pyrene

Concen: 17.92 ng

RT: 25.13 min Scan# 3793

Delta R.T. -0.01 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22





Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017

Tgt Ion:	252	Resp:	556931
Ion Ratio	Lower	Upper	
252	100		
253	23.2	19.2	28.8
125	4.6	5.4	8.0#



#90

Dibenzo(a,h)anthracene

Concen: 17.76 ng




RT: 29.26 min Scan# 4497


Delta R.T. -0.02 min

Lab File: BG025391.D

Acq: 22 Dec 2016 16:22

Tgt Ion:	278	Resp:	584904
Ion Ratio	Lower	Upper	
278	100		
139	8.9	7.7	11.5
279	23.1	19.1	28.7





#91

Benzo(g,h,i)perylene

Concen: 16.99 ng

RT: 30.45 min Scan# 4699

Delta R.T. -0.02 min

Lab File: BG025391.D


Acq: 22 Dec 2016 16:22

Instrument :

BNA_G

ClientSampleId :

LOQ-WATER-06-QT1-2017




Tgt Ion: 276 Resp: 556127

Ion Ratio Lower Upper

276 100

277 27.7 18.6 27.8

138 8.8 8.1 12.1



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

