

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG122217\  
 Data File : BG031717.D  
 Acq On : 22 Dec 2017 18:47  
 Operator : SJ/JU  
 Sample : I7002-07  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 EPE-106-3DS(60-62)

Quant Time: Dec 23 01:37:03 2017  
 Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG122117.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Dec 22 11:08:55 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)  |
|-----------------------------|-------|------|----------|--------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.83  | 152  | 46277    | 20.00  | ng    | -0.01     |
| 21) Naphthalene-d8          | 11.71 | 136  | 183269   | 20.00  | ng    | -0.01     |
| 38) Acenaphthene-d10        | 15.46 | 164  | 129629   | 20.00  | ng    | -0.01     |
| 63) Phenanthrene-d10        | 18.19 | 188  | 340307   | 20.00  | ng    | -0.01     |
| 75) Chrysene-d12            | 22.55 | 240  | 399830   | 20.00  | ng    | -0.02     |
| 86) Perylene-d12            | 26.34 | 264  | 425332   | 20.00  | ng    | -0.02     |
| System Monitoring Compounds |       |      |          |        |       |           |
| 5) 2-Fluorophenol           | 6.24  | 112  | 351216   | 138.23 | ng    | 0.00      |
| 7) Phenol-d6                | 7.92  | 99   | 449417   | 136.24 | ng    | 0.00      |
| 23) Nitrobenzene-d5         | 10.03 | 82   | 231441   | 95.36  | ng    | -0.01     |
| 41) 2,4,6-Tribromophenol    | 16.94 | 330  | 308461   | 155.95 | ng    | 0.00      |
| 44) 2-Fluorobiphenyl        | 14.09 | 172  | 874893   | 84.71  | ng    | -0.01     |
| 78) Terphenyl-d14           | 20.73 | 244  | 1528621  | 87.35  | ng    | -0.01     |
| Target Compounds            |       |      |          |        |       |           |
| 10) Phenol                  | 7.95  | 94   | 14506    | 4.356  | ng    | Qvalue 90 |
| 49) Dimethylphthalate       | 14.89 | 163  | 124223   | 10.878 | ng    | # 98      |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG122217\  
Data File : BG031717.D  
Acq On : 22 Dec 2017 18:47  
Operator : SJ/JU  
Sample : I7002-07  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
EPE-106-3DS(60-62)

Quant Time: Dec 23 01:37:03 2017  
Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG122117.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Dec 22 11:08:55 2017  
Response via : Initial Calibration

