


Data Path : Z:\HPCHEM1\BNA\_G\Data\BG122217\  
 Data File : BG031713.D  
 Acq On : 22 Dec 2017 16:16  
 Operator : SJ/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG122117.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Dec 22 11:08:55 2017



AutoFind: Scans 2601, 2602, 2603; Background Corrected with Scan 2594

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 16.9      | 9350    | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 18.8      | 10375   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 31.6      | 17480   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 55248   | PASS             |
| 199         | 198          | 5            | 9            | 6.3       | 3462    | PASS             |
| 275         | 198          | 10           | 60           | 31.4      | 17374   | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 1858    | PASS             |
| 441         | 442          | 0.01         | 24           | 16.4      | 7912    | PASS             |
| 442         | 198          | 50           | 100          | 87.2      | 48160   | PASS             |
| 443         | 442          | 15           | 24           | 18.8      | 9069    | PASS             |