



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122324\
Data File : BG063778.D
Acq On : 23 Dec 2024 10:21
Operator : RC/JU
Sample : DFTPP448
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP448

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Dec 24 00:44:39 2024



AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2468

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	38.6	109005	PASS
68	69	0.00	2	1.1	1367	PASS
69	198	0.00	100	42.3	119633	PASS
70	69	0.00	2	0.5	634	PASS
127	198	10	80	36.9	104299	PASS
197	198	0.00	2	0.2	476	PASS
198	198	100	100	100.0	282539	PASS
199	198	5	9	7.1	20006	PASS
275	198	10	60	28.8	81371	PASS
365	198	1	100	3.8	10678	PASS
441	198	0.01	100	12.7	35885	PASS
442	198	50	100	75.4	213077	PASS
443	442	15	24	19.8	42235	PASS

m/z	Abundance
35.20	389.0
35.90	185.0
40.10	19440.0
41.00	692.0
42.00	333.0
43.20	1092.0
44.10	5095.0
45.00	161.0
46.20	235.0
46.90	395.0
48.00	260.0
49.10	1206.0
50.00	215.0
51.10	350.0
55.10	986.0
56.20	302.0
57.10	726.0
61.00	173.0
62.90	191.0
67.00	314.0
69.10	650.0
71.20	351.0
72.90	286.0
80.90	185.0
83.00	693.0
84.00	1595.0
85.10	667.0
86.00	613.0
94.90	196.0
96.00	251.0
97.10	1047.0
98.00	386.0
110.10	215.0
111.10	245.0
117.10	251.0
118.90	209.0
124.10	158.0
129.00	162.0
129.40	188.0
184.10	270.0
188.20	172.0
189.10	196.0
190.80	236.0
194.40	216.0
207.00	154.0
209.10	252.0
212.20	170.0
218.00	161.0
232.20	248.0
266.10	292.0
298.40	160.0
328.60	161.0
332.10	161.0
415.70	242.0
450.00	184.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

m/z	Abundance
36.10	382.0
37.10	451.0
38.10	1990.0
39.10	9540.0
40.10	21624.0
41.10	849.0
42.10	186.0
43.10	928.0
44.00	5351.0
45.10	249.0
47.10	363.0
47.90	160.0
49.00	1885.0
50.10	32216.0
51.10	129960.0
52.10	6707.0
53.10	286.0
54.00	233.0
55.00	694.0
56.10	3508.0
57.10	8777.0
58.10	572.0
60.00	166.0
61.10	1223.0
62.10	1504.0
63.10	5017.0
64.10	1325.0
65.10	2374.0
67.10	625.0
68.00	2816.0
69.00	142720.0
70.00	703.0
70.90	651.0
73.20	1562.0
74.10	13431.0
75.10	22240.0
76.10	7864.0
77.10	168448.0
78.10	11875.0
79.00	10742.0
80.00	8180.0
81.10	9696.0
82.10	2543.0
83.00	2171.0
83.90	1400.0
85.00	1990.0
86.00	5064.0
87.00	1391.0
88.10	699.0
88.80	447.0
91.00	2514.0
92.00	3244.0
93.00	19144.0
94.10	743.0
95.10	449.0
96.20	841.0
98.10	14476.0
99.10	10135.0
100.10	937.0
101.10	5571.0
102.00	383.0
103.10	2077.0
104.00	4734.0
105.00	3658.0
106.20	2085.0
107.00	44456.0
108.00	7281.0
109.00	1951.0
110.00	73424.0
111.00	10085.0
112.00	1918.0
113.00	361.0
114.10	168.0
114.90	290.0
116.00	2240.0
117.00	33704.0
118.00	1851.0
119.20	428.0
119.90	541.0
121.30	347.0
122.00	2321.0
123.10	3656.0
124.00	1559.0
125.00	1988.0
127.10	118968.0
128.10	9179.0
129.00	45480.0
130.00	3586.0
131.00	753.0
132.00	401.0
133.00	491.0
134.00	1388.0

Instrument :

BNA_G

ClientSampleId :

DFTPP448

135.00	3034.0
136.20	1443.0
137.10	2325.0
138.70	217.0
139.00	246.0
141.10	5358.0
142.00	2149.0
142.90	1469.0
144.00	202.0
144.50	320.0
145.20	372.0
146.00	1021.0
147.10	3589.0
148.00	7440.0
149.10	1931.0
150.10	395.0
151.20	1251.0
151.90	543.0
153.10	1862.0
154.10	1828.0
155.00	4434.0
156.10	5316.0
157.20	1527.0
157.90	957.0
159.00	1006.0
160.10	2790.0
161.00	3041.0
162.10	1651.0
162.70	292.0
163.90	425.0
165.10	2822.0
166.10	2196.0
167.10	14155.0
168.10	5714.0
169.00	939.0
170.10	748.0
170.90	670.0
171.50	629.0
172.10	1381.0
173.00	1643.0
174.00	2703.0
175.10	6233.0
176.00	1666.0
177.00	3951.0
177.90	1065.0
179.00	12371.0
180.10	8027.0
181.10	3851.0
182.20	1166.0
184.10	1576.0
185.10	6709.0
186.10	42088.0
187.00	13028.0
187.90	1484.0
189.00	2994.0
190.10	690.0
191.00	1844.0
192.00	3840.0
193.10	4547.0
194.10	1227.0
194.90	837.0
196.10	9330.0
198.00	295296.0
199.00	22672.0
200.00	1949.0
201.30	1546.0
203.00	2822.0
204.00	12914.0
205.00	22328.0
206.00	80304.0
207.10	12096.0
208.00	3311.0
209.10	1418.0
209.90	1444.0
211.00	3212.0
212.00	684.0
212.90	196.0
213.90	287.0
215.10	1416.0
216.00	2140.0
217.00	27408.0
218.00	3475.0
219.20	529.0
220.00	548.0
221.00	11315.0
222.00	3317.0
223.00	6912.0
224.10	48944.0
225.00	14255.0
226.00	1060.0
227.00	25624.0
227.90	2606.0
229.00	4595.0
230.20	512.0
230.90	2237.0
232.20	462.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

232.80	226.0
233.20	219.0
234.10	1422.0
234.90	1538.0
235.80	880.0
236.20	700.0
237.00	1810.0
238.00	237.0
239.00	1349.0
240.00	799.0
241.00	1513.0
242.00	2460.0
243.00	2719.0
244.00	39040.0
245.10	4967.0
246.00	10003.0
247.00	1800.0
247.90	195.0
248.90	1062.0
250.20	456.0
251.20	559.0
251.90	932.0
253.10	1224.0
255.00	205504.0
256.00	30192.0
257.00	2623.0
258.00	12380.0
259.10	2202.0
259.90	746.0
261.20	539.0
263.90	377.0
265.00	5067.0
266.00	926.0
267.10	153.0
268.80	150.0
270.00	352.0
271.00	2939.0
271.90	652.0
273.00	6003.0
274.00	15362.0
275.00	76448.0
276.10	9906.0
277.00	7866.0
278.00	1173.0
278.90	468.0
281.20	214.0
281.80	313.0
283.00	518.0
284.00	362.0
285.00	1002.0
286.10	458.0
289.10	695.0
290.90	962.0
291.80	759.0
292.90	2720.0
294.10	676.0
295.00	1082.0
296.00	26448.0
297.00	2590.0
301.10	266.0
301.70	331.0
303.00	2415.0
304.10	757.0
308.10	281.0
308.60	177.0
310.00	286.0
313.10	235.0
314.00	999.0
315.00	2422.0
316.10	1489.0
317.00	262.0
321.20	641.0
322.00	183.0
323.00	7049.0
323.90	833.0
324.90	432.0
326.00	268.0
327.10	1534.0
328.00	623.0
332.00	677.0
333.00	670.0
334.00	4593.0
334.90	812.0
339.00	272.0
341.20	712.0
342.00	269.0
344.00	156.0
346.00	878.0
346.90	269.0
348.20	161.0
351.20	294.0
352.00	2427.0
353.10	793.0
354.00	1970.0
354.80	263.0
355.30	173.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

359.10	204.0
361.30	166.0
363.90	390.0
365.00	8854.0
366.00	1142.0
370.10	176.0
371.00	503.0
372.00	3669.0
373.00	1111.0
383.00	680.0
384.00	164.0
390.10	417.0
391.10	297.0
392.30	195.0
397.10	257.0
401.10	239.0
401.90	1612.0
402.90	1767.0
403.80	479.0
405.00	281.0
418.90	270.0
421.00	1985.0
422.00	1063.0
423.00	10622.0
424.00	1416.0
425.00	444.0
433.90	171.0
437.70	278.0
438.90	371.0
441.00	30976.0
442.00	179136.0
443.00	36328.0
444.00	3065.0
444.90	172.0
446.30	235.0
447.00	192.0
457.90	300.0
461.40	265.0
479.20	189.0
533.70	178.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

m/z	Abundance
35.50	158.0
37.20	312.0
38.00	1248.0
39.10	9213.0
40.10	19200.0
41.10	851.0
42.10	341.0
43.10	794.0
44.10	6103.0
45.10	446.0
46.20	150.0
46.90	794.0
49.10	1695.0
50.10	33552.0
51.10	126688.0
52.10	6055.0
53.00	436.0
53.90	217.0
55.00	1027.0
56.00	3666.0
57.10	7451.0
58.00	279.0
59.10	351.0
60.00	207.0
61.10	1469.0
62.10	1383.0
63.10	4957.0
64.10	568.0
65.20	3815.0
66.00	463.0
67.00	488.0
69.00	141824.0
70.20	1181.0
71.10	855.0
73.00	1474.0
74.10	14544.0
75.00	22976.0
76.10	7207.0
77.10	162752.0
78.10	11083.0
79.00	9319.0
80.00	6659.0
81.00	10137.0
82.10	2006.0
83.00	3989.0
84.00	1914.0
85.10	2446.0
86.00	3285.0
87.00	1599.0
88.10	1150.0
89.00	221.0
91.00	2051.0
92.00	2660.0
93.10	16688.0
94.00	1310.0
95.10	204.0
96.00	621.0
96.90	542.0
98.10	11357.0
99.10	9110.0
100.00	1127.0
101.10	6273.0
102.00	265.0
103.10	1172.0
104.00	4263.0
105.00	4022.0
107.00	41168.0
108.00	5835.0
110.00	75152.0
111.10	10947.0
112.10	1725.0
113.00	635.0
115.00	197.0
116.00	2094.0
117.00	34016.0
118.00	2125.0
119.20	357.0
120.00	865.0
121.20	384.0
122.00	1729.0
123.00	3141.0
124.10	1392.0
125.10	1838.0
125.90	772.0
127.10	117464.0
128.10	8709.0
129.00	49312.0
130.10	3515.0
131.00	1169.0
132.20	348.0
133.20	527.0
134.00	1144.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

135.00	4847.0
136.00	1583.0
137.10	3355.0
138.00	328.0
139.00	252.0
140.00	755.0
141.00	7383.0
142.00	1717.0
143.00	1592.0
143.80	661.0
145.00	631.0
146.10	1068.0
147.10	2388.0
148.00	9567.0
149.00	1851.0
150.00	445.0
151.30	952.0
153.00	2191.0
154.10	1430.0
155.00	3854.0
156.10	5133.0
157.10	1096.0
158.00	1825.0
159.10	1269.0
160.00	2176.0
161.00	4199.0
162.20	1080.0
162.90	359.0
163.60	266.0
164.10	637.0
165.00	2497.0
166.00	2974.0
167.10	13864.0
168.00	7351.0
169.00	2025.0
170.10	795.0
170.90	419.0
172.00	1784.0
173.00	2479.0
174.00	3440.0
175.10	6874.0
176.00	1965.0
176.90	4073.0
178.10	1631.0
179.00	12853.0
180.00	9439.0
181.10	4752.0
181.90	486.0
182.20	445.0
183.10	454.0
183.90	1522.0
185.10	5595.0
186.00	50208.0
187.10	15012.0
187.90	1117.0
189.00	3232.0
190.10	640.0
191.10	1795.0
192.00	4184.0
193.10	5209.0
194.00	1740.0
195.10	1312.0
196.00	8648.0
198.00	334848.0
199.00	21672.0
200.00	1624.0
200.70	494.0
201.50	2493.0
202.30	430.0
203.10	2244.0
204.10	15557.0
205.00	27376.0
206.00	102224.0
207.10	13411.0
208.00	3664.0
209.00	1751.0
210.10	1976.0
211.00	4997.0
212.00	814.0
212.90	572.0
213.60	248.0
215.00	1837.0
216.00	3071.0
217.00	31088.0
218.00	3629.0
219.00	187.0
219.90	793.0
221.00	14478.0
222.00	4219.0
223.00	6526.0
224.00	56344.0
225.00	15991.0
227.00	25320.0
228.00	3392.0
229.00	4630.0
230.20	574.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

230.90	1846.0
231.90	475.0
233.00	509.0
234.10	1274.0
235.00	2201.0
236.00	1518.0
237.00	2031.0
238.20	254.0
238.90	1456.0
240.00	977.0
240.90	1172.0
242.00	4617.0
243.00	4077.0
244.00	45840.0
245.10	5064.0
246.00	11071.0
246.90	1617.0
248.00	855.0
249.00	1173.0
250.90	619.0
252.00	1256.0
253.00	906.0
255.00	242432.0
256.00	34440.0
257.00	2561.0
258.00	14717.0
259.00	2498.0
259.80	506.0
260.90	666.0
262.30	348.0
263.90	365.0
265.00	6892.0
266.10	697.0
268.10	207.0
269.90	582.0
270.90	3549.0
272.20	1542.0
273.00	8213.0
274.00	17856.0
275.00	99888.0
276.10	11270.0
277.00	9825.0
278.00	1405.0
279.00	503.0
281.20	427.0
282.00	348.0
283.00	939.0
283.90	751.0
285.00	2170.0
286.00	698.0
287.00	216.0
288.90	885.0
289.90	295.0
291.00	371.0
292.20	722.0
293.10	1875.0
294.20	1245.0
295.10	984.0
296.00	34824.0
297.00	4641.0
298.20	242.0
298.90	317.0
301.00	295.0
302.00	746.0
303.00	3745.0
304.20	1073.0
305.20	244.0
308.10	563.0
308.80	454.0
310.00	228.0
310.70	174.0
310.90	186.0
311.90	302.0
312.90	459.0
314.10	993.0
315.10	4163.0
315.90	1900.0
317.00	247.0
320.80	1109.0
323.00	9049.0
324.10	1477.0
326.90	1904.0
328.10	880.0
329.10	208.0
330.90	296.0
332.00	800.0
333.00	1058.0
334.00	5399.0
335.00	1637.0
336.20	498.0
339.60	248.0
340.20	202.0
341.00	1353.0
342.00	641.0
346.00	2163.0
346.70	245.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

349.00	245.0
352.10	2644.0
352.90	1774.0
354.00	1957.0
355.00	1048.0
358.30	179.0
359.10	340.0
360.00	295.0
364.90	12364.0
365.80	1649.0
366.90	353.0
370.10	203.0
371.00	826.0
372.10	4449.0
373.10	1655.0
377.00	352.0
381.00	362.0
382.90	721.0
386.50	206.0
389.70	217.0
390.90	1092.0
391.90	602.0
402.10	1564.0
403.00	1506.0
404.10	1382.0
405.20	232.0
414.90	261.0
416.60	174.0
417.30	356.0
418.30	253.0
419.20	294.0
421.00	1692.0
422.00	1847.0
423.00	13302.0
424.00	2490.0
425.10	231.0
436.30	220.0
437.70	565.0
438.90	641.0
441.00	42112.0
442.00	249152.0
443.00	49288.0
444.10	3334.0
444.90	343.0
447.40	228.0
456.90	184.0
483.00	240.0
544.70	171.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

m/z	Abundance
35.10	372.0
37.20	452.0
38.10	677.0
39.10	4215.0
40.10	21128.0
41.10	570.0
42.20	231.0
43.10	1438.0
44.10	6181.0
45.10	164.0
46.90	329.0
48.10	269.0
49.10	1478.0
50.10	18128.0
51.10	71416.0
52.10	2915.0
54.00	213.0
55.00	1281.0
56.10	2436.0
57.10	5410.0
58.10	662.0
61.00	823.0
62.00	1437.0
63.10	3108.0
63.90	410.0
65.10	1136.0
66.20	198.0
67.30	437.0
68.00	1286.0
69.00	76304.0
70.10	722.0
71.20	369.0
73.00	919.0
74.10	6628.0
75.00	12683.0
76.10	4284.0
77.10	96944.0
78.10	4708.0
79.00	5357.0
80.10	4913.0
81.00	6145.0
82.00	1829.0
83.00	2215.0
84.00	1463.0
85.00	1679.0
86.00	3212.0
87.00	1235.0
88.10	495.0
91.10	1342.0
92.00	2035.0
93.00	10487.0
94.10	985.0
95.00	771.0
96.10	642.0
97.10	1023.0
98.10	7985.0
99.00	6071.0
100.10	405.0
101.00	3731.0
103.10	1201.0
104.00	2359.0
105.10	2518.0
106.00	708.0
107.00	26560.0
108.10	3300.0
109.10	840.0
110.00	38976.0
111.10	6087.0
112.20	726.0
112.90	277.0
113.10	304.0
113.50	374.0
116.00	2053.0
117.10	18736.0
118.10	1184.0
119.10	181.0
120.00	388.0
121.20	482.0
122.00	1738.0
123.00	1788.0
124.20	964.0
125.10	1086.0
127.00	76464.0
128.00	4833.0
129.00	29648.0
130.00	2634.0
131.00	330.0
132.10	256.0
132.80	522.0
134.00	852.0
135.10	2469.0
135.90	1121.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

137.10	1433.0
138.00	691.0
139.10	181.0
140.00	861.0
141.10	4431.0
142.00	1679.0
142.60	995.0
142.90	1003.0
145.00	489.0
146.00	1336.0
146.90	1817.0
148.00	6521.0
149.00	1092.0
149.90	198.0
151.20	583.0
151.70	536.0
153.00	1309.0
154.10	1553.0
155.10	2758.0
156.00	3754.0
157.00	878.0
158.00	478.0
159.00	782.0
160.10	1650.0
161.00	2507.0
162.00	657.0
163.10	245.0
163.90	244.0
165.00	1611.0
166.20	1714.0
167.10	9179.0
168.10	4301.0
169.00	866.0
169.90	303.0
170.80	375.0
172.00	1005.0
173.10	2410.0
174.10	2162.0
175.10	4905.0
176.10	996.0
176.90	2437.0
178.10	1087.0
179.00	9880.0
180.00	5990.0
181.00	1975.0
182.10	517.0
182.80	480.0
184.00	867.0
185.10	5094.0
186.00	31680.0
187.10	8428.0
188.00	494.0
189.00	2117.0
190.00	429.0
191.10	1311.0
192.00	3179.0
193.00	3161.0
194.00	889.0
195.00	544.0
196.00	6503.0
196.90	1429.0
198.00	217472.0
199.00	15675.0
200.10	1409.0
201.30	1391.0
202.10	334.0
203.00	2257.0
204.10	9474.0
205.00	15990.0
206.00	69904.0
207.10	7285.0
208.00	2994.0
208.90	883.0
209.10	894.0
210.00	948.0
211.10	2561.0
212.10	744.0
213.10	462.0
213.70	300.0
214.90	576.0
216.10	1344.0
217.00	19296.0
218.10	2219.0
218.80	394.0
219.70	208.0
221.00	8655.0
221.90	2345.0
223.00	3984.0
224.00	42856.0
225.00	10386.0
226.10	1393.0
227.00	18384.0
228.00	2823.0
229.00	3655.0
230.10	488.0
231.10	1875.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

232.20	376.0
233.00	343.0
234.00	1376.0
235.00	1221.0
236.10	919.0
237.00	1386.0
237.90	290.0
238.90	840.0
240.00	874.0
241.10	845.0
242.00	1811.0
243.10	2245.0
244.10	30472.0
245.00	5302.0
246.00	7692.0
246.90	1056.0
248.00	736.0
248.90	1606.0
249.80	456.0
250.50	451.0
251.00	339.0
252.20	488.0
253.00	771.0
255.00	170112.0
256.00	25168.0
257.00	2102.0
258.00	10018.0
259.00	2065.0
260.10	337.0
261.00	348.0
263.10	336.0
264.10	456.0
265.00	3667.0
266.00	405.0
267.00	518.0
268.00	291.0
271.00	2464.0
272.20	1224.0
273.00	5239.0
274.00	13764.0
275.00	67776.0
276.00	6876.0
277.00	6704.0
278.00	964.0
279.20	270.0
281.40	213.0
282.10	397.0
283.00	1189.0
284.00	698.0
284.90	1324.0
285.70	172.0
288.20	218.0
291.00	782.0
292.10	654.0
293.00	2035.0
294.10	426.0
296.00	23120.0
297.00	2728.0
298.00	315.0
300.90	306.0
301.90	487.0
303.10	2498.0
304.10	680.0
307.90	358.0
308.80	163.0
310.00	171.0
311.80	197.0
313.00	295.0
313.90	1202.0
314.90	1922.0
316.00	1552.0
317.00	218.0
319.20	180.0
320.00	265.0
321.00	1215.0
322.10	443.0
323.00	6939.0
324.10	1080.0
325.80	230.0
327.00	1163.0
328.20	684.0
328.80	201.0
329.10	232.0
332.10	709.0
333.20	887.0
334.00	5725.0
335.00	1126.0
338.30	260.0
341.00	1079.0
342.10	582.0
345.90	1697.0
347.30	203.0
349.70	178.0
351.90	2211.0
353.00	1085.0
354.10	2102.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

355.10	628.0
359.30	287.0
365.00	10815.0
366.20	750.0
367.00	327.0
368.20	162.0
370.00	241.0
370.80	857.0
372.00	3740.0
372.80	1157.0
382.90	966.0
384.90	273.0
390.20	306.0
390.90	278.0
392.00	356.0
400.90	328.0
402.00	1341.0
403.10	2577.0
403.90	664.0
412.70	179.0
414.50	206.0
419.30	163.0
421.00	2198.0
422.00	1884.0
423.00	11367.0
424.00	2692.0
424.90	172.0
432.10	258.0
438.10	579.0
438.70	253.0
439.20	413.0
441.00	34568.0
442.00	210944.0
443.00	41088.0
444.00	4704.0
444.90	504.0
457.70	291.0
475.00	177.0
518.80	209.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448