

Data Path : Z:\HPCHEM1\BNA G\Data\BG122515\
 Data File : BG020501.D
 Acq On : 25 Dec 2015 12:31
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02026

Quant Time: Dec 26 00:22:38 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG122415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 25 03:27:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.416	0.492	-18.3	105	0.00
3 S	1,4-Dioxane-d8	0.363	0.371	-2.2	92	0.00
4	Benzaldehyde	0.973	1.145	-17.7	97	0.00
5 S	Phenol-d5	1.605	1.759	-9.6	100	0.00
6	Phenol	1.671	1.854	-11.0	103	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.951	1.032	-8.5	98	0.00
8	Bis(2-Chloroethyl)ether	1.218	1.354	-11.2	104	0.00
9 S	2-Chlorophenol-d4	1.231	1.365	-10.9	97	0.00
10	2-Chlorophenol	1.285	1.420	-10.5	97	0.00
11	2-Methylphenol	1.297	1.426	-9.9	100	0.00
12	2,2'-oxybis(1-Chloropropane	1.579	1.718	-8.8	96	0.00
13 S	4-Methylphenol-d8	1.350	1.524	-12.9	102	0.00
14	Acetophenone	2.164	2.454	-13.4	102	0.00
15 P	N-Nitroso-di-n-propylamine	1.094	1.274	-16.5	103	0.00
16	4-Methylphenol	1.407	1.622	-15.3	104	0.00
17	Hexachloroethane	0.543	0.583	-7.4	96	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
19 S	Nitrobenzene-d5	0.157	0.168	-7.0	97	0.00
20	Nitrobenzene	0.391	0.425	-8.7	98	0.00
21	Isophorone	0.738	0.829	-12.3	101	0.00
22 S	2-Nitrophenol-d4	0.178	0.203	-14.0	102	0.00
23 C	2-Nitrophenol	0.188	0.207	-10.1	102	0.00
24	2,4-Dimethylphenol	0.400	0.447	-11.7	99	0.00
25	Bis(2-Chloroethoxy)methane	0.410	0.460	-12.2	102	0.00
26 S	2,4-Dichlorophenol-d3	0.345	0.394	-14.2	101	0.00
27 C	2,4-Dichlorophenol	0.354	0.398	-12.4	101	0.00
28	Naphthalene	1.063	1.165	-9.6	100	0.00
29 S	4-Chloroaniline-d4	0.394	0.446	-13.2	99	0.00
30	4-Chloroaniline	0.397	0.463	-16.6	102	0.00
31 C	Hexachlorobutadiene	0.296	0.316	-6.8	99	0.00
32	Caprolactam	0.114	0.126	-10.5	103	0.00
33 C	4-Chloro-3-methylphenol	0.371	0.435	-17.3	102	0.00
34	2-Methylnaphthalene	0.787	0.881	-11.9	102	0.00
35	1-Methylnaphthalene	0.747	0.849	-13.7	102	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
37	1,2,4,5-Tetrachlorobenzene	0.810	0.853	-5.3	101	0.00
38	Hexachlorocyclopentadiene	0.441	0.402	8.8	96	0.00
39 C	2,4,6-Trichlorophenol	0.481	0.523	-8.7	101	0.00
40	2,4,5-Trichlorophenol	0.526	0.592	-12.5	107	0.00
41	1,1'-Biphenyl	1.544	1.666	-7.9	103	0.00
42	2-Chloronaphthalene	1.245	1.337	-7.4	104	0.00
43	2-Nitroaniline	0.349	0.393	-12.6	105	0.00
44 S	Dimethylphthalate-d6	1.648	1.780	-8.0	101	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG122515\
 Data File : BG020501.D
 Acq On : 25 Dec 2015 12:31
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02026

Quant Time: Dec 26 00:22:38 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG122415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 25 03:27:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.676	1.810	-8.0	102	0.00
46	2,6-Dinitrotoluene	0.341	0.374	-9.7	101	0.00
47 S	Acenaphthylene-d8	1.910	2.068	-8.3	102	0.00
48	Acenaphthylene	2.008	2.193	-9.2	102	0.00
49	3-Nitroaniline	0.315	0.365	-15.9	107	0.00
50 C	Acenaphthene	1.357	1.475	-8.7	104	0.00
51	2,4-Dinitrophenol	0.218	0.185	15.1	91	0.00
52 S	4-Nitrophenol-d4	0.275	0.274	0.4	94	0.00
53	4-Nitrophenol	0.249	0.253	-1.6	98	0.00
54	Dibenzofuran	2.019	2.220	-10.0	103	0.00
55	2,4-Dinitrotoluene	0.510	0.569	-11.6	103	0.00
56	2,3,4,6-Tetrachlorophenol	0.518	0.555	-7.1	101	0.00
57	Diethylphthalate	1.720	1.877	-9.1	102	0.00
58 S	Fluorene-d10	1.491	1.599	-7.2	101	0.00
59	Fluorene	1.636	1.796	-9.8	102	0.00
60	4-Chlorophenyl-phenylether	0.947	1.021	-7.8	102	0.00
61	4-Nitroaniline	0.353	0.391	-10.8	105	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.133	0.126	5.3	96	-0.01
64	4,6-Dinitro-2-methylphenol	0.141	0.136	3.5	95	0.00
65	N-Nitrosodiphenylamine	0.559	0.606	-8.4	102	0.00
66	4-Bromophenyl-phenylether	0.249	0.265	-6.4	102	0.00
67	Hexachlorobenzene	0.278	0.288	-3.6	98	0.00
68	Atrazine	0.248	0.266	-7.3	102	0.00
69 C	Pentachlorophenol	0.150	0.150	0.0	102	0.00
70	Phenanthrene	1.113	1.191	-7.0	101	0.00
71 S	Anthracene-d10	0.937	1.016	-8.4	104	0.00
72	Anthracene	1.138	1.219	-7.1	101	0.00
73	1,2,3,4-Tetrachlorobenzene	0.334	0.349	-4.5	101	0.00
74	Pentachlorobenzene	0.310	0.331	-6.8	103	0.00
75	Carbazole	0.959	1.043	-8.8	103	0.00
76	Di-n-butylphthalate	1.143	1.203	-5.2	100	0.00
77 C	Fluoranthene	1.399	1.535	-9.7	100	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	98	-0.01
79 S	Pyrene-d10	0.866	0.984	-13.6	103	0.00
80	Pyrene	1.131	1.268	-12.1	101	0.00
81	Butylbenzylphthalate	0.402	0.440	-9.5	98	0.00
82	3,3'-Dichlorobenzidine	0.392	0.401	-2.3	90	0.00
83	Benzo(a)anthracene	1.175	1.294	-10.1	99	0.00
84	Bis(2-ethylhexyl)phthalate	0.588	0.641	-9.0	98	0.00
85	Chrysene	1.119	1.227	-9.7	98	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	-0.01
87	Di-n-octyl phthalate	1.035	1.137	-9.9	99	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG122515\
Data File : BG020501.D
Acq On : 25 Dec 2015 12:31
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02026

Quant Time: Dec 26 00:22:38 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG122415.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 25 03:27:34 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.200	1.303	-8.6	98	0.00
89	Benzo(k)fluoranthene	1.160	1.248	-7.6	100	-0.01
90 S	Benzo(a)pyrene-d12	1.011	1.103	-9.1	100	-0.01
91 C	Benzo(a)pyrene	1.154	1.259	-9.1	100	-0.01
92	Indeno(1,2,3-cd)pyrene	1.382	1.484	-7.4	100	-0.03
93	Dibenzo(a,h)anthracene	1.147	1.233	-7.5	100	-0.02
94	Benzo(a,h,i)perylene	1.135	1.199	-5.6	98	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0