


Data Path : Z:\HPCHEM1\BNA G\Data\BG122515\
 Data File : BG020500.D
 Acq On : 25 Dec 2015 11:52
 Operator : UM/SJ
 Sample : DFTPP15
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP15

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG122415.M
 Title : SVOA CALIBRATION
 Last Update : Mon Dec 28 18:11:16 2015



AutoFind: Scans 2558, 2559, 2560; Background Corrected with Scan 2553

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.1	8750	PASS
68	69	0.00	2	1.6	157	PASS
69	198	0.00	100	31.9	9911	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.1	12479	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	31114	PASS
199	198	5	9	7.2	2227	PASS
275	198	10	60	29.7	9240	PASS
365	198	1	100	3.5	1086	PASS
441	442	0.01	100	16.7	4307	PASS
442	198	50	100	83.1	25850	PASS
443	442	15	24	19.6	5072	PASS

Scan 2553 (17.837 min): BG020500.D
DFTPP15

m/z	Abundance
40.00	160.0
44.00	216.0
207.00	166.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

m/z	Abundance
39.00	511.0
39.90	234.0
44.00	173.0
50.00	2203.0
51.00	7346.0
52.00	524.0
56.00	257.0
56.90	571.0
63.00	409.0
65.00	154.0
68.00	218.0
68.90	8253.0
74.00	780.0
75.00	1361.0
76.00	375.0
77.00	9044.0
78.00	627.0
78.90	646.0
79.90	379.0
80.90	483.0
83.00	189.0
85.90	285.0
91.00	304.0
91.90	299.0
92.90	1069.0
97.90	675.0
98.90	670.0
100.90	504.0
103.80	247.0
105.00	297.0
107.00	3000.0
107.90	545.0
109.90	4674.0
110.90	992.0
116.00	152.0
117.00	2090.0
117.90	193.0
121.90	239.0
122.90	213.0
125.00	155.0
127.00	9498.0
127.90	791.0
128.90	3989.0
129.90	406.0
134.90	285.0
136.90	216.0
140.80	518.0
142.00	170.0
147.00	347.0
147.90	533.0
152.90	302.0
153.90	189.0
154.90	294.0
155.90	551.0
160.90	262.0
165.00	236.0
166.00	187.0
166.90	1175.0
167.90	501.0
172.80	170.0
173.90	260.0
175.00	563.0
175.90	155.0
176.90	269.0
178.90	726.0
179.90	659.0
180.80	200.0
184.90	426.0
186.00	3094.0
186.90	889.0
188.00	162.0
189.00	166.0
191.90	201.0
192.90	358.0
195.90	595.0
197.90	21824.0
198.90	1610.0
203.00	224.0
203.90	1089.0
204.90	1359.0
206.00	6772.0
206.90	1103.0
207.90	336.0
209.90	203.0
210.90	237.0
215.90	174.0
216.90	1888.0
217.90	187.0
220.90	956.0
222.10	162.0
222.90	393.0
223.90	3557.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

225.00	899.0
226.90	1428.0
228.90	273.0
242.00	196.0
242.90	225.0
243.90	2611.0
244.90	394.0
245.90	604.0
246.90	166.0
254.90	13494.0
255.90	2219.0
256.90	173.0
257.90	686.0
259.00	174.0
264.90	385.0
272.90	515.0
273.90	980.0
274.90	6084.0
275.90	851.0
276.90	411.0
295.90	1711.0
296.90	272.0
302.90	204.0
323.00	449.0
333.90	452.0
351.90	182.0
364.90	663.0
372.00	252.0
422.00	163.0
422.90	690.0
424.00	153.0
441.00	2082.0
442.00	13040.0
442.90	2749.0
444.00	331.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

m/z	Abundance
38.00	198.0
39.00	795.0
44.00	164.0
49.00	164.0
50.00	2586.0
51.00	10399.0
52.00	460.0
55.90	339.0
57.00	777.0
61.00	194.0
61.90	176.0
63.00	542.0
65.00	243.0
68.90	11531.0
74.00	1401.0
75.00	1873.0
76.00	726.0
77.00	13685.0
78.00	975.0
78.90	998.0
79.90	752.0
80.90	825.0
81.90	212.0
82.90	213.0
84.90	183.0
85.90	283.0
87.10	164.0
90.80	244.0
92.00	196.0
92.90	1494.0
93.90	156.0
98.00	1107.0
98.90	1144.0
100.90	629.0
102.90	156.0
104.00	371.0
104.90	395.0
106.90	4208.0
108.00	603.0
109.90	7747.0
110.90	1147.0
112.00	199.0
115.90	268.0
116.90	3618.0
117.90	204.0
122.00	298.0
122.90	337.0
123.90	192.0
125.00	217.0
127.00	14854.0
128.00	1103.0
128.90	5959.0
129.90	383.0
133.80	180.0
135.00	535.0
135.90	218.0
136.90	210.0
141.00	777.0
142.00	179.0
142.90	158.0
145.90	162.0
146.90	453.0
147.90	1177.0
148.90	202.0
153.00	313.0
153.90	288.0
154.90	564.0
156.00	783.0
157.90	206.0
159.90	168.0
160.90	439.0
165.00	321.0
165.90	243.0
167.00	1765.0
167.90	825.0
168.90	169.0
170.90	155.0
171.80	195.0
172.90	251.0
174.00	293.0
174.90	653.0
175.90	152.0
176.90	352.0
178.90	1224.0
180.00	1124.0
180.90	354.0
184.90	723.0
185.90	5126.0
186.90	1846.0
188.90	344.0
191.00	189.0
191.80	363.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

192.90	681.0
193.70	163.0
195.90	953.0
197.90	36832.0
198.90	2740.0
200.00	151.0
201.50	166.0
202.90	266.0
203.90	1639.0
204.90	2607.0
206.00	9434.0
207.00	1494.0
207.90	445.0
208.90	154.0
210.00	228.0
210.90	388.0
215.90	254.0
216.90	3103.0
217.90	375.0
220.90	1801.0
222.90	743.0
223.90	5723.0
225.00	1353.0
226.00	195.0
226.90	2592.0
227.90	360.0
228.90	627.0
231.00	165.0
233.90	154.0
234.90	158.0
236.90	430.0
240.90	214.0
241.90	265.0
243.00	435.0
244.00	4196.0
244.90	651.0
245.90	790.0
246.80	185.0
253.00	212.0
254.90	24368.0
255.90	3689.0
256.90	253.0
257.90	1395.0
258.90	216.0
264.90	466.0
273.00	722.0
273.90	1851.0
274.90	9904.0
276.00	1475.0
277.00	822.0
284.90	150.0
292.90	242.0
295.90	3035.0
297.00	402.0
302.90	351.0
314.90	327.0
322.90	822.0
324.00	219.0
334.00	583.0
334.90	160.0
340.80	161.0
345.90	165.0
351.90	179.0
354.00	217.0
364.90	1174.0
365.90	160.0
372.00	470.0
373.00	160.0
382.80	182.0
402.00	229.0
402.90	306.0
422.00	243.0
422.90	1380.0
424.00	315.0
440.90	4682.0
442.00	27752.0
443.00	5442.0
444.00	686.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

m/z	Abundance
39.00	634.0
39.90	176.0
44.00	290.0
49.00	189.0
50.00	1860.0
51.00	8505.0
52.00	265.0
55.90	269.0
57.00	508.0
63.00	338.0
64.90	230.0
68.00	253.0
68.90	9949.0
74.00	991.0
75.00	1633.0
76.00	668.0
77.00	11655.0
78.00	738.0
78.90	603.0
79.90	569.0
80.90	791.0
81.90	211.0
82.90	169.0
84.90	184.0
85.90	437.0
91.00	182.0
92.90	1441.0
97.90	982.0
98.90	883.0
100.90	512.0
102.90	168.0
103.90	384.0
104.90	348.0
107.00	3621.0
108.00	636.0
109.90	6515.0
110.90	924.0
116.00	262.0
117.00	2760.0
117.80	160.0
121.90	373.0
122.90	327.0
123.90	197.0
124.90	188.0
127.00	13086.0
127.90	1002.0
128.90	5488.0
129.90	471.0
133.90	158.0
134.90	547.0
137.00	226.0
140.90	689.0
142.00	209.0
142.90	180.0
147.00	414.0
147.90	799.0
148.90	221.0
152.80	291.0
154.00	183.0
154.90	402.0
155.90	703.0
159.90	255.0
160.90	339.0
164.90	356.0
165.90	296.0
166.90	1587.0
168.00	742.0
172.90	273.0
173.90	356.0
175.00	665.0
175.90	155.0
176.90	505.0
178.90	1548.0
179.90	946.0
180.90	461.0
185.00	662.0
185.90	4774.0
186.90	1442.0
188.00	170.0
188.90	236.0
190.90	187.0
191.90	403.0
193.00	626.0
195.90	938.0
197.90	34688.0
198.90	2332.0
199.90	184.0
202.90	372.0
203.90	1282.0
204.90	2307.0
206.00	9188.0
206.90	1604.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

207.90	379.0
209.00	222.0
209.90	267.0
211.00	373.0
214.80	211.0
216.90	3072.0
217.80	355.0
220.90	1660.0
221.90	345.0
222.90	852.0
224.00	5348.0
224.90	1802.0
226.90	2583.0
227.80	397.0
228.90	596.0
230.80	255.0
236.90	288.0
240.90	159.0
241.90	350.0
242.90	397.0
244.00	4831.0
244.90	574.0
245.90	757.0
246.90	202.0
248.90	254.0
254.90	24720.0
255.90	3597.0
256.90	371.0
257.90	1440.0
258.90	248.0
264.90	687.0
272.90	664.0
274.00	1723.0
274.90	11732.0
275.90	1440.0
276.90	989.0
284.90	161.0
292.90	246.0
295.90	3741.0
296.80	476.0
303.00	411.0
314.00	165.0
314.90	392.0
315.90	273.0
322.90	1206.0
323.90	227.0
326.90	176.0
333.90	636.0
341.00	206.0
345.90	235.0
352.00	337.0
352.90	209.0
354.00	322.0
364.90	1423.0
365.80	164.0
371.90	596.0
372.90	193.0
382.90	157.0
401.90	268.0
402.90	349.0
420.90	336.0
422.00	288.0
423.00	1916.0
423.90	508.0
441.00	6159.0
442.00	36760.0
443.00	7025.0
444.00	709.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15