

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122624\
 Data File : BG063832.D
 Acq On : 26 Dec 2024 13:07
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020583

Quant Time: Dec 27 01:20:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 24 00:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.794	152	69613	20.000	ng/u1	0.00
20) Naphthalene-d8	10.584	136	295678	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.439	164	235219	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.188	188	590054	20.000	ng/u1	0.00
79) Chrysene-d12	21.442	240	638728	20.000	ng/u1	0.00
88) Perylene-d12	24.456	264	690224	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	17537	9.872	ng/uL	0.00
4) Pyridine-d5	3.663	84	120579	21.517	ng/u1	0.00
7) Phenol-d5	6.983	99	138332	21.576	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.130	67	91643	20.696	ng/u1	0.00
11) 2-Chlorophenol-d4	7.335	132	88893	21.416	ng/u1	0.00
15) 4-Methylphenol-d8	8.516	113	102381	20.568	ng/u1	0.00
21) Nitrobenzene-d5	8.957	128	49007	23.213	ng/u1	0.00
24) 2-Nitrophenol-d4	9.674	143	54133	22.875	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.220	165	104067	22.713	ng/u1	0.00
31) 4-Chloroaniline-d4	10.725	131	149604	25.716	ng/u1	0.00
46) Dimethylphthalate-d6	13.845	166	358040	22.370	ng/u1	0.00
49) Acenaphthylene-d8	14.127	160	405011	22.198	ng/u1	0.00
54) 4-Nitrophenol-d4	14.674	143	56750	25.025	ng/u1	0.00
60) Fluorene-d10	15.438	176	319963	22.610	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.573	200	67002	21.870	ng/u1	0.00
73) Anthracene-d10	17.288	188	597305	23.545	ng/u1	0.00
81) Pyrene-d10	19.592	212	716655	21.561	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.245	264	734033	22.646	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	19349	9.204	ng/uL#	86
5) Pyridine	3.681	79	132033	21.997	ng/u1#	91
6) Benzaldehyde	6.936	77	86607	22.022	ng/u1	97
8) Phenol	7.006	94	142088	20.795	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.218	93	109397	20.926	ng/u1	96
12) 2-Chlorophenol	7.371	128	92962	22.126	ng/u1	96
13) 2-Methylphenol	8.252	108	103116	20.880	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.316	45	155883	20.085	ng/u1	99
16) Acetophenone	8.616	105	174864	20.712	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.599	70	100296	19.909	ng/u1	92
18) 4-Methylphenol	8.581	108	115614	21.390	ng/u1	98
19) Hexachloroethane	8.869	117	42264	20.472	ng/u1	98
22) Nitrobenzene	8.992	77	150195	21.305	ng/u1	94
23) Isophorone	9.509	82	288294	21.643	ng/u1	99
25) 2-Nitrophenol	9.703	139	57570	23.671	ng/u1#	85
26) 2,4-Dimethylphenol	9.774	107	121502	22.016	ng/u1	96
27) Bis(2-Chloroethoxy)met...	9.997	93	158694	22.027	ng/u1#	93
29) 2,4-Dichlorophenol	10.250	162	101083	22.203	ng/u1	98
30) Naphthalene	10.637	128	331860	22.449	ng/u1	98
32) 4-Chloroaniline	10.749	127	145134	26.220	ng/u1	96
33) Hexachlorobutadiene	10.919	225	71489	19.329	ng/u1	96
34) Caprolactam	11.507	113	42326	26.895	ng/u1	86
35) 4-Chloro-3-methylphenol	11.895	107	120185	22.486	ng/u1	98
36) 2-Methylnaphthalene	12.253	142	239004	22.495	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122624\
 Data File : BG063832.D
 Acq On : 26 Dec 2024 13:07
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD020583


Quant Time: Dec 27 01:20:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 24 00:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.470	142	244142	22.559	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.623	216	140187	18.617	ng/ul#	99
40) Hexachlorocyclopentadiene	12.600	237	47002	15.790	ng/ul	94
41) 2,4,6-Trichlorophenol	12.876	196	88902	20.889	ng/ul	91
42) 2,4,5-Trichlorophenol	12.952	196	96463	21.718	ng/ul	95
43) 1,1'-Biphenyl	13.264	154	330519	21.307	ng/ul	95
44) 2-Chloronaphthalene	13.311	162	261062	20.958	ng/ul	98
45) 2-Nitroaniline	13.516	65	97759	23.429	ng/ul	94
47) Dimethylphthalate	13.892	163	367218	22.892	ng/ul	98
48) 2,6-Dinitrotoluene	14.022	165	73090	24.534	ng/ul#	88
50) Acenaphthylene	14.157	152	435692	22.270	ng/ul	98
51) 3-Nitroaniline	14.351	138	53226	19.324	ng/ul	89
52) Acenaphthene	14.503	153	313598	22.616	ng/ul	96
53) 2,4-Dinitrophenol	14.574	184	34952	22.957	ng/ul	91
55) 4-Nitrophenol	14.691	109	59124	25.108	ng/ul	83
56) Dibenzofuran	14.838	168	451274	23.084	ng/ul	100
57) 2,4-Dinitrotoluene	14.815	165	116299	26.272	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.079	232	82232	21.478	ng/ul#	95
59) Diethylphthalate	15.261	149	359456	23.180	ng/ul	97
61) Fluorene	15.490	166	362372	23.214	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.485	204	186360	21.429	ng/ul	99
63) 4-Nitroaniline	15.514	138	44105	15.995	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.584	198	70675	22.393	ng/ul	95
67) N-Nitrosodiphenylamine	15.702	169	315857	22.106	ng/ul	96
68) 4-Bromophenyl-phenylether	16.378	248	125923	20.940	ng/ul	98
69) Hexachlorobenzene	16.507	284	145946	21.662	ng/ul#	91
70) Atrazine	16.654	200	135168	22.379	ng/ul	93
71) Pentachlorophenol	16.854	266	52650	18.534	ng/ul	98
72) Phenanthrene	17.235	178	675703	23.763	ng/ul	99
74) Anthracene	17.324	178	681655	23.685	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.234	216	145150	18.437	ng/ul	99
76) Pentachlorobenzene	14.762	250	157622	20.210	ng/ul	99
77) Carbazole	17.600	167	609986	26.184	ng/ul	98
78) Di-n-butylphthalate	18.158	149	710692	25.261	ng/ul	98
80) Fluoranthene	19.257	202	842936	22.313	ng/ul	97
82) Pyrene	19.621	202	913645	22.732	ng/ul	96
83) Butylbenzylphthalate	20.514	149	311668	25.261	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.342	252	213859	18.547	ng/ul#	92
85) Benzo(a)anthracene	21.425	228	885272	22.115	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.337	149	448858	25.009	ng/ul	98
87) Chrysene	21.483	228	860627	22.558	ng/ul	98
89) Di-n-octyl phthalate	22.465	149	790971	26.550	ng/ul	100
90) Benzo(b)fluoranthene	23.505	252	870243	22.184	ng/ul	99
91) Benzo(k)fluoranthene	23.563	252	887423	22.678	ng/ul	97
93) Benzo(a)pyrene	24.315	252	815129	22.137	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.852	276	1021375	22.713	ng/ul	97
95) Dibenzo(a,h)anthracene	27.899	278	831922	23.121	ng/ul	98
96) Benzo(g,h,i)perylene	28.910	276	795595	22.296	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 27 01:20:04 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 24 00:44:39 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.204 ng/uL
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion: 88 Resp: 19349
Ion Ratio Lower Upper
88 100
43 24.7 20.9 31.3
58 56.8 57.7 86.5#



#5
Pyridine
Concen: 21.997 ng/uL
RT: 3.681 min Scan# 101
Delta R.T. -0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion: 79 Resp: 132033
Ion Ratio Lower Upper
79 100
52 70.9 60.9 91.3
51 45.8 29.5 44.3#









#6
Benzaldehyde
Concen: 22.022 ng/ul
RT: 6.936 min Scan# 61
Delta R.T. -0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion: 77 Resp: 86607
Ion Ratio Lower Upper
77 100
105 83.2 65.2 97.8
106 79.5 61.0 91.4



#8
Phenol
Concen: 20.795 ng/ul
RT: 7.006 min Scan# 667
Delta R.T. -0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion: 94 Resp: 142088
Ion Ratio Lower Upper
94 100
65 28.3 25.6 38.4
66 39.8 34.2 51.2









#10
Bis(2-Chloroethyl)ether
Concen: 20.926 ng/ul
RT: 7.218 min Scan# 703
Delta R.T. -0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion: 93 Resp: 109397
Ion Ratio Lower Upper
93 100
63 83.8 63.8 95.6
95 36.0 27.6 41.4



#12
2-Chlorophenol
Concen: 22.126 ng/ul
RT: 7.371 min Scan# 729
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:128 Resp: 92962
Ion Ratio Lower Upper
128 100
64 56.2 48.9 73.3
130 30.4 24.2 36.2









#13
2-Methylphenol
Concen: 20.880 ng/ul
RT: 8.252 min Scan# 81
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:108 Resp: 103116
Ion Ratio Lower Upper
108 100
107 89.8 72.6 108.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.085 ng/ul
RT: 8.316 min Scan# 890
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion: 45 Resp: 155883
Ion Ratio Lower Upper
45 100
77 15.3 11.5 17.3
79 11.3 9.2 13.8









#16
Acetophenone
Concen: 20.712 ng/ul
RT: 8.616 min Scan# 941
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:105 Resp: 174864
Ion Ratio Lower Upper
105 100
77 87.0 73.6 110.4
51 31.1 25.4 38.2



#17
N-Nitroso-di-n-propylamine
Concen: 19.909 ng/ul
RT: 8.599 min Scan# 938
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion: 70 Resp: 100296
Ion Ratio Lower Upper
70 100
42 52.4 47.4 71.0
101 8.6 6.1 9.1
130 15.9 11.4 17.2









#18
4-Methylphenol
Concen: 21.390 ng/ul
RT: 8.581 min Scan# 911
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:108 Resp: 115614
Ion Ratio Lower Upper
108 100
107 108.9 89.2 133.8



#19
Hexachloroethane
Concen: 20.472 ng/ul
RT: 8.869 min Scan# 984
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:117 Resp: 42264
Ion Ratio Lower Upper
117 100
201 111.0 86.8 130.2
199 75.3 59.4 89.2









#22
Nitrobenzene
Concen: 21.305 ng/ul
RT: 8.992 min Scan# 1005
Delta R.T. -0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion: 77 Resp: 150195
Ion Ratio Lower Upper
77 100
123 38.8 27.0 40.4
65 13.4 10.8 16.2



#23
Isophorone
Concen: 21.643 ng/ul
RT: 9.509 min Scan# 1093
Delta R.T. -0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion: 82 Resp: 288294
Ion Ratio Lower Upper
82 100
95 6.5 5.8 8.6
138 13.4 10.3 15.5









#25
2-Nitrophenol
Concen: 23.671 ng/ul
RT: 9.703 min Scan# 1126
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:139 Resp: 57570
Ion Ratio Lower Upper
139 100
65 61.0 62.2 93.2#
109 35.8 26.9 40.3



#26
2,4-Dimethylphenol
Concen: 22.016 ng/ul
RT: 9.774 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:107 Resp: 121502
Ion Ratio Lower Upper
107 100
121 41.7 38.4 57.6
122 78.4 61.8 92.8





#27

Bis(2-Chloroethoxy)methane

Concen: 22.027 ng/ul

RT: 9.997 min Scan# 1176

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion: 93 Resp: 158694

Ion Ratio Lower Upper

93 100

95 33.2 23.5 35.3

123 10.1 6.2 9.4#



#29

2,4-Dichlorophenol

Concen: 22.203 ng/ul

RT: 10.250 min Scan# 1219

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:162 Resp: 101083


Ion Ratio Lower Upper

162 100

164 66.2 53.0 79.4

98 37.5 33.0 49.6









#30
Naphthalene
Concen: 22.449 ng/ul
RT: 10.637 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:128 Resp: 331860
Ion Ratio Lower Upper
128 100
129 10.8 8.2 12.2
127 11.3 9.8 14.8



#32
4-Chloroaniline
Concen: 26.220 ng/ul
RT: 10.749 min Scan# 1304
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:127 Resp: 145134
Ion Ratio Lower Upper
127 100
129 32.6 24.5 36.7









#33
Hexachlorobutadiene
Concen: 19.329 ng/ul
RT: 10.919 min Scan# 1333
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:225 Resp: 71489
Ion Ratio Lower Upper
225 100
223 60.4 51.2 76.8
227 61.6 51.1 76.7



#34
Caprolactam
Concen: 26.895 ng/ul
RT: 11.507 min Scan# 1433
Delta R.T. 0.012 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:113 Resp: 42326
Ion Ratio Lower Upper
113 100
55 178.9 162.3 243.5
56 121.4 107.1 160.7









#35
4-Chloro-3-methylphenol
Concen: 22.486 ng/ul
RT: 11.895 min Scan# 1499
Delta R.T. 0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:107 Resp: 120185
Ion Ratio Lower Upper
107 100
144 22.1 16.3 24.5
142 67.0 55.0 82.4



#36
2-Methylnaphthalene
Concen: 22.495 ng/ul
RT: 12.253 min Scan# 1560
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:142 Resp: 239004
Ion Ratio Lower Upper
142 100
141 86.9 69.5 104.3





#37

1-Methylnaphthalene

Concen: 22.559 ng/ul

RT: 12.470 min Scan# 1597

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:142 Resp: 244142

Ion Ratio Lower Upper

142 100

141 94.5 73.8 110.6

116 4.4 3.5 5.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.617 ng/ul

RT: 12.623 min Scan# 1623

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Tgt Ion:216 Resp: 140187




Ion Ratio Lower Upper


216 100

214 78.0 62.6 93.8

179 20.6 16.5 24.7

108 17.4 11.2 16.8#





#40

Hexachlorocyclopentadiene

Concen: 15.790 ng/ul

RT: 12.600 min Scan# 1619

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:237 Resp: 47002

Ion Ratio Lower Upper

237 100

235 69.3 51.0 76.6

272 11.4 9.7 14.5



#41

2,4,6-Trichlorophenol

Concen: 20.889 ng/ul

RT: 12.876 min Scan# 1666

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:196 Resp: 88902


Ion Ratio Lower Upper

196 100

198 101.0 72.4 108.6

200 29.9 22.5 33.7





#42

2,4,5-Trichlorophenol

Concen: 21.718 ng/ul

RT: 12.952 min Scan# 1

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:196 Resp: 96463

Ion Ratio Lower Upper

196 100

198 86.5 74.2 111.4

200 28.8 23.5 35.3



#43

1,1'-Biphenyl

Concen: 21.307 ng/ul

RT: 13.264 min Scan# 1732

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:154 Resp: 330519


Ion Ratio Lower Upper

154 100

153 40.1 35.1 52.7

76 18.4 13.4 20.2





#44

2-Chloronaphthalene

Concen: 20.958 ng/ul

RT: 13.311 min Scan# 1740

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:162 Resp: 261062

Ion Ratio Lower Upper

162 100

164 30.1 25.2 37.8

127 35.5 29.2 43.8



#45

2-Nitroaniline

Concen: 23.429 ng/ul

RT: 13.516 min Scan# 1775

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion: 65 Resp: 97759


Ion Ratio Lower Upper

65 100

92 60.9 49.5 74.3

138 85.6 61.8 92.8









#47
Dimethylphthalate
Concen: 22.892 ng/ul
RT: 13.892 min Scan# 1839
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:163 Resp: 367218
Ion Ratio Lower Upper
163 100
194 3.4 3.0 4.4
164 9.8 8.8 13.2



#48
2,6-Dinitrotoluene
Concen: 24.534 ng/ul
RT: 14.022 min Scan# 1861
Delta R.T. 0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:165 Resp: 73090
Ion Ratio Lower Upper
165 100
89 63.4 59.4 89.2
121 20.5 20.8 31.2#









#50
Acenaphthylene
Concen: 22.270 ng/ul
RT: 14.157 min Scan# 1884
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:152 Resp: 435692
Ion Ratio Lower Upper
152 100
151 18.7 14.6 22.0
153 13.5 10.0 15.0



#51
3-Nitroaniline
Concen: 19.324 ng/ul
RT: 14.351 min Scan# 1917
Delta R.T. 0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:138 Resp: 53226
Ion Ratio Lower Upper
138 100
108 12.0 8.4 12.6
92 135.6 120.4 180.6









#52
Acenaphthene
Concen: 22.616 ng/ul
RT: 14.503 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:153 Resp: 313598
Ion Ratio Lower Upper
153 100
152 45.9 40.2 60.4
154 86.4 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 22.957 ng/ul
RT: 14.574 min Scan# 1955
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:184 Resp: 34952
Ion Ratio Lower Upper
184 100
63 78.0 52.1 78.1
154 59.6 47.8 71.6









#55
4-Nitrophenol
Concen: 25.108 ng/ul
RT: 14.691 min Scan# 1975
Delta R.T. 0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Instrument :
BNA_G
ClientSampleId :
SSTD020583

Tgt Ion:109 Resp: 59124




Ion	Ratio	Lower	Upper
109	100		
139	118.5	86.4	129.6
65	130.9	127.0	190.4




#56
Dibenzofuran
Concen: 23.084 ng/ul
RT: 14.838 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:168 Resp: 451274

Ion	Ratio	Lower	Upper
168	100		
139	34.8	27.8	41.8









#57
2,4-Dinitrotoluene
Concen: 26.272 ng/ul
RT: 14.815 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:165 Resp: 116299
Ion Ratio Lower Upper
165 100
63 56.5 45.0 67.4
182 2.4 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.478 ng/ul
RT: 15.079 min Scan# 2041
Delta R.T. 0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:232 Resp: 82232
Ion Ratio Lower Upper
232 100
131 35.9 30.3 45.5
130 0.0 1.8 2.6#
166 27.8 24.9 37.3









#59
Diethylphthalate
Concen: 23.180 ng/ul
RT: 15.261 min Scan# 2072
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:149 Resp: 359456
Ion Ratio Lower Upper
149 100
177 23.7 17.6 26.4
150 11.4 9.0 13.4



#61
Fluorene
Concen: 23.214 ng/ul
RT: 15.490 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:166 Resp: 362372
Ion Ratio Lower Upper
166 100
165 98.0 81.9 122.9
167 13.5 11.4 17.0





#62

4-Chlorophenyl-phenylether

Concen: 21.429 ng/ul

RT: 15.485 min Scan# 2110

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:204 Resp: 186360

Ion Ratio Lower Upper

204 100

206 32.8 25.6 38.4

141 51.9 42.3 63.5



#63

4-Nitroaniline

Concen: 15.995 ng/ul

RT: 15.514 min Scan# 2115

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:138 Resp: 44105


Ion Ratio Lower Upper

138 100

92 71.5 53.2 79.8

108 93.8 79.4 119.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 22.393 ng/ul
 RT: 15.584 min Scan# 2127
 Delta R.T. -0.000 min
 Lab File: BG063832.D
 Acq: 26 Dec 2024 13:07




Instrument :
 BNA_G
 ClientSampleId :
 SST020583


Tgt Ion:198 Resp: 70675
 Ion Ratio Lower Upper
 198 100
 182 4.7 3.4 5.2
 77 28.3 25.2 37.8



#67
 N-Nitrosodiphenylamine
 Concen: 22.106 ng/ul
 RT: 15.702 min Scan# 2147
 Delta R.T. 0.006 min
 Lab File: BG063832.D
 Acq: 26 Dec 2024 13:07

Tgt Ion:169 Resp: 315857
 Ion Ratio Lower Upper
 169 100
 168 71.4 54.0 81.0
 167 35.9 29.8 44.8









#68
4-Bromophenyl-phenylether
Concen: 20.940 ng/ul
RT: 16.378 min Scan# 211
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:248 Resp: 125923
Ion Ratio Lower Upper
248 100
250 93.1 76.1 114.1
141 58.8 46.0 69.0



#69
Hexachlorobenzene
Concen: 21.662 ng/ul
RT: 16.507 min Scan# 2284
Delta R.T. 0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:284 Resp: 145946
Ion Ratio Lower Upper
284 100
142 23.0 23.9 35.9#
249 28.3 24.6 36.8





#70

Atrazine

Concen: 22.379 ng/ul

RT: 16.654 min Scan# 21

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:200 Resp: 135168

Ion Ratio Lower Upper

200 100

173 23.0 21.3 31.9

215 51.5 37.8 56.8



#71

Pentachlorophenol

Concen: 18.534 ng/ul

RT: 16.854 min Scan# 2343

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:266 Resp: 52650


Ion Ratio Lower Upper

266 100

264 60.5 49.3 73.9

268 60.5 50.5 75.7





#72

Phenanthrene

Concen: 23.763 ng/ul

RT: 17.235 min Scan# 2408

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:178 Resp: 675703

Ion Ratio Lower Upper

178 100

179 15.0 12.2 18.2

176 19.0 15.8 23.6



#74

Anthracene

Concen: 23.685 ng/ul

RT: 17.324 min Scan# 2423

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:178 Resp: 681655


Ion Ratio Lower Upper

178 100

179 16.4 12.4 18.6

176 18.5 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.437 ng/ul
RT: 13.234 min Scan# 1727
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:216 Resp: 145150
Ion Ratio Lower Upper
216 100
214 75.1 60.2 90.4
218 47.6 37.6 56.4



#76
Pentachlorobenzene
Concen: 20.210 ng/ul
RT: 14.762 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:250 Resp: 157622
Ion Ratio Lower Upper
250 100
248 63.7 49.8 74.8
252 63.4 50.4 75.6





#77

Carbazole

Concen: 26.184 ng/ul

RT: 17.600 min Scan# 2470

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:167 Resp: 609986

Ion Ratio Lower Upper

167 100

166 22.1 18.2 27.4

139 11.3 10.1 15.1



#78

Di-n-butylphthalate

Concen: 25.261 ng/ul

RT: 18.158 min Scan# 2565

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:149 Resp: 710692


Ion Ratio Lower Upper

149 100

150 8.7 7.2 10.8

104 5.4 5.4 8.0









#80
Fluoranthene
Concen: 22.313 ng/ul
RT: 19.257 min Scan# 21
Delta R.T. 0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:202 Resp: 842936
Ion Ratio Lower Upper
202 100
101 11.6 8.1 12.1
100 7.8 6.5 9.7



#82
Pyrene
Concen: 22.732 ng/ul
RT: 19.621 min Scan# 2814
Delta R.T. 0.006 min
Lab File: BG063832.D
Acq: 26 Dec 2024 13:07

Tgt Ion:202 Resp: 913645
Ion Ratio Lower Upper
202 100
101 12.7 8.7 13.1
100 10.9 7.8 11.6





#83

Butylbenzylphthalate

Concen: 25.261 ng/ul

RT: 20.514 min Scan# 2966

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:149 Resp: 311668

Ion Ratio Lower Upper

149 100

91 82.4 68.7 103.1

206 22.2 17.3 25.9



#84

3,3'-Dichlorobenzidine

Concen: 18.547 ng/ul

RT: 21.342 min Scan# 3107

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:252 Resp: 213859


Ion Ratio Lower Upper

252 100

254 62.7 55.8 83.6

126 9.8 6.4 9.6#





#85

Benzo(a)anthracene

Concen: 22.115 ng/ul

RT: 21.425 min Scan# 3121

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:228 Resp: 885272

Ion Ratio Lower Upper

228 100

229 19.4 15.7 23.5

226 27.0 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 25.009 ng/ul

RT: 21.337 min Scan# 3106

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:149 Resp: 448858


Ion Ratio Lower Upper

149 100

167 26.6 22.1 33.1

279 3.6 2.9 4.3





#87

Chrysene

Concen: 22.558 ng/ul

RT: 21.483 min Scan# 3131

Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:228 Resp: 860627

Ion Ratio Lower Upper

228 100

226 29.7 23.4 35.0

229 20.4 14.8 22.2



#89

Di-n-octyl phthalate

Concen: 26.550 ng/ul




RT: 22.465 min Scan# 3298


Delta R.T. -0.000 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Tgt Ion:149 Resp: 790971





#90

Benzo(b)fluoranthene

Concen: 22.184 ng/ul

RT: 23.505 min Scan# 3475

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:252 Resp: 870243

Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 9.1 7.0 10.4



#91

Benzo(k)fluoranthene

Concen: 22.678 ng/ul

RT: 23.563 min Scan# 3485

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:252 Resp: 887423


Ion Ratio Lower Upper

252 100

253 23.4 17.7 26.5

125 9.6 6.5 9.7





#93

Benzo(a)pyrene

Concen: 22.137 ng/ul

RT: 24.315 min Scan# 3613

Delta R.T. 0.012 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:252 Resp: 815129

Ion Ratio Lower Upper

252 100

253 22.7 16.6 24.8

125 10.0 8.0 12.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.713 ng/ul

RT: 27.852 min Scan# 4215

Delta R.T. 0.006 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07




Tgt Ion:276 Resp: 1021375


Ion Ratio Lower Upper

276 100

138 18.7 13.7 20.5

277 23.8 18.4 27.6





#95

Dibenzo(a,h)anthracene

Concen: 23.121 ng/ul

RT: 27.899 min Scan# 4223

Delta R.T. 0.012 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:278 Resp: 831922

Ion Ratio Lower Upper

278 100

139 13.3 9.7 14.5

279 22.3 17.3 25.9



#96

Benzo(g,h,i)perylene

Concen: 22.296 ng/ul

RT: 28.910 min Scan# 4395

Delta R.T. 0.012 min

Lab File: BG063832.D

Acq: 26 Dec 2024 13:07

Tgt Ion:276 Resp: 795595

Ion Ratio Lower Upper

276 100

138 19.5 13.8 20.8

277 23.4 17.8 26.8

