

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122624\
 Data File : BG063873.D
 Acq On : 27 Dec 2024 18:03
 Operator : RC/JU
 Sample : P5323-17
 Misc : GCMS CONFIRMATION
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 YE639

Quant Time: Dec 27 23:31:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 24 00:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	112337	20.000	ng/u1	0.00
20) Naphthalene-d8	10.584	136	420487	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.438	164	268785	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.194	188	630235	20.000	ng/u1	0.00
79) Chrysene-d12	21.442	240	766175	20.000	ng/u1	0.00
88) Perylene-d12	24.468	264	780240	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.187	96	95	0.033	ng/uL	-0.07
4) Pyridine-d5	3.669	84	145	0.016	ng/u1	0.00
7) Phenol-d5	6.906	99	162	0.016	ng/u1	-0.08
9) Bis-(2-Chloroethyl)eth...	7.164	67	209	0.029	ng/u1	0.03
11) 2-Chlorophenol-d4	0.000	132	0	0.000	ng/u1	
15) 4-Methylphenol-d8	0.000	113	0	0.000	ng/u1	
21) Nitrobenzene-d5	9.109	128	82	0.027	ng/u1	0.16
24) 2-Nitrophenol-d4	0.000	143	0	0.000	ng/u1	
28) 2,4-Dichlorophenol-d3	10.132	165	115	0.018	ng/u1	-0.09
31) 4-Chloroaniline-d4	10.725	131	72	0.009	ng/u1	0.00
46) Dimethylphthalate-d6	0.000	166	0	0.000	ng/u1	
49) Acenaphthylene-d8	0.000	160	0	0.000	ng/u1	
54) 4-Nitrophenol-d4	0.000	143	0	0.000	ng/u1	
60) Fluorene-d10	0.000	176	0	0.000	ng/u1	
65) 4,6-Dinitro-2-methylph...	0.000	200	0	0.000	ng/u1	
73) Anthracene-d10	17.300	188	2334	0.086	ng/u1	0.01
81) Pyrene-d10	19.556	212	74	0.002	ng/u1	-0.03
92) Benzo(a)pyrene-d12	0.000	264	0	0.000	ng/u1	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122624\
Data File : BG063873.D
Acq On : 27 Dec 2024 18:03
Operator : RC/JU
Sample : P5323-17
Misc : GCMS CONFIRMATION
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
YE639

Quant Time: Dec 27 23:31:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 24 00:44:39 2024
Response via : Initial Calibration

