

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122624\
 Data File : BG063876.D
 Acq On : 27 Dec 2024 20:33
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc : GCMS CONFIRMATION
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020586

Quant Time: Dec 27 23:32:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 24 00:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.800	152	115473	20.000	ng/u1	0.00
20) Naphthalene-d8	10.585	136	479320	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.439	164	347729	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.195	188	804312	20.000	ng/u1	0.00
79) Chrysene-d12	21.449	240	846302	20.000	ng/u1	0.01
88) Perylene-d12	24.469	264	854579	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	25335	8.598	ng/uL	0.00
4) Pyridine-d5	3.670	84	191175	20.566	ng/u1	0.00
7) Phenol-d5	6.984	99	204219	19.203	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.130	67	142350	19.380	ng/u1	0.00
11) 2-Chlorophenol-d4	7.336	132	134224	19.495	ng/u1	0.00
15) 4-Methylphenol-d8	8.517	113	162465	19.676	ng/u1	0.00
21) Nitrobenzene-d5	8.952	128	72560	21.202	ng/u1	0.00
24) 2-Nitrophenol-d4	9.674	143	81735	21.306	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.221	165	155403	20.923	ng/u1	0.00
31) 4-Chloroaniline-d4	10.726	131	205682	21.810	ng/u1	0.00
46) Dimethylphthalate-d6	13.846	166	495348	20.935	ng/u1	0.00
49) Acenaphthylene-d8	14.134	160	563436	20.889	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	66577	19.859	ng/u1	0.00
60) Fluorene-d10	15.438	176	433691	20.731	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.573	200	84679	20.277	ng/u1	0.00
73) Anthracene-d10	17.295	188	750961	21.716	ng/u1	0.00
81) Pyrene-d10	19.592	212	890940	20.231	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.263	264	866454	21.590	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.300	88	28858	8.275	ng/uL	97
5) Pyridine	3.687	79	204925	20.582	ng/u1	95
6) Benzaldehyde	6.937	77	133289	20.432	ng/u1	97
8) Phenol	7.013	94	219134	19.334	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.224	93	165392	19.072	ng/u1	93
12) 2-Chlorophenol	7.365	128	140150	20.109	ng/u1	90
13) 2-Methylphenol	8.253	108	159411	19.459	ng/u1	93
14) 2,2'-oxybis(1-Chloropr...	8.317	45	250133	19.429	ng/u1	97
16) Acetophenone	8.617	105	272766	19.477	ng/u1	94
17) N-Nitroso-di-n-propyla...	8.599	70	155601	18.621	ng/u1	92
18) 4-Methylphenol	8.582	108	173924	19.399	ng/u1	98
19) Hexachloroethane	8.870	117	65221	19.046	ng/u1	97
22) Nitrobenzene	8.993	77	229798	20.107	ng/u1	98
23) Isophorone	9.516	82	431463	19.981	ng/u1	98
25) 2-Nitrophenol	9.710	139	84511	21.435	ng/u1	88
26) 2,4-Dimethylphenol	9.774	107	185123	20.692	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.998	93	242819	20.790	ng/u1	94
29) 2,4-Dichlorophenol	10.244	162	154104	20.881	ng/u1	100
30) Naphthalene	10.638	128	492721	20.560	ng/u1	99
32) 4-Chloroaniline	10.750	127	197392	21.999	ng/u1	98
33) Hexachlorobutadiene	10.920	225	114068	19.025	ng/u1	97
34) Caprolactam	11.502	113	34632	13.575	ng/u1	87
35) 4-Chloro-3-methylphenol	11.895	107	168416	19.437	ng/u1	96
36) 2-Methylnaphthalene	12.248	142	347416	20.171	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122624\
 Data File : BG063876.D
 Acq On : 27 Dec 2024 20:33
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc : GCMS CONFIRMATION
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020586

Quant Time: Dec 27 23:32:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 24 00:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.471	142	356820	20.339	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.630	216	220445	19.803	ng/ul	98
40) Hexachlorocyclopentadiene	12.600	237	77704	17.658	ng/ul	95
41) 2,4,6-Trichlorophenol	12.877	196	127498	20.265	ng/ul	91
42) 2,4,5-Trichlorophenol	12.953	196	131208	19.983	ng/ul	99
43) 1,1'-Biphenyl	13.264	154	487484	21.258	ng/ul	93
44) 2-Chloronaphthalene	13.311	162	386556	20.991	ng/ul	99
45) 2-Nitroaniline	13.523	65	135068	21.897	ng/ul	97
47) Dimethylphthalate	13.893	163	490874	20.700	ng/ul	99
48) 2,6-Dinitrotoluene	14.022	165	99125	22.507	ng/ul#	88
50) Acenaphthylene	14.157	152	599056	20.713	ng/ul	98
51) 3-Nitroaniline	14.351	138	72265	17.747	ng/ul	92
52) Acenaphthene	14.504	153	429429	20.949	ng/ul	95
53) 2,4-Dinitrophenol	14.575	184	37422	16.627	ng/ul	92
55) 4-Nitrophenol	14.686	109	65914	18.934	ng/ul	93
56) Dibenzofuran	14.839	168	624913	21.623	ng/ul	96
57) 2,4-Dinitrotoluene	14.816	165	147844	22.592	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.080	232	114572	20.243	ng/ul	93
59) Diethylphthalate	15.262	149	482034	21.027	ng/ul	98
61) Fluorene	15.491	166	480283	20.812	ng/ul	94
62) 4-Chlorophenyl-phenyle...	15.485	204	258925	20.140	ng/ul	95
63) 4-Nitroaniline	15.521	138	62915	15.434	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.591	198	86378	20.078	ng/ul#	89
67) N-Nitrosodiphenylamine	15.703	169	412237	21.166	ng/ul	95
68) 4-Bromophenyl-phenylether	16.384	248	171875	20.968	ng/ul	99
69) Hexachlorobenzene	16.508	284	188675	20.544	ng/ul	97
70) Atrazine	16.655	200	178605	21.693	ng/ul	95
71) Pentachlorophenol	16.860	266	60077	15.515	ng/ul	97
72) Phenanthrene	17.236	178	858340	22.145	ng/ul	99
74) Anthracene	17.330	178	869577	22.166	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.235	216	225030	20.970	ng/ul	98
76) Pentachlorobenzene	14.769	250	222637	20.942	ng/ul	96
77) Carbazole	17.600	167	761565	23.982	ng/ul	98
78) Di-n-butylphthalate	18.159	149	888470	23.167	ng/ul	99
80) Fluoranthene	19.257	202	1047878	20.935	ng/ul	97
82) Pyrene	19.622	202	1120353	21.038	ng/ul	95
83) Butylbenzylphthalate	20.521	149	386820	23.662	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.349	252	288258	18.868	ng/ul	97
85) Benzo(a)anthracene	21.431	228	1078467	20.334	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.343	149	571694	24.040	ng/ul	100
87) Chrysene	21.490	228	1003797	19.857	ng/ul	96
89) Di-n-octyl phthalate	22.477	149	984683	26.696	ng/ul	100
90) Benzo(b)fluoranthene	23.511	252	1015876	20.916	ng/ul	99
91) Benzo(k)fluoranthene	23.576	252	1086277	22.421	ng/ul	99
93) Benzo(a)pyrene	24.328	252	970939	21.297	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	27.871	276	1189354	21.362	ng/ul	96
95) Dibenzo(a,h)anthracene	27.924	278	957731	21.498	ng/ul#	96
96) Benzo(g,h,i)perylene	28.940	276	925135	20.940	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.275 ng/uL
RT: 3.300 min Scan# 30
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion: 88 Resp: 28858
Ion Ratio Lower Upper
88 100
43 26.0 20.9 31.3
58 75.9 57.7 86.5



#5
Pyridine
Concen: 20.582 ng/uL
RT: 3.687 min Scan# 102
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion: 79 Resp: 204925
Ion Ratio Lower Upper
79 100
52 73.3 60.9 91.3
51 32.5 29.5 44.3









#6
Benzaldehyde
Concen: 20.432 ng/ul
RT: 6.937 min Scan# 611
Delta R.T. -0.005 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion: 77 Resp: 133289
Ion Ratio Lower Upper
77 100
105 83.5 65.2 97.8
106 79.7 61.0 91.4



#8
Phenol
Concen: 19.334 ng/ul
RT: 7.013 min Scan# 668
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion: 94 Resp: 219134
Ion Ratio Lower Upper
94 100
65 32.4 25.6 38.4
66 38.9 34.2 51.2









#10
Bis(2-Chloroethyl)ether
Concen: 19.072 ng/ul
RT: 7.224 min Scan# 704
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion: 93 Resp: 165392
Ion Ratio Lower Upper
93 100
63 86.3 63.8 95.6
95 30.3 27.6 41.4



#12
2-Chlorophenol
Concen: 20.109 ng/ul
RT: 7.365 min Scan# 728
Delta R.T. -0.005 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:128 Resp: 140150
Ion Ratio Lower Upper
128 100
64 71.6 48.9 73.3
130 31.0 24.2 36.2





#13

2-Methylphenol

Concen: 19.459 ng/ul

RT: 8.253 min Scan# 81

Delta R.T. 0.001 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :





SSTD020586

Tgt Ion:108 Resp: 159411

Ion Ratio Lower Upper

108 100

107 84.4 72.6 108.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.429 ng/ul

RT: 8.317 min Scan# 890

Delta R.T. 0.001 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33




Tgt Ion: 45 Resp: 250133


Ion Ratio Lower Upper

45 100

77 16.0 11.5 17.3

79 11.0 9.2 13.8









#16
Acetophenone
Concen: 19.477 ng/ul
RT: 8.617 min Scan# 941
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:105 Resp: 272766
Ion Ratio Lower Upper
105 100
77 87.6 73.6 110.4
51 27.3 25.4 38.2



#17
N-Nitroso-di-n-propylamine
Concen: 18.621 ng/ul
RT: 8.599 min Scan# 938
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion: 70 Resp: 155601
Ion Ratio Lower Upper
70 100
42 51.2 47.4 71.0
101 7.8 6.1 9.1
130 13.8 11.4 17.2









#18
4-Methylphenol
Concen: 19.399 ng/ul
RT: 8.582 min Scan# 913
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:108 Resp: 173924
Ion Ratio Lower Upper
108 100
107 109.0 89.2 133.8



#19
Hexachloroethane
Concen: 19.046 ng/ul
RT: 8.870 min Scan# 984
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:117 Resp: 65221
Ion Ratio Lower Upper
117 100
201 109.8 86.8 130.2
199 78.9 59.4 89.2









#22
Nitrobenzene
Concen: 20.107 ng/ul
RT: 8.993 min Scan# 1005
Delta R.T. -0.005 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion: 77 Resp: 229798
Ion Ratio Lower Upper
77 100
123 35.3 27.0 40.4
65 13.4 10.8 16.2



#23
Isophorone
Concen: 19.981 ng/ul
RT: 9.516 min Scan# 1094
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion: 82 Resp: 431463
Ion Ratio Lower Upper
82 100
95 6.0 5.8 8.6
138 13.6 10.3 15.5









#25
2-Nitrophenol
Concen: 21.435 ng/ul
RT: 9.710 min Scan# 1126
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:139 Resp: 84511
Ion Ratio Lower Upper
139 100
65 65.1 62.2 93.2
109 37.3 26.9 40.3



#26
2,4-Dimethylphenol
Concen: 20.692 ng/ul
RT: 9.774 min Scan# 1138
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:107 Resp: 185123
Ion Ratio Lower Upper
107 100
121 45.5 38.4 57.6
122 76.6 61.8 92.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.790 ng/ul
 RT: 9.998 min Scan# 1176
 Delta R.T. 0.001 min
 Lab File: BG063876.D
 Acq: 27 Dec 2024 20:33

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020586

Tgt Ion: 93 Resp: 242819




Ion	Ratio	Lower	Upper
93	100		
95	32.9	23.5	35.3
123	8.5	6.2	9.4




#29
 2,4-Dichlorophenol
 Concen: 20.881 ng/ul
 RT: 10.244 min Scan# 1218
 Delta R.T. 0.001 min
 Lab File: BG063876.D
 Acq: 27 Dec 2024 20:33

Tgt Ion:162 Resp: 154104

Ion	Ratio	Lower	Upper
162	100		
164	66.4	53.0	79.4
98	41.0	33.0	49.6









#30
Naphthalene
Concen: 20.560 ng/ul
RT: 10.638 min Scan# 1285
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:128 Resp: 492721
Ion Ratio Lower Upper
128 100
129 10.8 8.2 12.2
127 12.1 9.8 14.8



#32
4-Chloroaniline
Concen: 21.999 ng/ul
RT: 10.750 min Scan# 1304
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:127 Resp: 197392
Ion Ratio Lower Upper
127 100
129 31.9 24.5 36.7









#33
Hexachlorobutadiene
Concen: 19.025 ng/ul
RT: 10.920 min Scan# 11
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:225 Resp: 114068
Ion Ratio Lower Upper
225 100
223 66.3 51.2 76.8
227 61.0 51.1 76.7



#34
Caprolactam
Concen: 13.575 ng/ul
RT: 11.502 min Scan# 1432
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:113 Resp: 34632
Ion Ratio Lower Upper
113 100
55 172.0 162.3 243.5
56 130.6 107.1 160.7









#35
4-Chloro-3-methylphenol
Concen: 19.437 ng/ul
RT: 11.895 min Scan# 1499
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:107 Resp: 168416
Ion Ratio Lower Upper
107 100
144 21.8 16.3 24.5
142 72.2 55.0 82.4



#36
2-Methylnaphthalene
Concen: 20.171 ng/ul
RT: 12.248 min Scan# 1559
Delta R.T. -0.005 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:142 Resp: 347416
Ion Ratio Lower Upper
142 100
141 86.4 69.5 104.3





#37

1-Methylnaphthalene

Concen: 20.339 ng/ul

RT: 12.471 min Scan# 1597

Delta R.T. 0.001 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :

SSTD020586





Tgt Ion:142 Resp: 356820

Ion Ratio Lower Upper

142 100

141 89.9 73.8 110.6

116 3.6 3.5 5.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.803 ng/ul

RT: 12.630 min Scan# 1624

Delta R.T. 0.006 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Tgt Ion:216 Resp: 220445




Ion Ratio Lower Upper


216 100

214 77.0 62.6 93.8

179 18.9 16.5 24.7

108 16.2 11.2 16.8





#40

Hexachlorocyclopentadiene

Concen: 17.658 ng/ul

RT: 12.600 min Scan# 1

Delta R.T. 0.001 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :

SSTD020586





Tgt Ion:237 Resp: 77704

Ion Ratio Lower Upper

237 100

235 60.4 51.0 76.6

272 10.0 9.7 14.5



#41

2,4,6-Trichlorophenol

Concen: 20.265 ng/ul

RT: 12.877 min Scan# 1666

Delta R.T. 0.006 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33




Tgt Ion:196 Resp: 127498


Ion Ratio Lower Upper

196 100

198 99.7 72.4 108.6

200 32.6 22.5 33.7









#42
2,4,5-Trichlorophenol
Concen: 19.983 ng/ul
RT: 12.953 min Scan# 1679
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:196 Resp: 131208
Ion Ratio Lower Upper
196 100
198 92.6 74.2 111.4
200 32.0 23.5 35.3



#43
1,1'-Biphenyl
Concen: 21.258 ng/ul
RT: 13.264 min Scan# 1732
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:154 Resp: 487484
Ion Ratio Lower Upper
154 100
153 38.1 35.1 52.7
76 16.3 13.4 20.2





#44

2-Chloronaphthalene

Concen: 20.991 ng/ul

RT: 13.311 min Scan# 1

Delta R.T. 0.001 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :

SSTD020586





Tgt Ion:162 Resp: 386556

Ion Ratio Lower Upper

162 100

164 31.7 25.2 37.8

127 35.2 29.2 43.8



#45

2-Nitroaniline

Concen: 21.897 ng/ul

RT: 13.523 min Scan# 1776

Delta R.T. 0.006 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33




Tgt Ion: 65 Resp: 135068


Ion Ratio Lower Upper

65 100

92 64.2 49.5 74.3

138 79.3 61.8 92.8









#47
Dimethylphthalate
Concen: 20.700 ng/ul
RT: 13.893 min Scan# 1839
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:163 Resp: 490874
Ion Ratio Lower Upper
163 100
194 3.2 3.0 4.4
164 10.5 8.8 13.2



#48
2,6-Dinitrotoluene
Concen: 22.507 ng/ul
RT: 14.022 min Scan# 1861
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:165 Resp: 99125
Ion Ratio Lower Upper
165 100
89 64.5 59.4 89.2
121 19.1 20.8 31.2#









#50
Acenaphthylene
Concen: 20.713 ng/ul
RT: 14.157 min Scan# 1884
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:152 Resp: 599056
Ion Ratio Lower Upper
152 100
151 19.4 14.6 22.0
153 12.7 10.0 15.0



#51
3-Nitroaniline
Concen: 17.747 ng/ul
RT: 14.351 min Scan# 1917
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:138 Resp: 72265
Ion Ratio Lower Upper
138 100
108 9.7 8.4 12.6
92 140.0 120.4 180.6









#52
Acenaphthene
Concen: 20.949 ng/ul
RT: 14.504 min Scan# 1943
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:153 Resp: 429429
Ion Ratio Lower Upper
153 100
152 48.6 40.2 60.4
154 82.6 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 16.627 ng/ul
RT: 14.575 min Scan# 1955
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:184 Resp: 37422
Ion Ratio Lower Upper
184 100
63 74.4 52.1 78.1
154 56.6 47.8 71.6









#55
4-Nitrophenol
Concen: 18.934 ng/ul
RT: 14.686 min Scan# 1974
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:109 Resp: 65914
Ion Ratio Lower Upper
109 100
139 103.2 86.4 129.6
65 147.9 127.0 190.4



#56
Dibenzofuran
Concen: 21.623 ng/ul
RT: 14.839 min Scan# 2000
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:168 Resp: 624913
Ion Ratio Lower Upper
168 100
139 32.2 27.8 41.8





#57

2,4-Dinitrotoluene

Concen: 22.592 ng/ul

RT: 14.816 min Scan# 1996

Delta R.T. 0.001 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :

SSTD020586





Tgt Ion:165 Resp: 147844

Ion Ratio Lower Upper

165 100

63 51.8 45.0 67.4

182 2.1 1.8 2.8



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.243 ng/ul

RT: 15.080 min Scan# 2041

Delta R.T. 0.006 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Tgt Ion:232 Resp: 114572




Ion Ratio Lower Upper


232 100

131 33.7 30.3 45.5

130 1.9 1.8 2.6

166 26.8 24.9 37.3









#59
Diethylphthalate
Concen: 21.027 ng/ul
RT: 15.262 min Scan# 2072
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:149 Resp: 482034
Ion Ratio Lower Upper
149 100
177 23.0 17.6 26.4
150 11.9 9.0 13.4



#61
Fluorene
Concen: 20.812 ng/ul
RT: 15.491 min Scan# 2111
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:166 Resp: 480283
Ion Ratio Lower Upper
166 100
165 96.3 81.9 122.9
167 13.1 11.4 17.0









#62
4-Chlorophenyl-phenylether
Concen: 20.140 ng/ul
RT: 15.485 min Scan# 2110
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:204 Resp: 258925
Ion Ratio Lower Upper
204 100
206 33.6 25.6 38.4
141 48.9 42.3 63.5



#63
4-Nitroaniline
Concen: 15.434 ng/ul
RT: 15.521 min Scan# 2116
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:138 Resp: 62915
Ion Ratio Lower Upper
138 100
92 64.4 53.2 79.8
108 86.4 79.4 119.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.078 ng/ul
 RT: 15.591 min Scan# 2128
 Delta R.T. 0.006 min
 Lab File: BG063876.D
 Acq: 27 Dec 2024 20:33




Instrument :
 BNA_G
 ClientSampleId :
 SST020586


Tgt Ion:198 Resp: 86378
 Ion Ratio Lower Upper
 198 100
 182 5.3 3.4 5.2#
 77 24.7 25.2 37.8#



#67
 N-Nitrosodiphenylamine
 Concen: 21.166 ng/ul
 RT: 15.703 min Scan# 2147
 Delta R.T. 0.006 min
 Lab File: BG063876.D
 Acq: 27 Dec 2024 20:33

Tgt Ion:169 Resp: 412237
 Ion Ratio Lower Upper
 169 100
 168 73.3 54.0 81.0
 167 37.4 29.8 44.8









#68
4-Bromophenyl-phenylether
Concen: 20.968 ng/ul
RT: 16.384 min Scan# 21
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:248 Resp: 171875
Ion Ratio Lower Upper
248 100
250 95.3 76.1 114.1
141 55.4 46.0 69.0



#69
Hexachlorobenzene
Concen: 20.544 ng/ul
RT: 16.508 min Scan# 2284
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:284 Resp: 188675
Ion Ratio Lower Upper
284 100
142 28.2 23.9 35.9
249 32.7 24.6 36.8





#70

Atrazine

Concen: 21.693 ng/ul

RT: 16.655 min Scan# 21

Delta R.T. 0.001 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :

SSTD020586





Tgt Ion:200 Resp: 178605

Ion Ratio Lower Upper

200 100

173 21.6 21.3 31.9

215 45.5 37.8 56.8



#71

Pentachlorophenol

Concen: 15.515 ng/ul

RT: 16.860 min Scan# 2344

Delta R.T. 0.006 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33




Tgt Ion:266 Resp: 60077


Ion Ratio Lower Upper

266 100

264 64.1 49.3 73.9

268 60.6 50.5 75.7









#72
Phenanthrene
Concen: 22.145 ng/ul
RT: 17.236 min Scan# 2408
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:178 Resp: 858340
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.2
176 20.1 15.8 23.6



#74
Anthracene
Concen: 22.166 ng/ul
RT: 17.330 min Scan# 2424
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:178 Resp: 869577
Ion Ratio Lower Upper
178 100
179 15.4 12.4 18.6
176 17.1 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.970 ng/ul
RT: 13.235 min Scan# 1727
Delta R.T. 0.001 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:216 Resp: 225030
Ion Ratio Lower Upper
216 100
214 76.9 60.2 90.4
218 46.0 37.6 56.4



#76
Pentachlorobenzene
Concen: 20.942 ng/ul
RT: 14.769 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:250 Resp: 222637
Ion Ratio Lower Upper
250 100
248 59.0 49.8 74.8
252 66.5 50.4 75.6









#77
 Carbazole
 Concen: 23.982 ng/ul
 RT: 17.600 min Scan# 2470
 Delta R.T. 0.006 min
 Lab File: BG063876.D
 Acq: 27 Dec 2024 20:33




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020586


Tgt Ion:167 Resp: 761565
 Ion Ratio Lower Upper
 167 100
 166 22.4 18.2 27.4
 139 11.3 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 23.167 ng/ul
 RT: 18.159 min Scan# 2565
 Delta R.T. 0.006 min
 Lab File: BG063876.D
 Acq: 27 Dec 2024 20:33

Tgt Ion:149 Resp: 888470
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.2 10.8
 104 6.0 5.4 8.0





#80

Fluoranthene

Concen: 20.935 ng/uI

RT: 19.257 min Scan# 21

Delta R.T. 0.006 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :

SSTD020586





Tgt Ion:202 Resp: 1047878

Ion Ratio Lower Upper

202 100

101 11.5 8.1 12.1

100 9.0 6.5 9.7



#82

Pyrene

Concen: 21.038 ng/uI

RT: 19.622 min Scan# 2814

Delta R.T. 0.006 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33




Tgt Ion:202 Resp: 1120353


Ion Ratio Lower Upper

202 100

101 13.0 8.7 13.1

100 11.0 7.8 11.6





#83

Butylbenzylphthalate

Concen: 23.662 ng/ul

RT: 20.521 min Scan# 2967

Delta R.T. 0.012 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :

SSTD020586





Tgt Ion:149 Resp: 386820

Ion Ratio Lower Upper

149 100

91 81.1 68.7 103.1

206 22.2 17.3 25.9



#84

3,3'-Dichlorobenzidine

Concen: 18.868 ng/ul

RT: 21.349 min Scan# 3108

Delta R.T. 0.006 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33




Tgt Ion:252 Resp: 288258


Ion Ratio Lower Upper

252 100

254 66.8 55.8 83.6

126 8.1 6.4 9.6









#85
Benzo(a)anthracene
Concen: 20.334 ng/ul
RT: 21.431 min Scan# 3122
Delta R.T. 0.012 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:228 Resp: 1078467
Ion Ratio Lower Upper
228 100
229 19.2 15.7 23.5
226 27.6 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 24.040 ng/ul
RT: 21.343 min Scan# 3107
Delta R.T. 0.012 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:149 Resp: 571694
Ion Ratio Lower Upper
149 100
167 27.8 22.1 33.1
279 3.8 2.9 4.3





#87

Chrysene

Concen: 19.857 ng/ul

RT: 21.490 min Scan# 3132

Delta R.T. 0.006 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :

SSTD020586





Tgt Ion:228 Resp: 1003797

Ion Ratio Lower Upper

228 100

226 31.2 23.4 35.0

229 20.9 14.8 22.2



#89

Di-n-octyl phthalate

Concen: 26.696 ng/ul




RT: 22.477 min Scan# 3300


Delta R.T. 0.012 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Tgt Ion:149 Resp: 984683









#90
Benzo(b)fluoranthene
Concen: 20.916 ng/ul
RT: 23.511 min Scan# 3476
Delta R.T. 0.012 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33




Instrument :
BNA_G
ClientSampleId :
SSTD020586


Tgt Ion:252 Resp: 1015876
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 9.2 7.0 10.4



#91
Benzo(k)fluoranthene
Concen: 22.421 ng/ul
RT: 23.576 min Scan# 3487
Delta R.T. 0.018 min
Lab File: BG063876.D
Acq: 27 Dec 2024 20:33

Tgt Ion:252 Resp: 1086277
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 8.8 6.5 9.7





#93

Benzo(a)pyrene

Concen: 21.297 ng/ul

RT: 24.328 min Scan# 3615

Delta R.T. 0.024 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :

SSTD020586





Tgt Ion:252 Resp: 970939

Ion Ratio Lower Upper

252 100

253 21.8 16.6 24.8

125 10.5 8.0 12.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.362 ng/ul

RT: 27.871 min Scan# 4218

Delta R.T. 0.024 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33




Tgt Ion:276 Resp: 1189354


Ion Ratio Lower Upper

276 100

138 19.0 13.7 20.5

277 24.6 18.4 27.6





#95

Dibenzo(a,h)anthracene

Concen: 21.498 ng/ul

RT: 27.924 min Scan# 41

Delta R.T. 0.036 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Instrument :

BNA_G

ClientSampleId :

SSTD020586





Tgt Ion:278 Resp: 957731

Ion Ratio Lower Upper

278 100

139 14.6 9.7 14.5#

279 23.0 17.3 25.9



#96

Benzo(g,h,i)perylene

Concen: 20.940 ng/ul

RT: 28.940 min Scan# 4400

Delta R.T. 0.042 min

Lab File: BG063876.D

Acq: 27 Dec 2024 20:33

Tgt Ion:276 Resp: 925135

Ion Ratio Lower Upper

276 100

138 17.4 13.8 20.8

277 23.2 17.8 26.8

