

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG122718\
 Data File : BG038850.D
 Acq On : 27 Dec 2018 19:19
 Operator : JU/SJ
 Sample : J6533-16
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 COMP-4

Manual Integrations
 APPROVED

Sohil
 12/28/2018 4:03:24 PM

Quant Time: Dec 28 01:09:52 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.06	152	25095	20.00	ng	0.00
21) Naphthalene-d8	10.87	136	110237	20.00	ng	-0.01
39) Acenaphthene-d10	14.69	164	79231	20.00	ng	0.00
64) Phenanthrene-d10	17.44	188	179935	20.00	ng	0.00
76) Chrysene-d12	21.72	240	188991	20.00	ng	-0.01
87) Perylene-d12	25.02	264	204160	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	91234	64.48	ng	0.00
7) Phenol-d6	7.22	99	97776	50.34	ng	0.01
23) Nitrobenzene-d5	9.24	82	159103	73.73	ng	0.00
42) 2,4,6-Tribromophenol	16.19	330	154296	141.30	ng	0.00
45) 2-Fluorobiphenyl	13.31	172	432898	74.95	ng	0.00
79) Terphenyl-d14	20.04	244	602981	69.44	ng	0.00
Target Compounds						Qvalue
9) Benzaldehyde	7.20	77	12694	10.074	ng	87
10) Phenol	7.24	94	21552	10.444	ng	99
15) Benzyl Alcohol	8.30	79	144695	95.440	ng	99
20) 3+4-Methylphenols	8.83	107	20054	10.503	ng	97
27) 2,4-Dimethylphenol	10.04	122	7005m	4.375	ng	
32) Benzoic acid	10.41	122	345577	277.994	ng	95
50) Dimethylphthalate	14.14	163	40224	6.217	ng	97
66) n-Nitrosodiphenylamine	15.95	169	16029	3.032	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG122718\
Data File : BG038850.D
Acq On : 27 Dec 2018 19:19
Operator : JU/SJ
Sample : J6533-16
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
COMP-4

Quant Time: Dec 28 01:09:52 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 15:26:27 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
12/28/2018 4:03:24 PM

