

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG122718\
 Data File : BG038840.D
 Acq On : 27 Dec 2018 10:38
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 28 00:26:02 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.525	0.424	19.2	66	0.00
3	Pyridine	1.445	1.220	15.6	59	0.00
4	n-Nitrosodimethylamine	0.708	0.607	14.3	66	0.00
5 S	2-Fluorophenol	1.128	1.117	1.0	82	0.00
6	Aniline	1.976	2.030	-2.7	85	0.00
7 S	Phenol-d6	1.548	1.688	-9.0	91	0.00
8	2-Chlorophenol	1.305	1.369	-4.9	87	0.00
9	Benzaldehyde	1.004	0.951	5.3	85	0.00
10 C	Phenol	1.645	1.760	-7.0	89	0.00
11	bis(2-Chloroethyl)ether	1.309	1.262	3.6	82	0.00
12	1,3-Dichlorobenzene	1.543	1.545	-0.1	84	0.00
13 C	1,4-Dichlorobenzene	1.580	1.596	-1.0	86	0.00
14	1,2-Dichlorobenzene	1.490	1.541	-3.4	89	0.00
15	Benzyl Alcohol	1.208	1.361	-12.7	90	0.00
16	2,2'-oxybis(1-Chloropropane	2.999	2.667	11.1	75	0.00
17	2-Methylphenol	1.102	1.221	-10.8	91	0.00
18	Hexachloroethane	0.577	0.545	5.5	80	0.00
19 P	n-Nitroso-di-n-propylamine	1.087	1.159	-6.6	88	0.00
20	3+4-Methylphenols	1.522	1.741	-14.4	94	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
22	Acetophenone	0.500	0.502	-0.4	93	0.00
23 S	Nitrobenzene-d5	0.391	0.373	4.6	88	0.00
24	Nitrobenzene	0.396	0.369	6.8	87	0.00
25	Isophorone	0.703	0.649	7.7	73	0.00
26 C	2-Nitrophenol	0.203	0.209	-3.0	96	0.00
27	2,4-Dimethylphenol	0.291	0.287	1.4	91	0.00
28	bis(2-Chloroethoxy)methane	0.397	0.385	3.0	89	0.00
29 C	2,4-Dichlorophenol	0.323	0.370	-14.6	102	0.00
30	1,2,4-Trichlorobenzene	0.390	0.405	-3.8	97	0.00
31	Naphthalene	0.985	0.992	-0.7	94	0.00
32	Benzoic acid	0.154	0.213	-38.3#	117	0.00
33	4-Chloroaniline	0.428	0.471	-10.0	99	0.00
34 C	Hexachlorobutadiene	0.264	0.272	-3.0	94	0.00
35	Caprolactam	0.134	0.140	-4.5	101	0.01
36 C	4-Chloro-3-methylphenol	0.369	0.400	-8.4	102	0.00
37	2-Methylnaphthalene	0.703	0.757	-7.7	100	0.00
38	1-Methylnaphthalene	0.682	0.733	-7.5	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
40	1,2,4,5-Tetrachlorobenzene	0.748	0.724	3.2	105	0.00
41 P	Hexachlorocyclopentadiene	0.411	0.364	11.4	94	0.00
42 S	2,4,6-Tribromophenol	0.276	0.322	-16.7	125	0.00
43 C	2,4,6-Trichlorophenol	0.460	0.489	-6.3	110	0.00
44	2,4,5-Trichlorophenol	0.487	0.515	-5.7	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG122718\
 Data File : BG038840.D
 Acq On : 27 Dec 2018 10:38
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 28 00:26:02 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.458	1.427	2.1	107	0.00
46	1,1'-Biphenyl	1.677	1.617	3.6	104	0.00
47	2-Chloronaphthalene	1.295	1.252	3.3	106	0.00
48	2-Nitroaniline	0.413	0.399	3.4	102	0.00
49	Acenaphthylene	1.936	1.917	1.0	107	0.00
50	Dimethylphthalate	1.633	1.656	-1.4	109	0.00
51	2,6-Dinitrotoluene	0.370	0.370	0.0	108	0.00
52 C	Acenaphthene	1.158	1.138	1.7	107	0.00
53	3-Nitroaniline	0.377	0.394	-4.5	111	0.00
54 P	2,4-Dinitrophenol	0.193	0.235	-21.8	129	0.00
55	Dibenzofuran	1.985	2.007	-1.1	112	0.00
56 P	4-Nitrophenol	0.286	0.264	7.7	95	0.01
57	2,4-Dinitrotoluene	0.521	0.532	-2.1	108	0.00
58	Fluorene	1.470	1.514	-3.0	111	0.00
59	2,3,4,6-Tetrachlorophenol	0.438	0.481	-9.8	115	0.00
60	Diethylphthalate	1.793	1.747	2.6	107	0.00
61	4-Chlorophenyl-phenylether	0.917	0.952	-3.8	112	0.00
62	4-Nitroaniline	0.388	0.401	-3.4	107	0.00
63	Azobenzene	1.498	1.389	7.3	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
65	4,6-Dinitro-2-methylphenol	0.143	0.141	1.4	110	0.00
66 c	n-Nitrosodiphenylamine	0.588	0.591	-0.5	114	0.00
67	4-Bromophenyl-phenylether	0.244	0.254	-4.1	116	0.00
68	Hexachlorobenzene	0.253	0.270	-6.7	118	0.00
69	Atrazine	0.218	0.205	6.0	103	0.00
70 C	Pentachlorophenol	0.150	0.157	-4.7	114	0.00
71	Phenanthrene	1.037	1.045	-0.8	114	0.00
72	Anthracene	1.024	1.030	-0.6	112	0.00
73	Carbazole	1.013	1.019	-0.6	113	0.00
74	Di-n-butylphthalate	1.162	1.110	4.5	106	0.00
75 C	Fluoranthene	1.283	1.263	1.6	114	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
77	Benzidine	0.591	0.614	-3.9	97	0.00
78	Pyrene	1.321	1.285	2.7	111	0.00
79 S	Terphenyl-d14	0.919	0.941	-2.4	115	0.00
80	Butylbenzylphthalate	0.516	0.493	4.5	104	0.00
81	Benzo(a)anthracene	1.219	1.240	-1.7	112	0.00
82	3,3'-Dichlorobenzidine	0.483	0.498	-3.1	110	0.00
83	Chrysene	1.174	1.170	0.3	111	0.00
84	Bis(2-ethylhexyl)phthalate	0.732	0.692	5.5	102	0.00
85 c	Di-n-octyl phthalate	1.226	1.159	5.5	101	-0.01
86	Indeno(1,2,3-cd)pyrene	1.380	1.404	-1.7	110	0.00
87 I	Perylene-d12	1.000	1.000	0.0	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG122718\
Data File : BG038840.D
Acq On : 27 Dec 2018 10:38
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Dec 28 00:26:02 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 15:26:27 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.098	1.080	1.6	107	0.00
89	Benzo(k)fluoranthene	1.093	1.114	-1.9	110	0.00
90 C	Benzo(a)pyrene	1.059	1.065	-0.6	109	0.00
91	Dibenzo(a,h)anthracene	1.055	1.060	-0.5	109	0.00
92	Benzo(q,h,i)perylene	1.039	1.045	-0.6	109	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0