


Data Path : Z:\HPCHEM1\BNA G\Data\BG122815\
Data File : BG020566.D
Acq On : 28 Dec 2015 11:40
Operator : UM/SJ
Sample : DFTPP17
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP17

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG122415.M
Title : SVOA CALIBRATION
Last Update : Tue Dec 29 05:04:12 2015



AutoFind: Scans 2557, 2558, 2559; Background Corrected with Scan 2552

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.1	9769	PASS
68	69	0.00	2	0.7	75	PASS
69	198	0.00	100	30.3	11370	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	41.1	15425	PASS
197	198	0.00	2	0.4	155	PASS
198	198	100	100	100.0	37496	PASS
199	198	5	9	6.7	2499	PASS
275	198	10	60	30.4	11399	PASS
365	198	1	100	3.9	1473	PASS
441	442	0.01	100	15.8	5220	PASS
442	198	50	100	88.3	33109	PASS
443	442	15	24	19.9	6580	PASS

Scan 2552 (17.833 min): BG020566.D
DFTPP17

m/z	Abundance
39.90	177.0
44.00	172.0
48.90	632.0
50.90	165.0
83.90	478.0
85.90	439.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

m/z	Abundance
39.00	669.0
40.00	214.0
43.90	263.0
49.00	644.0
50.00	2384.0
51.00	10016.0
52.00	424.0
56.00	274.0
57.00	771.0
62.00	154.0
63.00	571.0
65.00	273.0
68.90	10643.0
73.00	150.0
74.00	1069.0
75.00	1945.0
76.00	692.0
77.00	12376.0
78.00	880.0
79.00	891.0
79.90	609.0
81.00	898.0
81.90	296.0
83.00	316.0
83.90	458.0
85.00	211.0
85.90	481.0
92.00	280.0
93.00	1524.0
97.90	1085.0
99.00	958.0
101.00	597.0
104.00	259.0
105.00	283.0
106.00	218.0
107.00	3844.0
108.00	827.0
109.90	6958.0
110.90	1004.0
115.90	157.0
116.90	3244.0
117.90	237.0
121.90	269.0
122.90	549.0
123.90	150.0
125.00	242.0
127.00	14066.0
128.00	1184.0
128.90	5848.0
130.00	354.0
134.00	191.0
134.90	525.0
136.90	324.0
140.90	813.0
142.00	187.0
142.90	247.0
145.90	166.0
146.90	387.0
147.90	808.0
152.90	249.0
154.00	170.0
154.90	445.0
156.00	718.0
157.00	214.0
157.90	235.0
159.90	305.0
161.00	389.0
164.90	279.0
165.90	252.0
166.90	1561.0
167.90	612.0
173.00	211.0
174.00	610.0
175.00	710.0
176.90	321.0
178.90	1333.0
180.00	967.0
180.90	417.0
184.90	558.0
186.00	4161.0
186.90	1276.0
188.00	215.0
188.90	287.0
190.90	157.0
191.90	387.0
193.00	501.0
194.00	166.0
195.90	1108.0
197.90	32552.0
198.90	2422.0
199.90	212.0
202.90	226.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

203.90	1520.0
204.90	2061.0
206.00	9736.0
206.90	1365.0
207.90	428.0
209.90	209.0
210.90	350.0
215.90	240.0
216.90	2681.0
217.90	377.0
220.90	1504.0
221.80	460.0
222.90	599.0
224.00	5352.0
225.00	1350.0
227.00	2432.0
227.90	310.0
228.90	522.0
235.00	214.0
236.90	205.0
242.00	348.0
242.90	374.0
244.00	3609.0
244.90	553.0
246.00	817.0
254.90	20560.0
255.90	2907.0
257.00	176.0
257.90	1102.0
259.00	180.0
264.90	377.0
272.90	596.0
273.90	1325.0
274.90	9214.0
275.90	1236.0
276.90	728.0
292.90	184.0
295.90	2312.0
296.90	249.0
303.00	222.0
314.90	278.0
322.90	674.0
323.90	158.0
326.90	191.0
334.00	478.0
345.70	164.0
352.00	299.0
352.90	205.0
353.90	231.0
364.90	1191.0
365.90	204.0
372.00	378.0
402.90	166.0
421.00	242.0
422.00	154.0
423.00	1130.0
424.00	241.0
441.00	3840.0
442.00	21816.0
443.00	4865.0
443.90	491.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

m/z	Abundance
39.00	846.0
39.90	158.0
43.90	308.0
49.00	589.0
50.00	3076.0
51.00	12046.0
52.00	597.0
55.00	172.0
55.90	354.0
57.00	946.0
61.90	233.0
63.00	451.0
65.00	284.0
67.90	225.0
68.90	13898.0
73.00	153.0
74.00	1475.0
75.00	2273.0
76.00	691.0
77.00	14956.0
78.00	917.0
79.00	1098.0
80.00	722.0
81.00	1318.0
81.90	223.0
82.90	376.0
83.90	351.0
85.00	266.0
85.90	810.0
90.90	212.0
92.00	320.0
93.00	1639.0
98.00	1266.0
98.90	1039.0
101.00	498.0
102.90	297.0
103.90	488.0
104.90	441.0
107.00	5116.0
107.90	807.0
109.00	216.0
110.00	8385.0
110.90	1287.0
111.90	177.0
116.00	324.0
116.90	3501.0
118.00	242.0
121.90	236.0
123.00	646.0
124.00	278.0
125.00	189.0
127.00	18400.0
128.00	1549.0
129.00	6907.0
130.00	665.0
133.90	158.0
135.00	432.0
136.00	222.0
136.90	303.0
139.90	213.0
140.90	746.0
142.00	301.0
142.90	217.0
146.90	475.0
147.90	909.0
148.90	186.0
151.70	154.0
152.90	293.0
154.00	223.0
154.90	598.0
156.00	898.0
157.90	210.0
159.00	159.0
159.90	316.0
161.00	530.0
164.90	473.0
166.00	309.0
166.90	1762.0
167.90	634.0
168.90	151.0
172.90	277.0
174.00	374.0
174.90	996.0
176.00	279.0
176.90	526.0
178.90	1663.0
179.90	1136.0
181.00	632.0
184.90	1026.0
186.00	6380.0
187.00	1641.0
187.90	177.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

188.90	487.0
190.90	203.0
191.90	615.0
193.00	644.0
195.90	1274.0
196.70	466.0
197.90	44288.0
198.90	3017.0
199.80	240.0
202.90	451.0
203.90	1561.0
204.90	2985.0
206.00	12864.0
206.90	1634.0
207.90	483.0
208.90	159.0
209.80	175.0
210.90	498.0
214.90	170.0
215.90	266.0
216.90	3459.0
217.90	447.0
220.90	2198.0
222.90	877.0
224.00	7781.0
224.90	1862.0
226.00	212.0
226.90	3154.0
227.90	513.0
228.90	791.0
230.90	313.0
233.90	174.0
234.90	175.0
235.90	154.0
236.90	250.0
238.90	153.0
242.00	371.0
242.90	505.0
244.00	5522.0
245.00	702.0
245.90	1019.0
246.90	267.0
248.90	227.0
254.90	29432.0
255.90	4668.0
256.90	349.0
257.90	1495.0
258.90	284.0
264.90	782.0
272.90	927.0
273.90	2099.0
274.90	13290.0
276.00	1623.0
277.00	1084.0
277.90	185.0
285.00	156.0
292.90	287.0
295.90	3329.0
296.90	615.0
302.90	290.0
313.90	222.0
314.90	383.0
316.00	221.0
323.00	1240.0
324.00	217.0
326.80	255.0
334.00	912.0
335.00	232.0
345.90	229.0
352.00	381.0
352.90	237.0
353.90	299.0
364.90	1683.0
365.90	239.0
371.90	645.0
372.90	164.0
382.90	214.0
401.90	156.0
403.00	220.0
420.90	209.0
421.90	304.0
423.00	2131.0
423.90	396.0
441.00	5670.0
442.00	37800.0
443.00	7625.0
444.00	734.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

m/z	Abundance
39.00	526.0
39.90	166.0
41.10	150.0
43.00	189.0
43.90	229.0
48.90	539.0
50.00	2009.0
51.00	7740.0
52.00	474.0
55.90	369.0
57.00	547.0
63.00	428.0
68.90	9570.0
74.00	1119.0
74.90	1618.0
76.00	521.0
77.00	10522.0
78.00	699.0
79.00	862.0
79.90	455.0
80.90	850.0
82.00	272.0
82.90	232.0
83.90	438.0
85.90	621.0
90.90	154.0
91.90	248.0
92.90	1150.0
97.90	820.0
98.90	842.0
100.90	522.0
103.00	171.0
103.90	379.0
104.90	306.0
107.00	3965.0
107.90	593.0
109.90	6252.0
110.90	912.0
115.90	296.0
116.90	2840.0
117.90	285.0
122.00	206.0
122.90	409.0
124.00	212.0
125.00	155.0
127.00	13811.0
128.00	978.0
129.00	5139.0
129.90	534.0
133.90	212.0
135.00	421.0
135.90	172.0
137.00	198.0
139.90	161.0
140.90	785.0
146.90	343.0
147.90	613.0
148.90	205.0
150.90	155.0
152.90	197.0
153.90	258.0
155.00	463.0
155.90	611.0
157.80	173.0
159.90	252.0
160.90	428.0
164.90	354.0
165.90	412.0
166.90	1686.0
167.90	702.0
171.90	159.0
172.90	224.0
174.10	233.0
174.90	604.0
175.90	282.0
176.90	338.0
178.90	1405.0
179.90	1047.0
181.00	468.0
184.90	700.0
186.00	5139.0
187.00	1447.0
187.90	202.0
188.90	285.0
191.90	339.0
193.00	407.0
195.90	927.0
197.90	35648.0
198.90	2059.0
199.90	172.0
201.30	189.0
202.90	258.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17

204.00	1309.0
205.00	2500.0
206.00	10159.0
207.00	1328.0
207.90	302.0
210.00	181.0
211.00	384.0
216.00	287.0
216.90	2677.0
217.90	351.0
220.90	1452.0
221.50	432.0
222.90	857.0
224.00	5581.0
224.90	1480.0
226.00	190.0
226.90	2654.0
227.90	368.0
229.00	547.0
230.90	231.0
233.90	191.0
234.90	202.0
236.90	177.0
240.90	162.0
242.00	365.0
243.00	361.0
244.00	4545.0
244.90	614.0
245.90	1110.0
246.90	159.0
248.90	230.0
253.90	252.0
254.90	25312.0
256.00	3536.0
257.00	253.0
257.90	1430.0
258.90	209.0
264.90	695.0
265.90	154.0
272.90	750.0
273.90	2039.0
274.90	11694.0
275.90	1650.0
276.90	859.0
284.90	162.0
292.90	213.0
295.90	3349.0
296.90	428.0
303.00	351.0
314.00	182.0
314.90	334.0
316.00	241.0
323.00	1205.0
323.90	207.0
326.90	183.0
333.90	785.0
334.90	163.0
345.80	231.0
352.00	478.0
353.00	312.0
354.00	280.0
364.90	1545.0
365.90	153.0
372.00	488.0
401.90	167.0
402.90	328.0
421.00	324.0
421.90	283.0
423.00	2040.0
423.90	370.0
441.00	6150.0
442.00	39712.0
443.00	7251.0
444.00	819.0

Instrument :
BNA_G
ClientSampleId :
DFTPP17