

Data Path : Z:\HPCHEM1\BNA G\DATA\BG123015\
 Data File : BG020611.D
 Acq On : 30 Dec 2015 16:35
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02006

Quant Time: Dec 31 01:01:25 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG123015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 30 23:49:10 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.422	0.461	-9.2	123	0.00
3 S	1,4-Dioxane-d8	0.334	0.353	-5.7	100	0.00
4	Benzaldehyde	1.035	1.190	-15.0	112	0.00
5 S	Phenol-d5	1.689	1.827	-8.2	117	0.00
6	Phenol	1.820	1.908	-4.8	111	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.022	1.076	-5.3	112	0.00
8	Bis(2-Chloroethyl)ether	1.298	1.381	-6.4	111	0.00
9 S	2-Chlorophenol-d4	1.292	1.426	-10.4	115	0.00
10	2-Chlorophenol	1.350	1.464	-8.4	115	0.00
11	2-Methylphenol	1.402	1.458	-4.0	111	0.00
12	2,2'-oxybis(1-Chloropropane	1.697	1.860	-9.6	114	0.00
13 S	4-Methylphenol-d8	1.449	1.569	-8.3	115	0.00
14	Acetophenone	2.375	2.548	-7.3	113	0.00
15 P	N-Nitroso-di-n-propylamine	1.199	1.303	-8.7	111	0.00
16	4-Methylphenol	1.551	1.658	-6.9	114	0.00
17	Hexachloroethane	0.567	0.601	-6.0	112	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
19 S	Nitrobenzene-d5	0.155	0.165	-6.5	111	0.00
20	Nitrobenzene	0.386	0.431	-11.7	116	0.00
21	Isophorone	0.751	0.821	-9.3	110	0.00
22 S	2-Nitrophenol-d4	0.179	0.198	-10.6	110	0.00
23 C	2-Nitrophenol	0.189	0.210	-11.1	118	0.00
24	2,4-Dimethylphenol	0.398	0.445	-11.8	114	0.00
25	Bis(2-Chloroethoxy)methane	0.415	0.446	-7.5	110	0.00
26 S	2,4-Dichlorophenol-d3	0.341	0.379	-11.1	112	0.00
27 C	2,4-Dichlorophenol	0.352	0.381	-8.2	111	0.00
28	Naphthalene	1.059	1.148	-8.4	112	0.00
29 S	4-Chloroaniline-d4	0.405	0.465	-14.8	113	0.00
30	4-Chloroaniline	0.416	0.457	-9.9	108	0.00
31 C	Hexachlorobutadiene	0.284	0.306	-7.7	113	0.00
32	Caprolactam	0.121	0.133	-9.9	117	0.00
33 C	4-Chloro-3-methylphenol	0.381	0.428	-12.3	115	0.00
34	2-Methylnaphthalene	0.793	0.867	-9.3	113	0.00
35	1-Methylnaphthalene	0.765	0.826	-8.0	111	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
37	1,2,4,5-Tetrachlorobenzene	0.770	0.823	-6.9	113	0.00
38	Hexachlorocyclopentadiene	0.277	0.227	18.1	111	0.00
39 C	2,4,6-Trichlorophenol	0.431	0.464	-7.7	111	0.00
40	2,4,5-Trichlorophenol	0.473	0.507	-7.2	108	0.00
41	1,1'-Biphenyl	1.505	1.594	-5.9	111	0.00
42	2-Chloronaphthalene	1.211	1.293	-6.8	112	0.00
43	2-Nitroaniline	0.348	0.387	-11.2	112	0.00
44 S	Dimethylphthalate-d6	1.637	1.745	-6.6	111	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG123015\
 Data File : BG020611.D
 Acq On : 30 Dec 2015 16:35
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02006

Quant Time: Dec 31 01:01:25 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG123015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 30 23:49:10 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.672	1.778	-6.3	111	0.00
46	2,6-Dinitrotoluene	0.347	0.372	-7.2	110	0.00
47 S	Acenaphthylene-d8	1.859	1.982	-6.6	111	0.00
48	Acenaphthylene	1.980	2.099	-6.0	111	0.00
49	3-Nitroaniline	0.316	0.348	-10.1	112	0.00
50 C	Acenaphthene	1.340	1.415	-5.6	111	0.00
51	2,4-Dinitrophenol	0.146	0.140	4.1	126	0.00
52 S	4-Nitrophenol-d4	0.257	0.268	-4.3	110	0.00
53	4-Nitrophenol	0.227	0.239	-5.3	113	0.00
54	Dibenzofuran	2.025	2.136	-5.5	110	0.00
55	2,4-Dinitrotoluene	0.506	0.553	-9.3	111	0.00
56	2,3,4,6-Tetrachlorophenol	0.441	0.488	-10.7	111	0.00
57	Diethylphthalate	1.723	1.815	-5.3	111	0.00
58 S	Fluorene-d10	1.484	1.554	-4.7	111	0.00
59	Fluorene	1.629	1.729	-6.1	110	0.00
60	4-Chlorophenyl-phenylether	0.941	0.994	-5.6	109	0.00
61	4-Nitroaniline	0.345	0.372	-7.8	111	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.121	0.126	-4.1	116	0.00
64	4,6-Dinitro-2-methylphenol	0.130	0.137	-5.4	113	0.00
65	N-Nitrosodiphenylamine	0.551	0.594	-7.8	110	0.00
66	4-Bromophenyl-phenylether	0.243	0.257	-5.8	109	0.00
67	Hexachlorobenzene	0.267	0.287	-7.5	111	0.00
68	Atrazine	0.243	0.259	-6.6	108	0.00
69 C	Pentachlorophenol	0.107	0.108	-0.9	116	0.00
70	Phenanthrene	1.087	1.159	-6.6	109	0.00
71 S	Anthracene-d10	0.932	0.997	-7.0	110	0.00
72	Anthracene	1.112	1.195	-7.5	110	0.00
73	1,2,3,4-Tetrachlorobenzene	0.318	0.338	-6.3	112	0.00
74	Pentachlorobenzene	0.303	0.327	-7.9	111	0.00
75	Carbazole	0.938	1.006	-7.2	107	0.00
76	Di-n-butylphthalate	1.101	1.189	-8.0	109	0.00
77 C	Fluoranthene	1.328	1.446	-8.9	107	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
79 S	Pyrene-d10	0.897	0.964	-7.5	105	0.00
80	Pyrene	1.158	1.261	-8.9	106	0.00
81	Butylbenzylphthalate	0.408	0.438	-7.4	107	0.00
82	3,3'-Dichlorobenzidine	0.381	0.418	-9.7	106	0.00
83	Benzo(a)anthracene	1.170	1.268	-8.4	108	0.00
84	Bis(2-ethylhexyl)phthalate	0.586	0.640	-9.2	108	0.00
85	Chrysene	1.101	1.199	-8.9	108	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Di-n-octyl phthalate	1.050	1.133	-7.9	108	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG123015\
Data File : BG020611.D
Acq On : 30 Dec 2015 16:35
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02006

Quant Time: Dec 31 01:01:25 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG123015.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 30 23:49:10 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.188	1.294	-8.9	107	0.00
89	Benzo(k)fluoranthene	1.138	1.235	-8.5	108	0.00
90 S	Benzo(a)pyrene-d12	0.998	1.089	-9.1	109	0.00
91 C	Benzo(a)pyrene	1.141	1.223	-7.2	107	0.00
92	Indeno(1,2,3-cd)pyrene	1.361	1.475	-8.4	108	-0.01
93	Dibenzo(a,h)anthracene	1.135	1.237	-9.0	108	0.00
94	Benzo(a,h,i)perylene	1.118	1.208	-8.1	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0