


Data Path : Z:\HPCHEM1\BNA G\Data\BG123015\
 Data File : BG020591.D
 Acq On : 30 Dec 2015 3:15
 Operator : UM/SJ
 Sample : DFTPP18
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP18

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG123015.M
 Title : SVOA CALIBRATION
 Last Update : Thu Dec 31 03:43:50 2015



AutoFind: Scans 2556, 2557, 2558; Background Corrected with Scan 2550

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.1	12366	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	29.2	14423	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	39.2	19326	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	49341	PASS
199	198	5	9	7.1	3489	PASS
275	198	10	60	30.9	15239	PASS
365	198	1	100	3.8	1885	PASS
441	442	0.01	100	15.9	7502	PASS
442	198	50	100	95.6	47184	PASS
443	442	15	24	18.3	8653	PASS

Scan 2550 (17.821 min): BG020591.D
DFTPP18

m/z	Abundance
39.90	188.0
43.90	180.0
289.00	203.0
290.10	348.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

m/z	Abundance
38.00	222.0
39.00	1194.0
43.90	171.0
50.00	4053.0
51.00	15859.0
52.00	888.0
55.90	494.0
56.90	1355.0
61.00	169.0
62.00	258.0
63.00	690.0
64.00	185.0
65.00	343.0
69.00	18072.0
74.00	1847.0
75.00	2779.0
76.00	1002.0
77.00	20568.0
78.00	1381.0
79.00	1250.0
80.00	1102.0
81.00	1483.0
81.90	446.0
83.00	320.0
84.90	241.0
85.90	442.0
87.00	221.0
91.00	328.0
92.00	394.0
92.90	2434.0
94.00	211.0
98.00	1941.0
99.00	1164.0
99.90	170.0
100.90	829.0
102.90	230.0
103.90	569.0
105.00	500.0
107.00	6784.0
107.90	954.0
108.90	255.0
110.00	11380.0
111.00	1963.0
112.00	178.0
116.00	326.0
117.00	4648.0
118.00	363.0
119.90	166.0
121.90	443.0
123.00	746.0
123.90	334.0
125.00	275.0
127.00	22664.0
127.90	1679.0
128.90	8784.0
130.00	753.0
130.80	203.0
134.00	227.0
134.90	841.0
136.00	404.0
137.00	400.0
141.00	1264.0
141.90	360.0
142.90	264.0
146.00	270.0
146.90	637.0
147.90	1398.0
149.00	255.0
150.90	183.0
152.90	554.0
154.00	471.0
155.00	802.0
156.00	1169.0
157.90	231.0
158.90	221.0
159.90	530.0
160.90	692.0
161.90	209.0
164.90	616.0
165.90	383.0
167.00	2431.0
167.90	1067.0
168.90	241.0
171.90	265.0
172.90	300.0
174.00	687.0
175.00	1154.0
175.90	214.0
176.90	566.0
177.90	262.0
178.90	2031.0
180.00	1529.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

180.90	735.0
183.90	223.0
185.00	1174.0
186.00	7989.0
186.90	2170.0
187.90	215.0
188.90	498.0
190.90	246.0
192.00	726.0
193.00	799.0
194.00	178.0
195.90	1571.0
197.90	54568.0
198.90	4006.0
199.90	394.0
201.70	213.0
202.90	557.0
203.90	2449.0
204.90	3294.0
206.00	15930.0
207.00	2063.0
208.00	593.0
208.90	185.0
209.90	169.0
210.90	509.0
214.90	213.0
215.90	463.0
216.90	4277.0
217.90	562.0
220.90	2549.0
221.70	630.0
222.90	1140.0
224.00	8706.0
225.00	2264.0
225.90	270.0
226.90	3819.0
228.00	661.0
228.90	885.0
230.00	161.0
230.90	348.0
233.90	302.0
234.90	304.0
235.80	193.0
236.90	371.0
239.80	190.0
240.90	152.0
242.00	468.0
243.00	560.0
244.00	6481.0
244.90	981.0
245.90	1326.0
247.00	306.0
249.00	295.0
253.00	177.0
254.90	36280.0
255.90	5100.0
256.90	416.0
258.00	1713.0
258.90	287.0
264.90	781.0
272.90	1302.0
273.90	2855.0
275.00	15509.0
276.00	2267.0
276.90	1541.0
278.00	282.0
282.90	230.0
285.00	362.0
288.90	214.0
289.90	295.0
292.90	351.0
295.90	4987.0
296.90	408.0
302.90	370.0
310.00	171.0
314.90	525.0
315.90	203.0
320.80	192.0
323.00	1521.0
324.00	211.0
326.90	324.0
327.90	159.0
332.00	157.0
334.00	922.0
335.00	271.0
340.90	193.0
345.90	336.0
352.00	447.0
352.90	411.0
354.00	387.0
365.00	1903.0
365.90	242.0
371.90	639.0
372.90	341.0
382.90	213.0
402.00	213.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

402.90	298.0
420.90	340.0
421.90	353.0
423.00	2221.0
424.00	483.0
441.00	6596.0
442.00	41744.0
443.00	7554.0
444.00	771.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

m/z	Abundance
37.90	189.0
39.00	1125.0
44.00	250.0
49.00	167.0
50.00	3665.0
51.00	14074.0
52.00	803.0
55.90	401.0
56.90	934.0
60.90	203.0
62.10	306.0
63.00	787.0
65.00	340.0
68.90	16576.0
73.00	187.0
74.00	1793.0
75.00	2769.0
76.00	919.0
77.00	19696.0
78.00	1283.0
78.90	1231.0
79.90	985.0
80.90	1268.0
81.90	375.0
83.00	358.0
84.90	302.0
86.00	428.0
90.90	452.0
92.00	398.0
92.90	2279.0
97.90	1810.0
98.90	1309.0
101.00	847.0
102.90	308.0
103.90	538.0
104.90	501.0
107.00	6448.0
107.90	871.0
108.90	163.0
109.90	10833.0
110.90	1656.0
111.90	252.0
115.90	298.0
116.90	4502.0
118.00	457.0
121.90	484.0
122.90	737.0
123.90	335.0
125.00	279.0
127.00	22256.0
128.00	1660.0
128.90	8952.0
129.90	809.0
131.00	235.0
133.90	276.0
135.00	843.0
135.90	326.0
137.00	503.0
141.00	1223.0
141.90	409.0
143.00	242.0
146.00	161.0
147.00	719.0
147.90	1407.0
148.90	221.0
152.90	417.0
153.90	380.0
155.00	669.0
155.90	1146.0
156.90	266.0
157.80	254.0
158.90	187.0
159.90	531.0
160.90	664.0
161.90	197.0
164.90	627.0
165.90	333.0
167.00	2469.0
168.00	1034.0
169.00	227.0
171.80	287.0
172.90	303.0
174.00	663.0
175.00	934.0
176.00	287.0
176.90	474.0
178.00	237.0
178.90	2112.0
179.90	1666.0
180.90	507.0
183.90	202.0
184.90	996.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

186.00	7994.0
187.00	2344.0
187.90	248.0
188.90	507.0
190.90	223.0
192.00	780.0
193.00	703.0
193.90	163.0
196.00	1653.0
197.90	57608.0
198.90	4199.0
199.90	336.0
201.40	235.0
202.90	493.0
204.00	2506.0
205.00	4068.0
206.00	16528.0
207.00	1935.0
207.90	684.0
208.90	318.0
210.00	295.0
211.00	581.0
211.90	155.0
215.00	221.0
215.90	424.0
216.90	4701.0
217.90	541.0
220.90	2953.0
221.80	567.0
222.90	1258.0
224.00	9387.0
224.90	2267.0
226.00	273.0
226.90	4073.0
227.90	634.0
229.00	843.0
231.00	373.0
233.90	273.0
234.90	360.0
236.90	312.0
238.80	191.0
240.80	276.0
241.90	527.0
242.90	605.0
244.00	7595.0
244.90	980.0
245.90	1426.0
246.90	280.0
248.90	204.0
252.90	203.0
254.90	39704.0
255.90	5892.0
256.90	629.0
257.90	1998.0
258.90	500.0
264.90	903.0
272.00	166.0
273.00	1431.0
274.00	3395.0
274.90	18856.0
275.90	2277.0
276.90	1392.0
277.90	251.0
282.90	242.0
285.00	225.0
289.90	157.0
292.90	354.0
295.00	209.0
295.90	4705.0
296.90	758.0
302.90	661.0
304.00	191.0
314.00	282.0
314.90	599.0
316.00	370.0
321.10	172.0
323.00	1677.0
324.00	283.0
326.90	192.0
327.90	151.0
333.00	284.0
334.00	1204.0
335.00	320.0
340.90	196.0
345.90	385.0
352.00	447.0
353.00	305.0
354.00	569.0
364.90	2220.0
365.90	379.0
370.90	183.0
372.00	865.0
372.90	194.0
382.90	265.0
390.00	186.0
401.90	433.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

402.90	538.0
403.90	224.0
421.00	338.0
421.90	478.0
423.00	3117.0
424.00	699.0
441.00	9033.0
442.00	57680.0
443.00	10944.0
444.00	877.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

m/z	Abundance
39.00	506.0
39.90	186.0
44.00	258.0
50.00	1595.0
51.00	7165.0
52.00	331.0
56.00	218.0
57.00	550.0
63.00	423.0
68.90	8621.0
74.00	793.0
74.90	1268.0
76.10	450.0
77.00	10151.0
78.00	848.0
78.90	747.0
80.00	531.0
80.90	732.0
81.90	168.0
82.90	183.0
84.90	162.0
85.90	169.0
86.90	196.0
91.00	213.0
91.90	237.0
92.90	1351.0
98.00	866.0
98.90	873.0
101.00	338.0
103.00	198.0
103.90	235.0
105.00	289.0
107.00	3332.0
108.00	533.0
110.00	6158.0
111.00	919.0
115.90	159.0
116.90	2833.0
118.00	263.0
121.80	246.0
123.00	292.0
123.90	189.0
127.00	13059.0
127.90	820.0
128.90	4486.0
129.90	513.0
134.90	476.0
135.90	236.0
136.90	243.0
140.90	612.0
142.00	163.0
145.90	195.0
146.90	337.0
147.90	824.0
148.90	202.0
153.00	365.0
154.00	196.0
155.00	481.0
156.00	554.0
158.80	150.0
160.00	183.0
161.00	387.0
164.90	200.0
165.90	308.0
167.00	1665.0
167.90	730.0
174.00	286.0
175.00	683.0
175.90	201.0
176.90	292.0
178.90	1189.0
180.00	680.0
180.90	309.0
184.00	168.0
184.90	715.0
186.00	4793.0
187.00	1503.0
187.90	226.0
188.90	305.0
190.90	167.0
191.90	503.0
192.90	350.0
193.80	155.0
195.90	1043.0
197.90	35848.0
198.90	2264.0
199.90	228.0
202.90	393.0
204.00	1583.0
204.90	2533.0
206.00	10160.0
206.90	1200.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

208.00	311.0
209.10	190.0
209.90	225.0
210.90	465.0
215.00	212.0
216.00	251.0
216.90	2513.0
217.90	401.0
221.00	1458.0
221.90	336.0
222.90	643.0
224.00	5821.0
224.90	1467.0
225.90	188.0
226.90	2593.0
227.90	264.0
228.90	410.0
231.00	262.0
234.90	201.0
236.90	204.0
242.00	344.0
242.90	462.0
244.00	4665.0
245.00	580.0
245.90	993.0
246.90	167.0
248.90	166.0
254.90	25984.0
255.90	3585.0
256.90	309.0
258.00	1450.0
259.00	299.0
265.00	653.0
272.90	939.0
273.90	1951.0
274.90	11352.0
275.90	1609.0
276.90	1000.0
278.00	172.0
283.00	204.0
285.00	231.0
292.90	226.0
295.90	3613.0
296.90	462.0
302.90	361.0
314.90	444.0
316.00	182.0
321.10	170.0
323.00	1092.0
324.00	166.0
326.90	208.0
332.90	196.0
334.00	714.0
335.00	200.0
346.00	187.0
351.90	338.0
352.90	197.0
354.00	482.0
364.90	1533.0
365.90	215.0
372.00	565.0
382.90	230.0
402.00	189.0
402.90	313.0
404.00	173.0
420.90	285.0
422.00	362.0
423.00	2214.0
423.90	415.0
441.00	6877.0
442.00	42128.0
443.00	7461.0
443.90	660.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18