



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG123021\
 Data File : BG051851.D
 Acq On : 30 Dec 2021 16:31
 Operator : CG/JU
 Sample : DFTPP447
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP447

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
 Title : SVOA CALIBRATION
 Last Update : Fri Dec 31 00:13:36 2021



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.6	37175	PASS
68	69	0.00	2	0.6	240	PASS
69	198	0.00	100	47.5	37901	PASS
70	69	0.00	2	0.9	349	PASS
127	198	10	80	48.1	38381	PASS
197	198	0.00	2	0.5	422	PASS
198	198	100	100	100.0	79736	PASS
199	198	5	9	7.5	5985	PASS
275	198	10	60	32.8	26165	PASS
365	198	1	100	4.4	3492	PASS
441	198	0.01	100	13.1	10435	PASS
442	198	50	100	80.5	64208	PASS
443	442	15	24	19.7	12632	PASS

m/z	Abundance
36.95	213.0
38.15	316.0
39.15	2263.0
40.05	2472.0
44.05	1113.0
47.05	222.0
49.05	1431.0
50.15	8107.0
51.15	30432.0
52.05	1830.0
55.15	331.0
56.15	803.0
57.05	2158.0
61.05	383.0
62.05	498.0
63.15	1080.0
64.15	331.0
65.05	676.0
69.05	30600.0
70.15	364.0
72.95	503.0
74.15	3146.0
75.15	5235.0
76.15	1874.0
77.15	32120.0
78.15	2310.0
79.05	2355.0
80.05	1970.0
81.05	2688.0
82.05	637.0
83.05	506.0
84.05	859.0
85.05	622.0
85.95	1343.0
87.05	228.0
91.05	755.0
92.15	665.0
93.05	4910.0
94.15	210.0
96.05	339.0
98.05	3810.0
99.05	3053.0
100.05	220.0
101.05	1443.0
102.15	163.0
103.05	442.0
104.05	799.0
105.05	1182.0
106.05	410.0
107.05	11669.0
108.05	1817.0
109.05	363.0
110.05	16808.0
111.05	2975.0
112.15	271.0
112.95	153.0
116.05	864.0
117.05	11368.0
118.05	693.0
120.15	196.0
122.15	765.0
123.05	1344.0
124.05	600.0
125.05	452.0
125.85	309.0
127.05	32640.0
128.05	2363.0
129.05	15797.0
130.05	1220.0
131.05	314.0
132.05	156.0
133.85	639.0
135.05	1243.0
136.05	672.0
137.05	883.0
137.85	157.0
141.05	2133.0
142.05	678.0
143.05	723.0
144.05	230.0
145.95	625.0
147.05	972.0
148.05	2607.0
149.05	515.0
151.15	349.0
151.85	189.0
152.15	216.0
153.05	896.0
153.95	455.0
155.05	1310.0
156.05	1543.0
156.95	324.0

Instrument :
BNA_G
ClientSampleId :
DFTPP447

157.85	595.0
160.05	768.0
161.15	993.0
161.95	163.0
165.05	942.0
166.05	627.0
167.05	4510.0
168.05	1908.0
169.15	306.0
170.85	339.0
172.05	359.0
173.05	646.0
173.95	828.0
175.05	1784.0
175.95	440.0
176.95	844.0
178.95	3409.0
180.05	2106.0
181.05	1005.0
181.95	152.0
184.15	257.0
185.05	1835.0
186.05	10390.0
187.05	3346.0
188.05	430.0
189.05	1030.0
190.05	215.0
191.15	557.0
191.95	904.0
193.05	1543.0
193.85	192.0
196.05	2122.0
197.95	71520.0
198.95	5475.0
200.05	566.0
201.55	268.0
203.05	603.0
204.05	2928.0
205.05	5354.0
206.05	21496.0
207.05	3520.0
208.05	678.0
208.95	519.0
210.25	299.0
211.05	1182.0
211.85	190.0
215.05	228.0
216.05	605.0
216.95	7033.0
217.95	920.0
221.05	2597.0
221.95	942.0
223.05	1726.0
224.05	14861.0
225.05	3286.0
225.95	373.0
227.05	6396.0
227.95	1057.0
228.95	1131.0
231.15	494.0
232.95	168.0
233.95	422.0
235.05	301.0
236.05	306.0
236.85	443.0
239.05	399.0
240.95	436.0
241.95	790.0
243.15	794.0
244.05	10929.0
244.95	865.0
245.95	2442.0
246.95	417.0
248.95	376.0
251.95	251.0
253.05	185.0
254.95	57296.0
256.05	8151.0
257.05	910.0
258.05	3714.0
258.95	806.0
263.15	150.0
263.75	168.0
264.95	1103.0
265.95	243.0
272.05	288.0
272.95	2024.0
274.05	4481.0
275.05	26104.0
275.95	2770.0
276.95	2225.0
277.95	362.0
283.05	163.0
285.15	435.0
292.05	156.0
292.85	499.0

Instrument :
BNA_G
ClientSampleId :
DFTPP447

294.85	188.0
296.05	8241.0
296.95	1214.0
300.85	199.0
302.95	850.0
304.05	319.0
313.95	318.0
315.05	885.0
316.05	551.0
320.95	221.0
323.05	2744.0
323.95	474.0
327.05	279.0
327.85	154.0
331.95	228.0
332.85	289.0
334.05	1404.0
334.85	494.0
340.85	518.0
345.95	738.0
351.95	835.0
352.95	533.0
354.15	859.0
355.15	254.0
359.65	151.0
364.95	3594.0
365.85	644.0
371.25	192.0
371.95	1505.0
372.95	388.0
383.05	344.0
401.85	845.0
402.85	620.0
420.85	547.0
422.05	785.0
422.95	4646.0
423.95	636.0
441.05	11508.0
441.95	71016.0
442.95	14055.0
444.05	1197.0

Instrument :
BNA_G
ClientSampleId :
DFTPP447

m/z	Abundance
35.25	168.0
36.95	217.0
38.15	685.0
39.15	3727.0
40.05	2629.0
43.05	151.0
44.05	856.0
47.05	374.0
48.15	208.0
49.05	1696.0
50.15	12825.0
51.15	44528.0
52.15	1980.0
55.05	301.0
56.15	1284.0
57.15	3350.0
58.05	227.0
61.15	521.0
62.05	663.0
63.15	2313.0
64.05	213.0
65.15	961.0
69.05	44480.0
69.95	308.0
73.15	394.0
74.15	4908.0
75.05	7433.0
76.15	2328.0
77.15	47200.0
78.05	2726.0
79.15	3886.0
80.05	3307.0
81.05	3974.0
82.05	1058.0
83.05	697.0
84.05	1010.0
85.15	486.0
86.05	1672.0
87.15	531.0
91.05	888.0
92.15	1238.0
93.05	7111.0
94.05	394.0
94.95	251.0
98.05	5891.0
99.05	4116.0
100.05	438.0
101.15	2050.0
103.15	263.0
103.95	1153.0
105.05	1659.0
106.05	477.0
107.05	16131.0
108.05	2400.0
109.05	842.0
110.05	23464.0
111.05	3914.0
112.05	559.0
112.95	164.0
115.25	232.0
117.05	16608.0
118.05	957.0
118.95	228.0
120.15	202.0
122.05	1375.0
123.05	1796.0
124.05	913.0
125.05	794.0
127.05	43952.0
128.05	3376.0
129.05	21432.0
130.05	1600.0
130.95	473.0
133.95	663.0
135.05	1874.0
136.05	731.0
136.95	705.0
138.25	163.0
140.05	350.0
141.05	2963.0
142.05	953.0
143.15	493.0
144.05	263.0
145.15	212.0
146.15	683.0
147.05	1589.0
148.05	3517.0
149.05	981.0
149.95	210.0
150.95	489.0
153.05	836.0
154.05	628.0

Instrument :

BNA_G

ClientSampleId :

DFTPP447

155.05	1694.0
156.05	1733.0
157.15	311.0
157.95	581.0
159.05	406.0
160.05	1290.0
161.05	1246.0
161.95	294.0
163.95	327.0
164.95	1455.0
166.05	1028.0
167.05	5450.0
168.05	2479.0
169.05	435.0
170.05	255.0
170.95	354.0
172.05	770.0
173.05	735.0
174.05	1227.0
175.05	2716.0
176.05	661.0
176.95	930.0
177.95	447.0
179.05	4444.0
180.05	2482.0
180.95	1297.0
182.05	186.0
183.95	536.0
185.05	2252.0
186.05	15592.0
187.05	4333.0
187.95	338.0
188.25	372.0
189.15	1272.0
189.95	266.0
191.05	718.0
192.05	1258.0
193.05	1332.0
194.05	432.0
195.95	2812.0
196.75	813.0
197.95	91488.0
198.95	6644.0
200.05	528.0
201.65	493.0
202.95	812.0
203.95	3981.0
205.05	7257.0
206.05	30240.0
207.05	4216.0
207.95	1090.0
208.95	492.0
209.95	446.0
210.25	472.0
211.05	1083.0
212.15	196.0
212.95	182.0
215.05	339.0
216.05	722.0
217.05	8225.0
217.95	1429.0
221.05	3732.0
222.05	859.0
222.95	2424.0
224.05	16936.0
225.05	4217.0
226.05	630.0
227.05	8909.0
228.05	1250.0
229.05	1266.0
229.95	199.0
230.95	598.0
233.05	281.0
234.05	421.0
234.95	475.0
235.95	371.0
237.05	545.0
238.95	459.0
239.95	402.0
241.05	365.0
242.05	838.0
243.05	838.0
244.05	13791.0
245.05	1930.0
246.05	2745.0
246.95	407.0
249.05	710.0
250.95	359.0
251.95	158.0
253.05	345.0
255.05	71000.0
256.05	11322.0
256.95	603.0
258.05	4512.0
258.95	585.0
261.25	176.0

Instrument :
BNA_G
ClientSampleId :
DFTPP447

264.95	1552.0
265.85	171.0
270.85	440.0
272.95	2452.0
273.95	5794.0
274.95	30608.0
276.05	4379.0
276.95	2683.0
278.05	460.0
282.05	186.0
283.05	553.0
284.05	260.0
285.05	659.0
290.95	155.0
293.05	651.0
295.95	10223.0
296.95	1604.0
302.95	1165.0
303.95	368.0
307.95	154.0
313.85	362.0
315.05	1001.0
316.05	707.0
320.85	213.0
323.05	2810.0
323.85	468.0
327.05	410.0
328.05	394.0
332.95	247.0
334.05	2143.0
335.05	577.0
340.95	417.0
345.95	600.0
351.95	1028.0
352.95	509.0
354.05	725.0
355.15	197.0
364.95	4133.0
365.95	397.0
370.85	255.0
371.95	1547.0
373.05	351.0
382.95	559.0
383.95	154.0
402.05	450.0
402.95	597.0
403.95	288.0
421.05	504.0
421.95	651.0
423.05	3964.0
424.05	740.0
441.05	12440.0
441.95	75672.0
442.95	14603.0
443.95	1475.0

Instrument :
BNA_G
ClientSampleId :
DFTPP447

m/z	Abundance
35.15	262.0
37.15	404.0
38.05	572.0
39.15	3197.0
40.05	2234.0
44.05	1065.0
44.95	156.0
47.05	400.0
48.05	167.0
49.05	1638.0
50.15	10262.0
51.15	38096.0
52.15	2288.0
54.95	197.0
56.05	1161.0
57.05	2748.0
61.05	807.0
62.15	629.0
63.15	1903.0
64.25	324.0
65.05	766.0
68.05	719.0
69.05	38624.0
70.05	375.0
73.25	589.0
74.15	4439.0
75.05	6614.0
76.15	2304.0
77.15	44288.0
78.15	3063.0
79.05	3057.0
80.05	2279.0
81.05	3843.0
82.05	810.0
82.95	724.0
84.05	1132.0
85.05	545.0
86.05	1415.0
87.15	428.0
91.05	1052.0
92.15	847.0
93.15	6251.0
94.15	464.0
94.95	158.0
96.05	253.0
98.05	4436.0
99.05	3249.0
100.05	256.0
101.05	1971.0
102.95	500.0
104.05	1372.0
105.05	1174.0
106.05	449.0
107.05	13895.0
108.05	2293.0
109.15	440.0
110.05	19832.0
111.05	3221.0
112.05	606.0
116.15	1121.0
117.05	13858.0
117.95	1136.0
119.95	288.0
122.05	900.0
123.05	1583.0
124.05	671.0
125.05	561.0
127.05	38552.0
128.05	3550.0
129.05	16448.0
130.05	1141.0
131.15	434.0
132.15	281.0
134.05	466.0
135.05	1551.0
135.95	647.0
137.05	830.0
140.05	266.0
141.05	2554.0
141.95	773.0
143.05	379.0
144.85	159.0
146.15	438.0
147.05	1509.0
148.05	3094.0
149.05	764.0
151.25	194.0
152.15	152.0
153.05	721.0
153.95	600.0
155.05	1215.0
156.05	1757.0

Instrument :
BNA_G
ClientSampleId :
DFTPP447

157.05	282.0
157.95	346.0
158.85	367.0
160.05	732.0
161.05	1223.0
162.05	247.0
163.15	188.0
165.05	1274.0
166.05	770.0
167.05	4431.0
168.05	2172.0
169.05	437.0
170.05	177.0
171.05	232.0
171.95	422.0
172.95	349.0
173.15	342.0
174.05	998.0
175.05	1689.0
176.05	397.0
176.95	972.0
178.15	441.0
179.05	4308.0
180.05	2106.0
181.05	1005.0
181.95	153.0
184.05	305.0
185.05	2212.0
186.05	12014.0
187.05	3894.0
188.05	204.0
189.05	907.0
189.95	173.0
191.05	500.0
191.95	1460.0
193.05	1366.0
193.95	364.0
195.95	2440.0
196.65	452.0
197.95	76200.0
198.95	5836.0
200.05	558.0
201.45	631.0
203.05	800.0
204.05	3885.0
205.05	5147.0
206.05	21944.0
207.05	3823.0
207.95	1245.0
209.05	478.0
210.05	432.0
211.05	799.0
211.85	199.0
214.95	277.0
215.85	401.0
216.95	7250.0
218.05	768.0
220.95	2956.0
222.05	692.0
222.95	1523.0
224.05	13415.0
225.05	3613.0
226.15	520.0
227.05	6254.0
228.05	918.0
229.05	1052.0
229.95	205.0
230.95	290.0
233.95	658.0
235.05	549.0
236.05	406.0
237.05	335.0
239.05	259.0
241.05	320.0
242.05	675.0
243.05	670.0
244.05	9514.0
245.05	1040.0
245.95	2588.0
246.95	532.0
247.95	168.0
248.85	375.0
251.05	192.0
251.95	188.0
252.95	335.0
255.05	49304.0
255.95	7756.0
256.95	698.0
258.05	3258.0
259.05	457.0
264.95	1356.0
266.05	280.0
271.05	161.0
271.95	253.0
273.05	1689.0
273.95	3608.0

Instrument :
BNA_G
ClientSampleId :
DFTPP447

274.95	21784.0
275.95	2788.0
276.95	1965.0
277.95	192.0
282.05	159.0
283.05	266.0
283.85	154.0
284.95	442.0
292.95	329.0
295.95	6981.0
296.85	792.0
302.95	967.0
303.95	306.0
313.95	168.0
315.05	602.0
315.95	466.0
317.25	159.0
320.85	290.0
323.05	1591.0
324.05	620.0
326.95	372.0
328.05	378.0
334.05	1282.0
335.15	346.0
341.05	255.0
345.95	437.0
351.95	581.0
353.05	562.0
353.95	688.0
364.95	2748.0
365.95	366.0
371.95	1009.0
373.05	235.0
383.05	274.0
401.95	329.0
402.95	579.0
403.95	211.0
420.85	328.0
422.05	329.0
422.95	3146.0
424.05	526.0
440.95	7356.0
441.95	45936.0
442.95	9238.0
443.95	772.0

Instrument :
BNA_G
ClientSampleId :
DFTPP447

m/z	Abundance
40.05	2316.0
44.05	509.0
47.15	270.0
47.85	158.0
49.05	1039.0
50.05	213.0
51.05	510.0
56.95	171.0
73.05	184.0
84.05	952.0
85.95	596.0
207.05	399.0
267.05	173.0
280.95	162.0

Instrument :

BNA_G

ClientSampleId :

DFTPP447