

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG123022\
 Data File : BG056177.D
 Acq On : 30 Dec 2022 19:17
 Operator : CG/JU
 Sample : N6220-01DL 5X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-1DL

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 01/03/2023
 Supervised By :Yogesh Patel 01/03/2023

Quant Time: Dec 31 00:20:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 05:20:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.341	152	52515	20.000	ng	0.00
21) Naphthalene-d8	11.167	136	185923	20.000	ng	0.00
39) Acenaphthene-d10	14.945	164	136506	20.000	ng	0.00
64) Phenanthrene-d10	17.683	188	320407	20.000	ng	0.00
76) Chrysene-d12	21.978	240	349645	20.000	ng	0.00
86) Perylene-d12	25.427	264	385448	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.855	112	69641	23.678	ng	0.00
7) Phenol-d6	7.465	99	98901	21.896	ng	0.00
23) Nitrobenzene-d5	9.510	82	51726	14.231	ng	0.00
42) 2,4,6-Tribromophenol	16.425	330	35521	20.546	ng	0.00
45) 2-Fluorobiphenyl	13.570	172	150276	15.202	ng	0.00
79) Terphenyl-d14	20.250	244	279104	14.297	ng	0.00
Target Compounds						Qvalue
37) 2-Methylnaphthalene	12.800	142	168582	20.405	ng	100
38) 1-Methylnaphthalene	13.012	142	120083	15.652	ng	95
58) Fluorene	15.985	166	31039	2.855	ng	# 89
71) Phenanthrene	17.724	178	134077	8.075	ng	99
75) Fluoranthene	19.710	202	64138	2.961	ng	95
78) Pyrene	20.068	202	80784m	3.278	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG123022\
Data File : BG056177.D
Acq On : 30 Dec 2022 19:17
Operator : CG/JU
Sample : N6220-01DL 5X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-1DL

Quant Time: Dec 31 00:20:14 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 28 05:20:14 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/03/2023
Supervised By :Yogesh Patel 01/03/2023

