

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG123114\
 Data File : BG015832.D
 Acq On : 31 Dec 2014 20:12
 Operator : TP/IZ
 Sample : SSTD00585
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jan 01 00:56:49 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG123114.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 01 00:49:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	26622	20.00	ng/ul	0.00
16) Naphthalene-d8	10.51	136	97261	20.00	ng/ul	0.00
33) Acenaphthene-d10	14.37	164	70532	20.00	ng/ul	0.00
59) Phenanthrene-d10	17.11	188	169378	20.00	ng/ul	0.00
73) Chrysene-d12	21.30	240	273611	20.00	ng/ul	0.00
81) Perylene-d12	23.56	264	259046	20.00	ng/ul	0.00

System Monitoring Compounds

3) Phenol-d5	6.91	99	10727	4.33	ng/ul	0.00
5) Bis-(2-Chloroethyl)ether-d	7.06	67	5641	4.09	ng/ul	0.00
7) 2-Chlorophenol-d4	7.26	132	7377	4.45	ng/ul	0.00
11) 4-Methylphenol-d8	8.45	113	8631	4.80	ng/ul	0.00
17) Nitrobenzene-d5	8.88	128	3793	4.83	ng/ul	0.00
20) 2-Nitrophenol-d4	9.60	143	3703	4.39	ng/ul	0.00
24) 2,4-Dichlorophenol-d3	10.14	165	7477	4.65	ng/ul	0.00
27) 4-Chloroaniline-d4	10.65	131	8903	4.99	ng/ul	0.00
41) Dimethylphthalate-d6	13.77	166	22744	4.72	ng/ul	0.00
44) Acenaphthylene-d8	14.06	160	28818	4.88	ng/ul	0.00
49) 4-Nitrophenol-d4	14.58	143	3570	4.04	ng/ul	0.00
55) Fluorene-d10	15.36	176	23225	4.87	ng/ul	0.00
60) 4,6-Dinitro-2-methylphenol	15.48	200	4073	4.21	ng/ul	0.00
68) Anthracene-d10	17.21	188	38028	4.86	ng/ul	0.00
74) Pyrene-d10	19.50	212	49931	4.52	ng/ul	0.00
85) Benzo(a)pyrene-d12	23.42	264	54605	4.67	ng/ul	0.00

Target Compounds

					Qvalue
2) Benzaldehyde	6.87	77	7724	4.76	ng/ul 90
4) Phenol	6.93	94	11562	4.43	ng/ul 95
6) Bis(2-Chloroethyl)ether	7.16	93	7528	3.91	ng/ul# 80
8) 2-Chlorophenol	7.29	128	7265	4.49	ng/ul 94
9) 2-Methylphenol	8.18	108	7839	4.26	ng/ul 90
10) 2,2'-oxybis(1-Chloropropan	8.27	45	3947	2.85	ng/ul# 91
12) Acetophenone	8.55	105	13258	4.40	ng/ul 84
13) N-Nitroso-di-n-propylamine	8.53	70	7257	4.39	ng/ul# 88
14) 4-Methylphenol	8.50	108	8134	3.97	ng/ul 97
15) Hexachloroethane	8.80	117	4287	4.91	ng/ul 96
18) Nitrobenzene	8.93	77	11290	4.50	ng/ul 98
19) Isophorone	9.44	82	20180	4.71	ng/ul# 96
21) 2-Nitrophenol	9.63	139	4423	4.80	ng/ul# 78
22) 2,4-Dimethylphenol	9.70	107	11054	4.63	ng/ul 90
23) Bis(2-Chloroethoxy)methane	9.93	93	10265	4.36	ng/ul# 95
25) 2,4-Dichlorophenol	10.17	162	7086	4.30	ng/ul# 82
26) Naphthalene	10.56	128	22227	4.51	ng/ul# 92
28) 4-Chloroaniline	10.68	127	8796	4.80	ng/ul# 74
29) Hexachlorobutadiene	10.85	225	8573	5.41	ng/ul# 84
30) Caprolactam	11.42	113	2864	4.20	ng/ul# 67
31) 4-Chloro-3-methylphenol	11.81	107	9821	4.45	ng/ul# 87
32) 2-Methylnaphthalene	12.18	142	16760	4.60	ng/ul 93
34) 1,2,4,5-Tetrachlorobenzene	12.55	216	13094	5.33	ng/ul 96
35) Hexachlorocyclopentadiene	12.53	237	8221	5.85	ng/ul# 80


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG123114\
 Data File : BG015832.D
 Acq On : 31 Dec 2014 20:12
 Operator : TP/IZ
 Sample : SSTD00585
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jan 01 00:56:49 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG123114.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 01 00:49:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

36) 2,4,6-Trichlorophenol	12.79	196	7247	4.84	ng/ul#	90
37) 2,4,5-Trichlorophenol	12.86	196	7733	4.75	ng/ul#	79
38) 1,1'-Biphenyl	13.20	154	24551	5.02	ng/ul	97
39) 2-Chloronaphthalene	13.24	162	18539	4.86	ng/ul	95
40) 2-Nitroaniline	13.45	65	6534	4.58	ng/ul	90
42) Dimethylphthalate	13.83	163	25151	4.99	ng/ul	99
43) 2,6-Dinitrotoluene	13.95	165	5770	5.24	ng/ul	92
45) Acenaphthylene	14.09	152	30735	4.81	ng/ul	96
46) 3-Nitroaniline	14.27	138	4564	4.68	ng/ul#	91
47) Acenaphthene	14.43	153	20499	4.95	ng/ul	94
48) 2,4-Dinitrophenol	14.49	184	1160	2.08	ng/ul#	56
51) Dibenzofuran	14.77	168	29068	4.67	ng/ul	95
52) 2,4-Dinitrotoluene	14.74	165	7250	4.59	ng/ul#	93
53) 2,3,4,6-Tetrachlorophenol	15.00	232	8529	4.96	ng/ul#	87
54) Diethylphthalate	15.19	149	26860	4.98	ng/ul#	94
56) Fluorene	15.41	166	23553	4.70	ng/ul	88
57) 4-Chlorophenyl-phenylether	15.41	204	16768	5.19	ng/ul#	92
58) 4-Nitroaniline	15.44	138	5289	4.51	ng/ul	85
61) 4,6-Dinitro-2-methylphenol	15.50	198	4030	3.95	ng/ul#	95
62) N-Nitrosodiphenylamine	15.63	169	22638	4.76	ng/ul	98
63) 4-Bromophenyl-phenylether	16.30	248	10412	4.73	ng/ul#	83
64) Hexachlorobenzene	16.41	284	12021	5.03	ng/ul#	74
65) Atrazine	16.58	200	10903	5.33	ng/ul	96
67) Phenanthrene	17.15	178	40152	4.68	ng/ul#	93
69) Anthracene	17.24	178	41176	4.69	ng/ul	99
70) Carbazole	17.52	167	38902	4.82	ng/ul	98
71) Di-n-butylphthalate	18.08	149	48801	5.19	ng/ul#	95
72) Fluoranthene	19.17	202	61060	5.18	ng/ul	98
75) Pyrene	19.53	202	66881	4.56	ng/ul	98
76) Butylbenzylphthalate	20.43	149	26593	4.80	ng/ul	95
77) 3,3'-Dichlorobenzidine	21.21	252	27609	5.48	ng/ul#	99
78) Benzo(a)anthracene	21.28	228	74627	4.82	ng/ul	97
79) Bis(2-ethylhexyl)phthalate	21.21	149	42165	5.52	ng/ul#	96
80) Chrysene	21.33	228	71465	4.92	ng/ul	96
82) Di-n-octyl phthalate	22.09	149	66171	5.30	ng/ul	100
83) Benzo(b)fluoranthene	22.88	252	73424	4.70	ng/ul#	97
84) Benzo(k)fluoranthene	22.92	252	72630	4.96	ng/ul#	95
86) Benzo(a)pyrene	23.46	252	67408	4.58	ng/ul	97
87) Indeno(1,2,3-cd)pyrene	25.86	276	77015	4.72	ng/ul#	96
88) Dibenzo(a,h)anthracene	25.87	278	64940	4.60	ng/ul#	89
89) Benzo(g,h,i)perylene	26.57	276	65010	4.79	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

Benzaldehyde

Concen: 4.76 ng/ul

RT: 6.87 min Scan# 712

Delta R.T. -0.01 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12


Tgt Ion: 77 Resp: 7724

Ion Ratio Lower Upper




77 100

105 73.1 71.1 106.7

106 68.7 55.8 83.8



Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 105.00 (104.70 to 105.70): BGC
Ion 106.00 (105.70 to 106.70): BGC



#4

Phenol

Concen: 4.43 ng/ul

RT: 6.93 min Scan# 722

Delta R.T. -0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12


Tgt Ion: 94 Resp: 11562

Ion Ratio Lower Upper



94 100


65 29.7 27.4 41.0

66 55.0 46.4 69.6



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC





#6

Bis(2-Chloroethyl)ether

Concen: 3.91 ng/ul


RT: 7.16 min Scan# 761

Delta R.T. -0.00 min




Lab File: BG015832.D

Acq: 31 Dec 2014 20:12

Tgt Ion:	93	Resp:	7528
Ion	Ratio	Lower	Upper
93	100		
63	69.9	45.0	67.6#
95	40.5	23.3	34.9#



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC



#8

2-Chlorophenol

Concen: 4.49 ng/ul


RT: 7.29 min Scan# 784

Delta R.T. -0.00 min



Lab File: BG015832.D


Acq: 31 Dec 2014 20:12

Tgt Ion:	128	Resp:	7265
Ion	Ratio	Lower	Upper
128	100		
64	51.1	45.4	68.0
130	39.4	29.5	44.3



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 130.00 (129.70 to 130.70): E





#9

2-Methylphenol

Concen: 4.26 ng/ul

RT: 8.18 min Scan# 934

Delta R.T. -0.00 min

Lab File: BG015832.D


Acq: 31 Dec 2014 20:12

Tgt Ion: 108 Resp: 7839

Ion Ratio Lower Upper

108 100

107 92.0 66.6 99.8





Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E

8.18

8.15 8.20

Time-->



#10

2,2'-oxybis(1-Chloropropane)

Concen: 2.85 ng/ul

RT: 8.27 min Scan# 950

Delta R.T. 0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12


Tgt Ion: 45 Resp: 3947

Ion Ratio Lower Upper

45 100

77 31.0 24.6 37.0

79 14.2 19.5 29.3#



Abundance Ion 45.00 (44.70 to 45.70): BGC


Ion 77.00 (76.70 to 77.70): BGC


Ion 79.00 (78.70 to 79.70): BGC

8.27

8.24 8.26 8.28 8.30

Time-->





#12

Acetophenone

Concen: 4.40 ng/ul

RT: 8.55 min Scan# 997

Delta R.T. -0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12


Tgt Ion: 105 Resp: 13258

Ion Ratio Lower Upper

105 100

77 87.5 86.6 129.8




51 27.0 22.2 33.4



Abundance Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BG0

Ion 51.00 (50.70 to 51.70): BG0



#13

N-Nitroso-di-n-propylamine

Concen: 4.39 ng/ul

RT: 8.53 min Scan# 995

Delta R.T. 0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12

Tgt Ion: 70 Resp: 7257


Ion Ratio Lower Upper

70 100

42 44.2 28.2 42.4#

101 11.1 8.7 13.1

130 17.8 18.9 28.3#






Abundance Ion 70.00 (69.70 to 70.70): BG0

Ion 42.00 (41.70 to 42.70): BG0

Ion 101.00 (100.70 to 101.70): E





Ion 130.00 (129.70 to 130.70): E








#14
4-Methylphenol
Concen: 3.97 ng/ul
RT: 8.50 min Scan# 989
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12


Tgt Ion:108 Resp: 8134
Ion Ratio Lower Upper
108 100
107 122.3 95.3 142.9



#15
Hexachloroethane
Concen: 4.91 ng/ul
RT: 8.80 min Scan# 1041
Delta R.T. 0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12





Tgt Ion:117 Resp: 4287
Ion Ratio Lower Upper
117 100
201 101.3 78.2 117.2
199 55.0 42.2 63.2








#18
Nitrobenzene
Concen: 4.50 ng/ul
RT: 8.93 min Scan# 1062
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12


Tgt Ion:	77	Resp:	11290
Ion	Ratio	Lower	Upper
77	100		
123	36.0	29.7	44.5
65	10.9	8.2	12.2



#19
Isophorone
Concen: 4.71 ng/ul
RT: 9.44 min Scan# 1150
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12





Tgt Ion:	82	Resp:	20180
Ion	Ratio	Lower	Upper
82	100		
95	7.4	8.7	13.1#
138	16.4	12.8	19.2








#21
 2-Nitrophenol
 Concen: 4.80 ng/ul
 RT: 9.63 min Scan# 1182
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12


Tgt Ion:139 Resp: 4423
 Ion Ratio Lower Upper
 139 100
 65 56.7 59.8 89.8#
 109 42.0 46.8 70.2#



#22
 2,4-Dimethylphenol
 Concen: 4.63 ng/ul
 RT: 9.70 min Scan# 1193
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion:107 Resp: 11054
 Ion Ratio Lower Upper
 107 100
 121 40.6 29.4 44.2
 122 54.2 50.9 76.3





#23

Bis(2-Chloroethoxy)methane

Concen: 4.36 ng/ul


RT: 9.93 min Scan# 1233

Delta R.T. 0.00 min




Lab File: BG015832.D

Acq: 31 Dec 2014 20:12

Tgt Ion:	93	Resp:	10265
Ion	Ratio	Lower	Upper
93	100		
95	33.0	25.1	37.7
123	17.0	10.6	15.8#



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 95.00 (94.70 to 95.70): BGC
Ion 123.00 (122.70 to 123.70): BGC



#25

2,4-Dichlorophenol

Concen: 4.30 ng/ul


RT: 10.17 min Scan# 1273

Delta R.T. 0.00 min



Lab File: BG015832.D


Acq: 31 Dec 2014 20:12

Tgt Ion:	162	Resp:	7086
Ion	Ratio	Lower	Upper
162	100		
164	78.1	50.6	76.0#
98	29.9	32.0	48.0#




Abundance Ion 162.00 (161.70 to 162.70): BGC
Ion 164.00 (163.70 to 164.70): BGC
Ion 98.00 (97.70 to 98.70): BGC








#26
Naphthalene
Concen: 4.51 ng/ul
RT: 10.56 min Scan# 1340
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12

Tgt Ion:128 Resp: 22227
Ion Ratio Lower Upper
128 100
129 12.0 7.1 10.7#
127 14.9 9.8 14.6#




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#28
4-Chloroaniline
Concen: 4.80 ng/ul
RT: 10.68 min Scan# 1360
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12

Tgt Ion:127 Resp: 8796
Ion Ratio Lower Upper
127 100
129 47.0 26.0 39.0#




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E








#29
Hexachlorobutadiene
Concen: 5.41 ng/ul
RT: 10.85 min Scan# 1389
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12

Tgt Ion: 225 Resp: 8573
Ion Ratio Lower Upper
225 100
223 55.1 49.0 73.4
227 79.8 49.3 73.9#




Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E






#30
Caprolactam
Concen: 4.20 ng/ul
RT: 11.42 min Scan# 1487
Delta R.T. -0.01 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12

Tgt Ion: 113 Resp: 2864
Ion Ratio Lower Upper
113 100
55 65.4 66.1 99.1#
56 43.9 69.0 103.6#




Abundance Ion 113.00 (112.70 to 113.70): E
Ion 55.00 (54.70 to 55.70): BG
Ion 56.00 (55.70 to 56.70): BG








#31
 4-Chloro-3-methylphenol
 Concen: 4.45 ng/ul
 RT: 11.81 min Scan# 1552
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion: 107 Resp: 9821
 Ion Ratio Lower Upper
 107 100
 144 28.4 14.6 22.0#
 142 53.0 48.2 72.2




Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E






#32
 2-Methylnaphthalene
 Concen: 4.60 ng/ul
 RT: 12.18 min Scan# 1616
 Delta R.T. 0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion: 142 Resp: 16760
 Ion Ratio Lower Upper
 142 100
 141 96.8 72.4 108.6




Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E








#34
 1,2,4,5-Tetrachlorobenzene
 Concen: 5.33 ng/ul
 RT: 12.55 min Scan# 1679
 Delta R.T. 0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion:	216	Resp:	13094
Ion	Ratio	Lower	Upper
216	100		
214	71.3	59.7	89.5
179	18.8	16.9	25.3
108	17.3	14.2	21.4




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#35
 Hexachlorocyclopentadiene
 Concen: 5.85 ng/ul
 RT: 12.53 min Scan# 1675
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion:	237	Resp:	8221
Ion	Ratio	Lower	Upper
237	100		
235	46.9	53.0	79.4#
272	15.7	13.0	19.4




Abundance Ion 237.00 (236.70 to 237.70): E
 Ion 235.00 (234.70 to 235.70): E
 Ion 272.00 (271.70 to 272.70): E








#36
 2,4,6-Trichlorophenol
 Concen: 4.84 ng/ul
 RT: 12.79 min Scan# 1720
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion	Ratio	Lower	Upper
196	100		
198	94.7	81.6	122.4
200	44.4	26.2	39.4




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E






#37
 2,4,5-Trichlorophenol
 Concen: 4.75 ng/ul
 RT: 12.86 min Scan# 1732
 Delta R.T. -0.01 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion	Ratio	Lower	Upper
196	100		
198	72.0	78.1	117.1
200	27.7	24.8	37.2




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E








#38
 1,1'-Biphenyl
 Concen: 5.02 ng/ul
 RT: 13.20 min Scan# 1789
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion:	154	Resp:	24551
Ion	Ratio	Lower	Upper
154	100		
153	34.3	28.7	43.1
76	9.6	8.2	12.4




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): B






#39
 2-Chloronaphthalene
 Concen: 4.86 ng/ul
 RT: 13.24 min Scan# 1796
 Delta R.T. 0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion:	162	Resp:	18539
Ion	Ratio	Lower	Upper
162	100		
164	27.0	25.8	38.6
127	40.7	33.3	49.9







Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 127.00 (126.70 to 127.70): E








#40
2-Nitroaniline
Concen: 4.58 ng/ul
RT: 13.45 min Scan# 1831
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12


Tgt Ion: 65 Resp: 6534
Ion Ratio Lower Upper
65 100
92 63.0 54.7 82.1
138 88.9 61.9 92.9



#42
Dimethylphthalate
Concen: 4.99 ng/ul
RT: 13.83 min Scan# 1896
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12


Tgt Ion:163 Resp: 25151
Ion Ratio Lower Upper
163 100
194 6.6 5.7 8.5
164 9.5 7.8 11.6








#43
 2,6-Dinitrotoluene
 Concen: 5.24 ng/ul
 RT: 13.95 min Scan# 1916
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion:165 Resp: 5770
 Ion Ratio Lower Upper
 165 100
 89 91.5 79.4 119.2
 121 24.5 23.9 35.9




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 89.00 (88.70 to 89.70): BGC
 Ion 121.00 (120.70 to 121.70): E






#45
 Acenaphthylene
 Concen: 4.81 ng/ul
 RT: 14.09 min Scan# 1940
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion:152 Resp: 30735
 Ion Ratio Lower Upper
 152 100
 151 21.2 15.4 23.0
 153 11.4 10.5 15.7







Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 151.00 (150.70 to 151.70): E
 Ion 153.00 (152.70 to 153.70): E








#46
 3-Nitroaniline
 Concen: 4.68 ng/ul
 RT: 14.27 min Scan# 1972
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12


Tgt Ion:138 Resp: 4564
 Ion Ratio Lower Upper
 138 100
 108 23.0 10.1 15.1#
 92 142.8 121.8 182.6



#47
 Acenaphthene
 Concen: 4.95 ng/ul
 RT: 14.43 min Scan# 1998
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12





Tgt Ion:153 Resp: 20499
 Ion Ratio Lower Upper
 153 100
 152 43.8 34.5 51.7
 154 82.7 73.0 109.4








#48
2,4-Dinitrophenol
Concen: 2.08 ng/ul
RT: 14.49 min Scan# 2008
Delta R.T. 0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12


Tgt Ion:184 Resp: 1160
Ion Ratio Lower Upper
184 100
63 83.8 42.0 63.0#
154 19.5 39.6 59.4#



#51
Dibenzofuran
Concen: 4.67 ng/ul
RT: 14.77 min Scan# 2056
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12





Tgt Ion:168 Resp: 29068
Ion Ratio Lower Upper
168 100
139 36.7 31.8 47.6








#52
 2,4-Dinitrotoluene
 Concen: 4.59 ng/ul
 RT: 14.74 min Scan# 2051
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12


Tgt Ion	Ratio	Lower	Upper
165	100		
63	49.1	43.0	64.6
182	8.0	3.4	5.2



#53
 2,3,4,6-Tetrachlorophenol
 Concen: 4.96 ng/ul
 RT: 15.00 min Scan# 2095
 Delta R.T. 0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12


Tgt Ion	Ratio	Lower	Upper
232	100		
131	28.0	32.0	48.0
130	7.0	1.4	2.2
166	23.4	17.7	26.5








#54
Diethylphthalate
Concen: 4.98 ng/ul
RT: 15.19 min Scan# 2128
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12

Tgt Ion:149 Resp: 26860
Ion Ratio Lower Upper
149 100
177 24.3 18.2 27.2
150 14.9 8.7 13.1#




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#56
Fluorene
Concen: 4.70 ng/ul
RT: 15.41 min Scan# 2166
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12

Tgt Ion:166 Resp: 23553
Ion Ratio Lower Upper
166 100
165 107.5 75.4 113.2
167 14.6 13.1 19.7




Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E








#57
 4-Chlorophenyl-phenylether
 Concen: 5.19 ng/ul
 RT: 15.41 min Scan# 2166
 Delta R.T. 0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion: 204 Resp: 16768
 Ion Ratio Lower Upper
 204 100
 206 45.1 29.6 44.4#
 141 61.7 46.8 70.2




Abundance Ion 204.00 (203.70 to 204.70): E
 Ion 206.00 (205.70 to 206.70): E
 Ion 141.00 (140.70 to 141.70): E






#58
 4-Nitroaniline
 Concen: 4.51 ng/ul
 RT: 15.44 min Scan# 2170
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion: 138 Resp: 5289
 Ion Ratio Lower Upper
 138 100
 92 76.6 57.5 86.3
 108 142.7 134.6 202.0




Abundance Ion 138.00 (137.70 to 138.70): E
 Ion 92.00 (91.70 to 92.70): BG
 Ion 108.00 (107.70 to 108.70): E








#61
 4,6-Dinitro-2-methylphenol
 Concen: 3.95 ng/ul
 RT: 15.50 min Scan# 2180
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion	Ratio	Lower	Upper
198	100		
182	0.0	4.6	6.8#
77	31.2	26.3	39.5




Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 77.00 (76.70 to 77.70): B






#62
 N-Nitrosodiphenylamine
 Concen: 4.76 ng/ul
 RT: 15.63 min Scan# 2202
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion	Ratio	Lower	Upper
169	100		
168	60.8	49.9	74.9
167	34.2	26.7	40.1







Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E








#63
4-Bromophenyl-phenylether
Concen: 4.73 ng/ul
RT: 16.30 min Scan# 2317
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12


Tgt Ion:248 Resp: 10412
Ion Ratio Lower Upper
248 100
250 86.8 78.4 117.6
141 54.2 60.5 90.7#



#64
Hexachlorobenzene
Concen: 5.03 ng/ul
RT: 16.41 min Scan# 2336
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12





Tgt Ion:284 Resp: 12021
Ion Ratio Lower Upper
284 100
142 36.2 25.0 37.4
249 3.1 21.6 32.4#








#65
Atrazine
Concen: 5.33 ng/ul
RT: 16.58 min Scan# 2364
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12


Tgt Ion	Ratio	Lower	Upper
200	100		
173	21.5	18.2	27.4
215	54.4	41.0	61.4



#67
Phenanthrene
Concen: 4.68 ng/ul
RT: 17.15 min Scan# 2462
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12


Tgt Ion	Ratio	Lower	Upper
178	100		
179	18.1	12.4	18.6
176	13.9	14.0	21.0








#69
 Anthracene
 Concen: 4.69 ng/ul
 RT: 17.24 min Scan# 2477
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion	Ratio	Lower	Upper
178	100		
179	15.6	12.0	18.0
176	18.7	15.0	22.6




Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 176.00 (175.70 to 176.70): E






#70
 Carbazole
 Concen: 4.82 ng/ul
 RT: 17.52 min Scan# 2524
 Delta R.T. 0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion	Ratio	Lower	Upper
167	100		
166	18.9	14.6	22.0
139	15.2	11.0	16.4







Abundance Ion 167.00 (166.70 to 167.70): E
 Ion 166.00 (165.70 to 166.70): E
 Ion 139.00 (138.70 to 139.70): E








#71
Di-n-butylphthalate
Concen: 5.19 ng/ul
RT: 18.08 min Scan# 2619
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12


Tgt Ion	Ratio	Lower	Upper
149	100		
150	9.8	7.8	11.8
104	5.3	7.1	10.7



#72
Fluoranthene
Concen: 5.18 ng/ul
RT: 19.17 min Scan# 2805
Delta R.T. -0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12

Tgt Ion	Ratio	Lower	Upper
202	100		
101	6.4	4.7	7.1
100	5.3	4.7	7.1





#75

Pyrene

Concen: 4.56 ng/ul


RT: 19.53 min Scan# 2867

Delta R.T. -0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12




Tgt Ion:	202	Resp:	66881
Ion	Ratio	Lower	Upper
202	100		
101	8.1	5.7	8.5
100	5.9	4.6	7.0



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#76

Butylbenzylphthalate

Concen: 4.80 ng/ul


RT: 20.43 min Scan# 3020

Delta R.T. -0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12



Tgt Ion:	149	Resp:	26593
Ion	Ratio	Lower	Upper
149	100		
91	89.6	68.6	102.8
206	25.3	23.7	35.5




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): B





Ion 206.00 (205.70 to 206.70): E








#77
 3,3'-Dichlorobenzidine
 Concen: 5.48 ng/ul
 RT: 21.21 min Scan# 3153
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12


Tgt Ion	Ratio	Lower	Upper
252	100		
254	65.9	53.1	79.7
126	8.2	4.3	6.5#



#78
 Benzo(a)anthracene
 Concen: 4.82 ng/ul
 RT: 21.28 min Scan# 3164
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12


Tgt Ion	Ratio	Lower	Upper
228	100		
229	18.8	15.8	23.8
226	25.9	21.9	32.9








#79
 Bis(2-ethylhexyl)phthalate
 Concen: 5.52 ng/ul
 RT: 21.21 min Scan# 3152
 Delta R.T. 0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion	Ratio	Lower	Upper
149	100		
167	33.5	25.6	38.4
279	14.1	9.3	13.9




Abundance Ion 149.00 (148.70 to 149.70): E
 Ion 167.00 (166.70 to 167.70): E
 Ion 279.00 (278.70 to 279.70): E






#80
 Chrysene
 Concen: 4.92 ng/ul
 RT: 21.33 min Scan# 3173
 Delta R.T. -0.00 min
 Lab File: BG015832.D
 Acq: 31 Dec 2014 20:12

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.1	24.3	36.5
229	21.9	16.0	24.0



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E









#82

Di-n-octyl phthalate
Concen: 5.30 ng/ul
RT: 22.09 min Scan# 3302
Delta R.T. 0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12

Tgt Ion:149 Resp: 66171




Abundance Ion 149.00 (148.70 to 149.70): E





#83


Benzo(b)fluoranthene
Concen: 4.70 ng/ul
RT: 22.88 min Scan# 3436
Delta R.T. 0.00 min
Lab File: BG015832.D
Acq: 31 Dec 2014 20:12

Tgt Ion:252 Resp: 73424
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 9.4 3.4 5.0#



Abundance Ion 252.00 (251.70 to 252.70): E





#84

Benzo(k)fluoranthene

Concen: 4.96 ng/ul

RT: 22.92 min Scan# 3443

Delta R.T. -0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12


Tgt Ion:252 Resp: 72630

Ion Ratio Lower Upper




252 100

253 23.5 17.4 26.0

125 8.1 2.9 4.3#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#86

Benzo(a)pyrene

Concen: 4.58 ng/ul

RT: 23.46 min Scan# 3536

Delta R.T. 0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12


Tgt Ion:252 Resp: 67408

Ion Ratio Lower Upper



252 100


253 22.7 17.0 25.4

125 4.6 4.2 6.4



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#87

Indeno(1,2,3-cd)pyrene

Concen: 4.72 ng/ul

RT: 25.86 min Scan# 3944

Delta R.T. -0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12


Tgt Ion:276 Resp: 77015

Ion Ratio Lower Upper

276 100

138 14.1 7.0 10.6#




277 27.7 22.0 33.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#88

Dibenzo(a,h)anthracene

Concen: 4.60 ng/ul

RT: 25.87 min Scan# 3946

Delta R.T. -0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12


Tgt Ion:278 Resp: 64940

Ion Ratio Lower Upper

278 100

139 5.0 5.4 8.2#



279 18.3 19.8 29.8#




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#89

Benzo(g,h,i)perylene

Concen: 4.79 ng/ul

RT: 26.57 min Scan# 4065

Delta R.T. 0.00 min

Lab File: BG015832.D

Acq: 31 Dec 2014 20:12

Tgt Ion:	276	Resp:	65010
Ion	Ratio	Lower	Upper
276	100		
138	7.9	6.9	10.3
277	18.6	19.4	29.2#

