

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG123114\
 Data File : BG015833.D
 Acq On : 31 Dec 2014 20:46
 Operator : TP/IZ
 Sample : SSTD01086
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jan 01 00:57:17 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG123114.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 01 00:49:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	25499	20.00	ng/ul	0.00
16) Naphthalene-d8	10.51	136	94806	20.00	ng/ul	0.00
33) Acenaphthene-d10	14.37	164	71198	20.00	ng/ul	0.00
59) Phenanthrene-d10	17.11	188	164257	20.00	ng/ul	0.00
73) Chrysene-d12	21.29	240	271649	20.00	ng/ul	0.00
81) Perylene-d12	23.56	264	256478	20.00	ng/ul	0.00

System Monitoring Compounds

3) Phenol-d5	6.91	99	21202	8.93	ng/ul	0.00
5) Bis-(2-Chloroethyl)ether-d	7.07	67	11183	8.46	ng/ul	0.00
7) 2-Chlorophenol-d4	7.26	132	15769	9.93	ng/ul	0.00
11) 4-Methylphenol-d8	8.43	113	16421	9.53	ng/ul	0.00
17) Nitrobenzene-d5	8.88	128	7160	9.36	ng/ul	0.00
20) 2-Nitrophenol-d4	9.61	143	8763	10.66	ng/ul	0.00
24) 2,4-Dichlorophenol-d3	10.14	165	15595	9.94	ng/ul	0.00
27) 4-Chloroaniline-d4	10.65	131	18510	10.65	ng/ul	0.00
41) Dimethylphthalate-d6	13.78	166	52210	10.74	ng/ul	0.00
44) Acenaphthylene-d8	14.06	160	58985	9.89	ng/ul	0.00
49) 4-Nitrophenol-d4	14.58	143	7977	8.94	ng/ul	0.00
55) Fluorene-d10	15.36	176	48329	10.03	ng/ul	0.00
60) 4,6-Dinitro-2-methylphenol	15.48	200	7622	8.12	ng/ul	0.00
68) Anthracene-d10	17.21	188	81228	10.70	ng/ul	0.00
74) Pyrene-d10	19.50	212	110338	10.05	ng/ul	0.00
85) Benzo(a)pyrene-d12	23.41	264	115124	9.93	ng/ul	0.00

Target Compounds

						Qvalue
2) Benzaldehyde	6.88	77	15034	9.67	ng/ul#	88
4) Phenol	6.94	94	21978	8.79	ng/ul#	88
6) Bis(2-Chloroethyl)ether	7.16	93	17586	9.55	ng/ul	95
8) 2-Chlorophenol	7.30	128	16488	10.64	ng/ul#	87
9) 2-Methylphenol	8.17	108	16668	9.46	ng/ul	92
10) 2,2'-oxybis(1-Chloropropan	8.26	45	8366	6.31	ng/ul	93
12) Acetophenone	8.55	105	27435	9.51	ng/ul	89
13) N-Nitroso-di-n-propylamine	8.53	70	15784	9.97	ng/ul#	93
14) 4-Methylphenol	8.50	108	17622	8.98	ng/ul	96
15) Hexachloroethane	8.80	117	9095	10.87	ng/ul#	91
18) Nitrobenzene	8.93	77	24687	10.10	ng/ul#	96
19) Isophorone	9.44	82	41141	9.86	ng/ul	95
21) 2-Nitrophenol	9.64	139	8661	9.65	ng/ul#	79
22) 2,4-Dimethylphenol	9.70	107	25318	10.88	ng/ul	98
23) Bis(2-Chloroethoxy)methane	9.93	93	20803	9.07	ng/ul	98
25) 2,4-Dichlorophenol	10.17	162	17166	10.70	ng/ul#	92
26) Naphthalene	10.56	128	47345	9.86	ng/ul	99
28) 4-Chloroaniline	10.67	127	18744	10.50	ng/ul	99
29) Hexachlorobutadiene	10.85	225	19158	12.40	ng/ul	89
30) Caprolactam	11.42	113	3725	5.60	ng/ul	94
31) 4-Chloro-3-methylphenol	11.81	107	19929	9.27	ng/ul#	97
32) 2-Methylnaphthalene	12.18	142	36968	10.42	ng/ul	81
34) 1,2,4,5-Tetrachlorobenzene	12.55	216	27480	11.08	ng/ul#	91
36) 2,4,6-Trichlorophenol	12.80	196	16362	10.82	ng/ul#	79


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG123114\
 Data File : BG015833.D
 Acq On : 31 Dec 2014 20:46
 Operator : TP/IZ
 Sample : SSTD01086
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jan 01 00:57:17 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG123114.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 01 00:49:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

37) 2,4,5-Trichlorophenol	12.87	196	15921	9.68	ng/ul#	95
38) 1,1'-Biphenyl	13.20	154	50512	10.24	ng/ul	92
39) 2-Chloronaphthalene	13.24	162	38019	9.87	ng/ul	97
40) 2-Nitroaniline	13.45	65	12574	8.73	ng/ul	97
42) Dimethylphthalate	13.83	163	51719	10.16	ng/ul	98
43) 2,6-Dinitrotoluene	13.94	165	11424	10.27	ng/ul#	70
45) Acenaphthylene	14.09	152	66430	10.29	ng/ul	97
46) 3-Nitroaniline	14.28	138	10419	10.58	ng/ul	89
47) Acenaphthene	14.43	153	43298	10.37	ng/ul	94
48) 2,4-Dinitrophenol	14.49	184	3002	5.33	ng/ul#	50
50) 4-Nitrophenol	14.59	109	12928	9.65	ng/ul#	79
51) Dibenzofuran	14.77	168	64699	10.30	ng/ul	95
52) 2,4-Dinitrotoluene	14.74	165	15842	9.94	ng/ul#	98
53) 2,3,4,6-Tetrachlorophenol	15.00	232	18381	10.59	ng/ul#	99
54) Diethylphthalate	15.19	149	54718	10.05	ng/ul#	95
56) Fluorene	15.41	166	50354	9.96	ng/ul	99
57) 4-Chlorophenyl-phenylether	15.41	204	35906	11.00	ng/ul#	80
58) 4-Nitroaniline	15.44	138	11552	9.76	ng/ul#	92
61) 4,6-Dinitro-2-methylphenol	15.50	198	10317	10.44	ng/ul#	86
62) N-Nitrosodiphenylamine	15.62	169	47139	10.22	ng/ul	95
63) 4-Bromophenyl-phenylether	16.31	248	21573	10.11	ng/ul	94
64) Hexachlorobenzene	16.42	284	23492	10.15	ng/ul#	91
65) Atrazine	16.58	200	22287	11.23	ng/ul#	87
66) Pentachlorophenol	16.76	266	12247	9.65	ng/ul#	89
67) Phenanthrene	17.15	178	83328	10.02	ng/ul	96
69) Anthracene	17.25	178	90844	10.66	ng/ul	98
70) Carbazole	17.52	167	82055	10.48	ng/ul#	92
71) Di-n-butylphthalate	18.08	149	99462	10.91	ng/ul	98
72) Fluoranthene	19.17	202	131106	11.47	ng/ul	99
75) Pyrene	19.53	202	140543	9.65	ng/ul	98
76) Butylbenzylphthalate	20.43	149	52225	9.49	ng/ul	93
77) 3,3'-Dichlorobenzidine	21.21	252	53648	10.73	ng/ul#	92
78) Benzo(a)anthracene	21.28	228	158669	10.31	ng/ul	97
79) Bis(2-ethylhexyl)phthalate	21.21	149	77716	10.26	ng/ul	95
80) Chrysene	21.33	228	148682	10.31	ng/ul	96
82) Di-n-octyl phthalate	22.09	149	130756	10.58	ng/ul	100
83) Benzo(b)fluoranthene	22.88	252	148405	9.60	ng/ul	97
84) Benzo(k)fluoranthene	22.92	252	150529	10.39	ng/ul	98
86) Benzo(a)pyrene	23.46	252	146068	10.01	ng/ul	99
87) Indeno(1,2,3-cd)pyrene	25.86	276	162303	10.05	ng/ul	96
88) Dibenzo(a,h)anthracene	25.87	278	137362	9.83	ng/ul#	94
89) Benzo(g,h,i)perylene	26.57	276	139236	10.37	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

Benzaldehyde

Concen: 9.67 ng/ul


RT: 6.88 min Scan# 713

Delta R.T. -0.00 min




Lab File: BG015833.D

Acq: 31 Dec 2014 20:46

Tgt Ion:	77	Resp:	15034
Ion	Ratio	Lower	Upper
77	100		
105	90.4	71.1	106.7
106	91.1	55.8	83.8#



Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 105.00 (104.70 to 105.70): BGC
Ion 106.00 (105.70 to 106.70): BGC



#4

Phenol

Concen: 8.79 ng/ul


RT: 6.94 min Scan# 723

Delta R.T. 0.00 min



Lab File: BG015833.D


Acq: 31 Dec 2014 20:46

Tgt Ion:	94	Resp:	21978
Ion	Ratio	Lower	Upper
94	100		
65	32.5	27.4	41.0
66	71.3	46.4	69.6#



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC





#6

Bis(2-Chloroethyl)ether

Concen: 9.55 ng/ul


RT: 7.16 min Scan# 761

Delta R.T. -0.00 min




Lab File: BG015833.D

Acq: 31 Dec 2014 20:46

Tgt Ion:	93	Resp:	17586
Ion	Ratio	Lower	Upper
93	100		
63	58.4	45.0	67.6
95	34.4	23.3	34.9



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC



#8

2-Chlorophenol

Concen: 10.64 ng/ul


RT: 7.30 min Scan# 785

Delta R.T. 0.00 min



Lab File: BG015833.D


Acq: 31 Dec 2014 20:46

Tgt Ion:	128	Resp:	16488
Ion	Ratio	Lower	Upper
128	100		
64	44.0	45.4	68.0#
130	33.3	29.5	44.3



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 130.00 (129.70 to 130.70): E





#9

2-Methylphenol

Concen: 9.46 ng/ul

RT: 8.17 min Scan# 934

Delta R.T. -0.00 min

Lab File: BG015833.D


Acq: 31 Dec 2014 20:46

Tgt Ion: 108 Resp: 16668

Ion Ratio Lower Upper




108 100

107 90.5 66.6 99.8



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E



#10

2,2'-oxybis(1-Chloropropane)

Concen: 6.31 ng/ul

RT: 8.26 min Scan# 949

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion: 45 Resp: 8366

Ion Ratio Lower Upper

45 100

77 34.5 24.6 37.0



79 21.0 19.5 29.3




Abundance Ion 45.00 (44.70 to 45.70): BGC

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC





#12

Acetophenone

Concen: 9.51 ng/ul

RT: 8.55 min Scan# 998

Delta R.T. 0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion: 105 Resp: 27435

Ion Ratio Lower Upper

105 100

77 94.5 86.6 129.8




51 26.4 22.2 33.4



Abundance Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BG015833.D (-963) (-)

Ion 51.00 (50.70 to 51.70): BG015833.D (-963) (-)



#13

N-Nitroso-di-n-propylamine

Concen: 9.97 ng/ul

RT: 8.53 min Scan# 994

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46

Tgt Ion: 70 Resp: 15784


Ion Ratio Lower Upper

70 100

42 37.9 28.2 42.4

101 9.1 8.7 13.1

130 17.4 18.9 28.3#






Abundance Ion 70.00 (69.70 to 70.70): BG015833.D (-960) (-)

Ion 42.00 (41.70 to 42.70): BG015833.D (-960) (-)

Ion 101.00 (100.70 to 101.70): BG015833.D (-960) (-)

Ion 130.00 (129.70 to 130.70): BG015833.D (-960) (-)





#14

4-Methylphenol

Concen: 8.98 ng/ul

RT: 8.50 min Scan# 990

Delta R.T. 0.00 min

Lab File: BG015833.D


Acq: 31 Dec 2014 20:46

Tgt Ion:108 Resp: 17622

Ion Ratio Lower Upper




108 100

107 123.8 95.3 142.9



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E



#15

Hexachloroethane

Concen: 10.87 ng/ul

RT: 8.80 min Scan# 1040

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:117 Resp: 9095

Ion Ratio Lower Upper

117 100

201 95.3 78.2 117.2



199 68.7 42.2 63.2#




Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E





Ion 199.00 (198.70 to 199.70): E








#18
Nitrobenzene
Concen: 10.10 ng/ul
RT: 8.93 min Scan# 1062
Delta R.T. -0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46


Tgt Ion: 77 Resp: 24687
Ion Ratio Lower Upper
77 100
123 35.9 29.7 44.5
65 15.1 8.2 12.2#



#19
Isophorone
Concen: 9.86 ng/ul
RT: 9.44 min Scan# 1150
Delta R.T. -0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46





Tgt Ion: 82 Resp: 41141
Ion Ratio Lower Upper
82 100
95 9.7 8.7 13.1
138 18.7 12.8 19.2








#21
2-Nitrophenol
Concen: 9.65 ng/ul
RT: 9.64 min Scan# 1183
Delta R.T. 0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46


Tgt Ion:139 Resp: 8661
Ion Ratio Lower Upper
139 100
65 45.0 59.8 89.8#
109 61.2 46.8 70.2



#22
2,4-Dimethylphenol
Concen: 10.88 ng/ul
RT: 9.70 min Scan# 1193
Delta R.T. -0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46


Tgt Ion:107 Resp: 25318
Ion Ratio Lower Upper
107 100
121 36.4 29.4 44.2
122 65.7 50.9 76.3








#23
 Bis(2-Chloroethoxy)methane
 Concen: 9.07 ng/ul
 RT: 9.93 min Scan# 1233
 Delta R.T. 0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46

Tgt Ion	Ratio	Lower	Upper
93	100		
95	31.8	25.1	37.7
123	11.0	10.6	15.8




Abundance Ion 93.00 (92.70 to 93.70): BGC
 Ion 95.00 (94.70 to 95.70): BGC
 Ion 123.00 (122.70 to 123.70): BGC






#25
 2,4-Dichlorophenol
 Concen: 10.70 ng/ul
 RT: 10.17 min Scan# 1273
 Delta R.T. 0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46

Tgt Ion	Ratio	Lower	Upper
162	100		
164	64.7	50.6	76.0
98	29.2	32.0	48.0#




Abundance Ion 162.00 (161.70 to 162.70): BGC
 Ion 164.00 (163.70 to 164.70): BGC
 Ion 98.00 (97.70 to 98.70): BGC








#26
Naphthalene
Concen: 9.86 ng/ul
RT: 10.56 min Scan# 1340
Delta R.T. -0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46

Tgt Ion:128 Resp: 47345
Ion Ratio Lower Upper
128 100
129 8.7 7.1 10.7
127 12.8 9.8 14.6




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#28
4-Chloroaniline
Concen: 10.50 ng/ul
RT: 10.67 min Scan# 1359
Delta R.T. -0.01 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46

Tgt Ion:127 Resp: 18744
Ion Ratio Lower Upper
127 100
129 32.1 26.0 39.0







Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E








#29
Hexachlorobutadiene
Concen: 12.40 ng/ul
RT: 10.85 min Scan# 1389
Delta R.T. -0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46


Tgt Ion: 225 Resp: 19158
Ion Ratio Lower Upper
225 100
223 56.5 49.0 73.4
227 50.2 49.3 73.9



#30
Caprolactam
Concen: 5.60 ng/ul
RT: 11.42 min Scan# 1486
Delta R.T. -0.01 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46





Tgt Ion: 113 Resp: 3725
Ion Ratio Lower Upper
113 100
55 77.0 66.1 99.1
56 91.5 69.0 103.6








#31
 4-Chloro-3-methylphenol
 Concen: 9.27 ng/ul
 RT: 11.81 min Scan# 1553
 Delta R.T. 0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46


Tgt Ion:107 Resp: 19929
 Ion Ratio Lower Upper
 107 100
 144 23.3 14.6 22.0#
 142 60.1 48.2 72.2



#32
 2-Methylnaphthalene
 Concen: 10.42 ng/ul
 RT: 12.18 min Scan# 1615
 Delta R.T. -0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46





Tgt Ion:142 Resp: 36968
 Ion Ratio Lower Upper
 142 100
 141 73.0 72.4 108.6








#34
 1,2,4,5-Tetrachlorobenzene
 Concen: 11.08 ng/ul
 RT: 12.55 min Scan# 1679
 Delta R.T. 0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46


Tgt Ion:	216	Resp:	27480
Ion	Ratio	Lower	Upper
216	100		
214	82.8	59.7	89.5
179	25.3	16.9	25.3#
108	15.7	14.2	21.4



#36
 2,4,6-Trichlorophenol
 Concen: 10.82 ng/ul
 RT: 12.80 min Scan# 1721
 Delta R.T. 0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46





Tgt Ion:	196	Resp:	16362
Ion	Ratio	Lower	Upper
196	100		
198	74.1	81.6	122.4#
200	33.5	26.2	39.4








#37
 2,4,5-Trichlorophenol
 Concen: 9.68 ng/ul
 RT: 12.87 min Scan# 1733
 Delta R.T. -0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46


Tgt Ion:196 Resp: 15921
 Ion Ratio Lower Upper
 196 100
 198 99.5 78.1 117.1
 200 39.6 24.8 37.2#



#38
 1,1'-Biphenyl
 Concen: 10.24 ng/ul
 RT: 13.20 min Scan# 1789
 Delta R.T. -0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46





Tgt Ion:154 Resp: 50512
 Ion Ratio Lower Upper
 154 100
 153 41.9 28.7 43.1
 76 10.0 8.2 12.4








#39
 2-Chloronaphthalene
 Concen: 9.87 ng/ul
 RT: 13.24 min Scan# 1796
 Delta R.T. 0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46


Tgt Ion: 162 Resp: 38019
 Ion Ratio Lower Upper
 162 100
 164 35.7 25.8 38.6
 127 41.5 33.3 49.9



#40
 2-Nitroaniline
 Concen: 8.73 ng/ul
 RT: 13.45 min Scan# 1831
 Delta R.T. -0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46

Tgt Ion: 65 Resp: 12574
 Ion Ratio Lower Upper
 65 100
 92 66.7 54.7 82.1
 138 81.3 61.9 92.9





#42

Dimethylphthalate

Concen: 10.16 ng/ul


RT: 13.83 min Scan# 1896

Delta R.T. -0.00 min




Lab File: BG015833.D

Acq: 31 Dec 2014 20:46

Tgt Ion:	163	Resp:	51719
Ion	Ratio	Lower	Upper
163	100		
194	7.9	5.7	8.5
164	10.3	7.8	11.6



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E



#43

2,6-Dinitrotoluene

Concen: 10.27 ng/ul


RT: 13.94 min Scan# 1916

Delta R.T. -0.00 min



Lab File: BG015833.D


Acq: 31 Dec 2014 20:46

Tgt Ion:	165	Resp:	11424
Ion	Ratio	Lower	Upper
165	100		
89	63.4	79.4	119.2#
121	24.9	23.9	35.9







Abundance Ion 165.00 (164.70 to 165.70): E
Ion 89.00 (88.70 to 89.70): BG
Ion 121.00 (120.70 to 121.70): E








#45
Acenaphthylene
Concen: 10.29 ng/ul
RT: 14.09 min Scan# 1940
Delta R.T. -0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46


Tgt Ion	Resp	Lower	Upper
152	100		
151	17.7	15.4	23.0
153	12.2	10.5	15.7



#46
3-Nitroaniline
Concen: 10.58 ng/ul
RT: 14.28 min Scan# 1973
Delta R.T. 0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46

Tgt Ion	Resp	Lower	Upper
138	100		
108	10.2	10.1	15.1
92	137.7	121.8	182.6





#47

Acenaphthene

Concen: 10.37 ng/ul

RT: 14.43 min Scan# 1999

Delta R.T. 0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:153 Resp: 43298

Ion Ratio Lower Upper

153 100

152 49.0 34.5 51.7




154 87.2 73.0 109.4



Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E



#48

2,4-Dinitrophenol

Concen: 5.33 ng/ul

RT: 14.49 min Scan# 2009

Delta R.T. 0.01 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:184 Resp: 3002

Ion Ratio Lower Upper

184 100

63 80.5 42.0 63.0#



154 91.1 39.6 59.4#




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BG0





Ion 154.00 (153.70 to 154.70): E








#50
4-Nitrophenol
Concen: 9.65 ng/ul
RT: 14.59 min Scan# 2026
Delta R.T. -0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46


Tgt Ion	Ratio	Lower	Upper
109	100		
139	81.9	50.1	75.1
65	97.6	65.4	98.2



#51
Dibenzofuran
Concen: 10.30 ng/ul
RT: 14.77 min Scan# 2056
Delta R.T. -0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46





Tgt Ion	Ratio	Lower	Upper
168	100		
139	42.5	31.8	47.6








#52
 2,4-Dinitrotoluene
 Concen: 9.94 ng/ul
 RT: 14.74 min Scan# 2051
 Delta R.T. -0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46


Tgt Ion:165 Resp: 15842
 Ion Ratio Lower Upper
 165 100
 63 55.3 43.0 64.6
 182 6.5 3.4 5.2#



#53
 2,3,4,6-Tetrachlorophenol
 Concen: 10.59 ng/ul
 RT: 15.00 min Scan# 2095
 Delta R.T. 0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46

Tgt Ion:232 Resp: 18381
 Ion Ratio Lower Upper
 232 100
 131 40.7 32.0 48.0
 130 4.0 1.4 2.2#
 166 21.4 17.7 26.5





#54

Diethylphthalate

Concen: 10.05 ng/ul


RT: 15.19 min Scan# 2128

Delta R.T. -0.00 min




Lab File: BG015833.D

Acq: 31 Dec 2014 20:46

Tgt Ion:	149	Resp:	54718
Ion	Ratio	Lower	Upper
149	100		
177	24.4	18.2	27.2
150	13.8	8.7	13.1#



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E



#56

Fluorene

Concen: 9.96 ng/ul


RT: 15.41 min Scan# 2166

Delta R.T. -0.00 min



Lab File: BG015833.D


Acq: 31 Dec 2014 20:46

Tgt Ion:	166	Resp:	50354
Ion	Ratio	Lower	Upper
166	100		
165	93.2	75.4	113.2
167	15.4	13.1	19.7



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E





#57

4-Chlorophenyl-phenylether

Concen: 11.00 ng/ul

RT: 15.41 min Scan# 2166

Delta R.T. 0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:204 Resp: 35906

Ion Ratio Lower Upper

204 100

206 28.3 29.6 44.4#




141 41.2 46.8 70.2#



Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E



#58

4-Nitroaniline

Concen: 9.76 ng/ul

RT: 15.44 min Scan# 2170

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:138 Resp: 11552

Ion Ratio Lower Upper

138 100

92 53.2 57.5 86.3#



108 169.9 134.6 202.0




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BG0

Ion 108.00 (107.70 to 108.70): E





#61

4,6-Dinitro-2-methylphenol

Concen: 10.44 ng/ul


RT: 15.50 min Scan# 2181

Delta R.T. 0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46




Tgt Ion	Ratio	Lower	Upper
198	100		
182	8.8	4.6	6.8#
77	24.8	26.3	39.5#



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 77.00 (76.70 to 77.70): B



#62

N-Nitrosodiphenylamine

Concen: 10.22 ng/ul


RT: 15.62 min Scan# 2202

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46



Tgt Ion	Ratio	Lower	Upper
169	100		
168	67.5	49.9	74.9
167	32.3	26.7	40.1




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E





#63

4-Bromophenyl-phenylether

Concen: 10.11 ng/ul

RT: 16.31 min Scan# 2318

Delta R.T. 0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:248 Resp: 21573

Ion Ratio Lower Upper

248 100

250 100.5 78.4 117.6




141 67.4 60.5 90.7



Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E



#64

Hexachlorobenzene

Concen: 10.15 ng/ul

RT: 16.42 min Scan# 2338

Delta R.T. 0.01 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:284 Resp: 23492

Ion Ratio Lower Upper

284 100

142 27.5 25.0 37.4



249 33.6 21.6 32.4#




Abundance Ion 284.00 (283.70 to 284.70): E

Ion 142.00 (141.70 to 142.70): E





Ion 249.00 (248.70 to 249.70): E








#65
Atrazine
Concen: 11.23 ng/ul
RT: 16.58 min Scan# 2364
Delta R.T. -0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46


Tgt Ion:200 Resp: 22287
Ion Ratio Lower Upper
200 100
173 23.4 18.2 27.4
215 64.0 41.0 61.4#



#66
Pentachlorophenol
Concen: 9.65 ng/ul
RT: 16.76 min Scan# 2396
Delta R.T. -0.00 min
Lab File: BG015833.D
Acq: 31 Dec 2014 20:46

Tgt Ion:266 Resp: 12247
Ion Ratio Lower Upper
266 100
264 57.3 58.6 87.8#
268 67.0 55.4 83.0





#67

Phenanthrene

Concen: 10.02 ng/ul

RT: 17.15 min Scan# 2462

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:178 Resp: 83328

Ion Ratio Lower Upper

178 100

179 16.9 12.4 18.6




176 19.5 14.0 21.0



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#69

Anthracene

Concen: 10.66 ng/ul

RT: 17.25 min Scan# 2478

Delta R.T. 0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:178 Resp: 90844

Ion Ratio Lower Upper

178 100

179 15.4 12.0 18.0



176 17.5 15.0 22.6




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E





Ion 176.00 (175.70 to 176.70): E








#70
 Carbazole
 Concen: 10.48 ng/ul
 RT: 17.52 min Scan# 2524
 Delta R.T. 0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46


Tgt Ion	Ratio	Lower	Upper
167	100		
166	23.9	14.6	22.0
139	12.9	11.0	16.4



#71
 Di-n-butylphthalate
 Concen: 10.91 ng/ul
 RT: 18.08 min Scan# 2620
 Delta R.T. 0.00 min
 Lab File: BG015833.D
 Acq: 31 Dec 2014 20:46

Tgt Ion	Ratio	Lower	Upper
149	100		
150	9.7	7.8	11.8
104	10.1	7.1	10.7





#72

Fluoranthene

Concen: 11.47 ng/ul

RT: 19.17 min Scan# 2805

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:202 Resp: 131106

Ion Ratio Lower Upper

202 100

101 5.6 4.7 7.1




100 6.2 4.7 7.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#75

Pyrene

Concen: 9.65 ng/ul

RT: 19.53 min Scan# 2867

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:202 Resp: 140543

Ion Ratio Lower Upper

202 100

101 6.9 5.7 8.5



100 6.7 4.6 7.0




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B





#76

Butylbenzylphthalate

Concen: 9.49 ng/ul

RT: 20.43 min Scan# 3020

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:149 Resp: 52225

Ion Ratio Lower Upper

149 100

91 93.5 68.6 102.8




206 27.5 23.7 35.5



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BG0

Ion 206.00 (205.70 to 206.70): E



#77

3,3'-Dichlorobenzidine

Concen: 10.73 ng/ul

RT: 21.21 min Scan# 3153

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:252 Resp: 53648

Ion Ratio Lower Upper

252 100

254 60.3 53.1 79.7



126 7.4 4.3 6.5#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#78

Benzo(a)anthracene

Concen: 10.31 ng/ul

RT: 21.28 min Scan# 3164


Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46

Tgt Ion: 228 Resp: 158669


Ion	Ratio	Lower	Upper
228	100		
229	21.4	15.8	23.8
226	29.1	21.9	32.9





Abundance Ion 228.00 (227.70 to 228.70): E

Ion 229.00 (228.70 to 229.70): E

Ion 226.00 (225.70 to 226.70): E



Time-->



#79

Bis(2-ethylhexyl)phthalate

Concen: 10.26 ng/ul

RT: 21.21 min Scan# 3152


Delta R.T. 0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46

Tgt Ion: 149 Resp: 77716


Ion	Ratio	Lower	Upper
149	100		
167	28.9	25.6	38.4
279	13.0	9.3	13.9




Abundance Ion 149.00 (148.70 to 149.70): E


Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



Time-->





#80

Chrysene

Concen: 10.31 ng/ul

RT: 21.33 min Scan# 3173


Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46

Tgt Ion: 228 Resp: 148682




Ion	Ratio	Lower	Upper
228	100		
226	27.9	24.3	36.5
229	18.5	16.0	24.0



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

Di-n-octyl phthalate

Concen: 10.58 ng/ul


RT: 22.09 min Scan# 3302

Delta R.T. 0.00 min



Lab File: BG015833.D


Acq: 31 Dec 2014 20:46

Tgt Ion: 149 Resp: 130756



Abundance Ion 149.00 (148.70 to 149.70): E





#83

Benzo(b)fluoranthene

Concen: 9.60 ng/ul


RT: 22.88 min Scan# 3436

Delta R.T. 0.00 min




Lab File: BG015833.D

Acq: 31 Dec 2014 20:46

Tgt Ion:	252	Resp:	148405
Ion	Ratio	Lower	Upper
252	100		
253	20.2	17.7	26.5
125	3.9	3.4	5.0



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#84

Benzo(k)fluoranthene

Concen: 10.39 ng/ul


RT: 22.92 min Scan# 3443

Delta R.T. -0.00 min



Lab File: BG015833.D


Acq: 31 Dec 2014 20:46

Tgt Ion:	252	Resp:	150529
Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.4	26.0
125	3.9	2.9	4.3



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#86

Benzo(a)pyrene

Concen: 10.01 ng/ul


RT: 23.46 min Scan# 3535

Delta R.T. -0.00 min




Lab File: BG015833.D

Acq: 31 Dec 2014 20:46

Tgt Ion:	252	Resp:	146068
Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.0	25.4
125	4.5	4.2	6.4



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#87

Indeno(1,2,3-cd)pyrene

Concen: 10.05 ng/ul


RT: 25.86 min Scan# 3944

Delta R.T. -0.00 min



Lab File: BG015833.D


Acq: 31 Dec 2014 20:46

Tgt Ion:	276	Resp:	162303
Ion	Ratio	Lower	Upper
276	100		
138	9.9	7.0	10.6
277	30.0	22.0	33.0



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E





#88

Dibenzo(a,h)anthracene

Concen: 9.83 ng/ul


RT: 25.87 min Scan# 3945

Delta R.T. -0.01 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:	278	Resp:	137362
Ion	Ratio	Lower	Upper
278	100		
139	8.5	5.4	8.2#
279	21.9	19.8	29.8





Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



Time-->



#89

Benzo(g,h,i)perylene

Concen: 10.37 ng/ul


RT: 26.57 min Scan# 4064

Delta R.T. -0.00 min

Lab File: BG015833.D

Acq: 31 Dec 2014 20:46


Tgt Ion:	276	Resp:	139236
Ion	Ratio	Lower	Upper
276	100		
138	8.7	6.9	10.3
277	21.9	19.4	29.2



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



Time-->

