

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG123114\
 Data File : BG015835.D
 Acq On : 31 Dec 2014 21:55
 Operator : TP/IZ
 Sample : SSTD04088
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jan 01 00:58:19 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG123114.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 01 00:49:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	26111	20.00	ng/ul	0.00
16) Naphthalene-d8	10.51	136	96712	20.00	ng/ul	0.00
33) Acenaphthene-d10	14.37	164	72273	20.00	ng/ul	0.00
59) Phenanthrene-d10	17.11	188	173509	20.00	ng/ul	0.00
73) Chrysene-d12	21.30	240	281266	20.00	ng/ul	0.00
81) Perylene-d12	23.56	264	266645	20.00	ng/ul	0.00

System Monitoring Compounds

3) Phenol-d5	6.91	99	87967	36.19	ng/ul	0.00
5) Bis-(2-Chloroethyl)ether-d	7.07	67	47939	35.44	ng/ul	0.00
7) 2-Chlorophenol-d4	7.27	132	66058	40.61	ng/ul	0.00
11) 4-Methylphenol-d8	8.45	113	68034	38.57	ng/ul	0.00
17) Nitrobenzene-d5	8.89	128	32835	42.08	ng/ul	0.00
20) 2-Nitrophenol-d4	9.60	143	37291	44.47	ng/ul	0.00
24) 2,4-Dichlorophenol-d3	10.14	165	68905	43.06	ng/ul	0.00
27) 4-Chloroaniline-d4	10.66	131	70404	39.72	ng/ul	0.00
41) Dimethylphthalate-d6	13.78	166	212661	43.11	ng/ul	0.00
44) Acenaphthylene-d8	14.06	160	259539	42.89	ng/ul	0.00
49) 4-Nitrophenol-d4	14.59	143	36187	39.97	ng/ul	0.00
55) Fluorene-d10	15.36	176	195996	40.07	ng/ul	0.00
60) 4,6-Dinitro-2-methylphenol	15.49	200	45917	46.32	ng/ul	0.00
68) Anthracene-d10	17.21	188	341949	42.64	ng/ul	0.00
74) Pyrene-d10	19.51	212	454875	40.02	ng/ul	0.00
85) Benzo(a)pyrene-d12	23.42	264	502725	41.73	ng/ul	0.00

Target Compounds

						Qvalue
2) Benzaldehyde	6.88	77	49874	31.33	ng/ul#	87
4) Phenol	6.94	94	97396	38.02	ng/ul	94
6) Bis(2-Chloroethyl)ether	7.17	93	71594	37.95	ng/ul	93
8) 2-Chlorophenol	7.30	128	65767	41.45	ng/ul	91
9) 2-Methylphenol	8.18	108	68909	38.19	ng/ul	99
10) 2,2'-oxybis(1-Chloropropan	8.26	45	44982	33.12	ng/ul	96
12) Acetophenone	8.55	105	112891	38.23	ng/ul	87
13) N-Nitroso-di-n-propylamine	8.54	70	63287	39.06	ng/ul	92
14) 4-Methylphenol	8.50	108	75815	37.72	ng/ul	81
15) Hexachloroethane	8.80	117	36527	42.62	ng/ul	97
18) Nitrobenzene	8.93	77	97901	39.26	ng/ul#	93
19) Isophorone	9.45	82	173812	40.83	ng/ul	99
21) 2-Nitrophenol	9.64	139	38989	42.59	ng/ul#	88
22) 2,4-Dimethylphenol	9.70	107	102663	43.23	ng/ul	96
23) Bis(2-Chloroethoxy)methane	9.93	93	87852	37.53	ng/ul#	91
25) 2,4-Dichlorophenol	10.17	162	70239	42.90	ng/ul	96
26) Naphthalene	10.57	128	200609	40.94	ng/ul#	98
28) 4-Chloroaniline	10.68	127	71336	39.15	ng/ul	95
29) Hexachlorobutadiene	10.85	225	75874	48.16	ng/ul	93
30) Caprolactam	11.44	113	19925	29.38	ng/ul	88
31) 4-Chloro-3-methylphenol	11.81	107	90507	41.26	ng/ul	99
32) 2-Methylnaphthalene	12.18	142	152486	42.11	ng/ul	97
34) 1,2,4,5-Tetrachlorobenzene	12.55	216	114754	45.57	ng/ul	92
35) Hexachlorocyclopentadiene	12.53	237	78466	54.50	ng/ul	95


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG123114\
 Data File : BG015835.D
 Acq On : 31 Dec 2014 21:55
 Operator : TP/IZ
 Sample : SSTD04088
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jan 01 00:58:19 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG123114.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 01 00:49:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

36) 2,4,6-Trichlorophenol	12.87	196	71974	46.87	ng/ul	87
37) 2,4,5-Trichlorophenol	12.87	196	72514	43.44	ng/ul	91
38) 1,1'-Biphenyl	13.20	154	207701	41.48	ng/ul	94
39) 2-Chloronaphthalene	13.24	162	160205	40.96	ng/ul	94
40) 2-Nitroaniline	13.45	65	58973	40.34	ng/ul	98
42) Dimethylphthalate	13.83	163	219622	42.51	ng/ul	99
43) 2,6-Dinitrotoluene	13.95	165	48180	42.68	ng/ul#	82
45) Acenaphthylene	14.09	152	271577	41.46	ng/ul	97
46) 3-Nitroaniline	14.28	138	40022	40.03	ng/ul#	95
47) Acenaphthene	14.43	153	174840	41.23	ng/ul	97
48) 2,4-Dinitrophenol	14.49	184	20252	35.41	ng/ul#	90
50) 4-Nitrophenol	14.60	109	58606	43.09	ng/ul	100
51) Dibenzofuran	14.77	168	266286	41.78	ng/ul	91
52) 2,4-Dinitrotoluene	14.74	165	70792	43.77	ng/ul#	93
53) 2,3,4,6-Tetrachlorophenol	15.00	232	82555	46.87	ng/ul#	95
54) Diethylphthalate	15.20	149	241969	43.77	ng/ul	96
56) Fluorene	15.42	166	216886	42.25	ng/ul	98
57) 4-Chlorophenyl-phenylether	15.41	204	143772	43.40	ng/ul	92
58) 4-Nitroaniline	15.44	138	46804	38.94	ng/ul	98
61) 4,6-Dinitro-2-methylphenol	15.50	198	46810	44.84	ng/ul#	93
62) N-Nitrosodiphenylamine	15.63	169	206513	42.41	ng/ul	97
63) 4-Bromophenyl-phenylether	16.31	248	100951	44.80	ng/ul	97
64) Hexachlorobenzene	16.42	284	109316	44.69	ng/ul#	94
65) Atrazine	16.58	200	91303	43.56	ng/ul	90
66) Pentachlorophenol	16.77	266	65522	48.85	ng/ul	92
67) Phenanthrene	17.15	178	367005	41.78	ng/ul	98
69) Anthracene	17.25	178	379065	42.11	ng/ul	98
70) Carbazole	17.52	167	356607	43.12	ng/ul	99
71) Di-n-butylphthalate	18.08	149	432677	44.92	ng/ul	99
72) Fluoranthene	19.17	202	560013	46.40	ng/ul	98
75) Pyrene	19.54	202	590516	39.16	ng/ul	99
76) Butylbenzylphthalate	20.44	149	226503	39.76	ng/ul	90
77) 3,3'-Dichlorobenzidine	21.21	252	227039	43.84	ng/ul#	99
78) Benzo(a)anthracene	21.28	228	653342	41.01	ng/ul	99
79) Bis(2-ethylhexyl)phthalate	21.21	149	330776	42.16	ng/ul	95
80) Chrysene	21.33	228	616370	41.30	ng/ul	97
82) Di-n-octyl phthalate	22.09	149	543281	42.27	ng/ul	100
83) Benzo(b)fluoranthene	22.88	252	661847	41.17	ng/ul	98
84) Benzo(k)fluoranthene	22.92	252	630440	41.84	ng/ul#	98
86) Benzo(a)pyrene	23.46	252	610235	40.24	ng/ul	98
87) Indeno(1,2,3-cd)pyrene	25.87	276	697123	41.53	ng/ul	95
88) Dibenzo(a,h)anthracene	25.88	278	585040	40.28	ng/ul	97
89) Benzo(g,h,i)perylene	26.57	276	583912	41.82	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

Benzaldehyde

Concen: 31.33 ng/ul

RT: 6.88 min Scan# 713





Delta R.T. -0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55

Tgt Ion: 77 Resp: 49874

Ion	Ratio	Lower	Upper
77	100		
105	91.6	71.1	106.7
106	90.9	55.8	83.8#



#4

Phenol

Concen: 38.02 ng/ul

RT: 6.94 min Scan# 723




Delta R.T. 0.00 min


Lab File: BG015835.D

Acq: 31 Dec 2014 21:55

Tgt Ion: 94 Resp: 97396

Ion	Ratio	Lower	Upper
94	100		
65	33.1	27.4	41.0
66	51.4	46.4	69.6





#6

Bis(2-Chloroethyl)ether

Concen: 37.95 ng/ul

RT: 7.17 min Scan# 762

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion: 93 Resp: 71594

Ion Ratio Lower Upper




93 100

63 62.1 45.0 67.6

95 31.5 23.3 34.9



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC



#8

2-Chlorophenol

Concen: 41.45 ng/ul

RT: 7.30 min Scan# 785

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion: 128 Resp: 65767

Ion Ratio Lower Upper



128 100


64 49.9 45.4 68.0

130 32.2 29.5 44.3



Abundance Ion 128.00 (127.70 to 128.70): BGC
Ion 64.00 (63.70 to 64.70): BGC
Ion 130.00 (129.70 to 130.70): BGC





#9

2-Methylphenol

Concen: 38.19 ng/ul

RT: 8.18 min Scan# 934

Delta R.T. -0.00 min

Lab File: BG015835.D


Acq: 31 Dec 2014 21:55

Tgt Ion: 108 Resp: 68909

Ion Ratio Lower Upper

108 100

107 84.5 66.6 99.8



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E

8.18

40000



30000

20000

10000

0

Time--> 8.10 8.15 8.20 8.25



#10

2,2'-oxybis(1-Chloropropane)

Concen: 33.12 ng/ul

RT: 8.26 min Scan# 949

Delta R.T. -0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion: 45 Resp: 44982

Ion Ratio Lower Upper

45 100

77 26.9 24.6 37.0

79 25.0 19.5 29.3



Abundance Ion 45.00 (44.70 to 45.70): BGC

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC

8.26


30000


20000

10000

0

Time--> 8.20 8.25 8.30





#12

Acetophenone

Concen: 38.23 ng/ul


RT: 8.55 min Scan# 998

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55




Tgt Ion:	105	Resp:	112891
Ion	Ratio	Lower	Upper
105	100		
77	91.9	86.6	129.8
51	25.3	22.2	33.4



Abundance Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BG

Ion 51.00 (50.70 to 51.70): BG



#13

N-Nitroso-di-n-propylamine

Concen: 39.06 ng/ul


RT: 8.54 min Scan# 996

Delta R.T. 0.01 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55

Tgt Ion:	70	Resp:	63287
Ion	Ratio	Lower	Upper
70	100		
42	42.2	28.2	42.4
101	13.0	8.7	13.1
130	21.7	18.9	28.3






Abundance Ion 70.00 (69.70 to 70.70): BG

Ion 42.00 (41.70 to 42.70): BG

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E





#14

4-Methylphenol

Concen: 37.72 ng/ul

RT: 8.50 min Scan# 990

Delta R.T. 0.00 min

Lab File: BG015835.D


Acq: 31 Dec 2014 21:55

Tgt Ion:108 Resp: 75815

Ion Ratio Lower Upper

108 100

107 98.3 95.3 142.9



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E

8.50

50000

40000



30000

20000

10000

0

Time-->



#15

Hexachloroethane

Concen: 42.62 ng/ul

RT: 8.80 min Scan# 1040

Delta R.T. -0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion:117 Resp: 36527

Ion Ratio Lower Upper

117 100

201 95.1 78.2 117.2

199 54.4 42.2 63.2



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E

8.80


30000


20000

10000

0


Time-->








#18
 Nitrobenzene
 Concen: 39.26 ng/ul
 RT: 8.93 min Scan# 1062
 Delta R.T. -0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion	Ratio	Lower	Upper
77	100		
123	40.3	29.7	44.5
65	15.4	8.2	12.2




Abundance Ion 77.00 (76.70 to 77.70): BGC
 Ion 123.00 (122.70 to 123.70): BGC
 Ion 65.00 (64.70 to 65.70): BGC






#19
 Isophorone
 Concen: 40.83 ng/ul
 RT: 9.45 min Scan# 1151
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion	Ratio	Lower	Upper
82	100		
95	10.5	8.7	13.1
138	16.2	12.8	19.2







Abundance Ion 82.00 (81.70 to 82.70): BGC
 Ion 95.00 (94.70 to 95.70): BGC
 Ion 138.00 (137.70 to 138.70): BGC








#21
 2-Nitrophenol
 Concen: 42.59 ng/ul
 RT: 9.64 min Scan# 1183
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55


Tgt Ion:139 Resp: 38989
 Ion Ratio Lower Upper
 139 100
 65 56.9 59.8 89.8#
 109 59.1 46.8 70.2



#22
 2,4-Dimethylphenol
 Concen: 43.23 ng/ul
 RT: 9.70 min Scan# 1194
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion:107 Resp: 102663
 Ion Ratio Lower Upper
 107 100
 121 35.0 29.4 44.2
 122 67.7 50.9 76.3





#23

Bis(2-Chloroethoxy)methane

Concen: 37.53 ng/ul





RT: 9.93 min Scan# 1233

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55

Tgt Ion:	93	Resp:	87852
Ion	Ratio	Lower	Upper
93	100		
95	36.4	25.1	37.7
123	16.2	10.6	15.8#



#25

2,4-Dichlorophenol

Concen: 42.90 ng/ul




RT: 10.17 min Scan# 1273


Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55





Tgt Ion:	162	Resp:	70239
Ion	Ratio	Lower	Upper
162	100		
164	66.4	50.6	76.0
98	37.9	32.0	48.0








#26
Naphthalene
Concen: 40.94 ng/ul
RT: 10.57 min Scan# 1341
Delta R.T. 0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55


Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.8	7.1	10.7#
127	12.3	9.8	14.6



#28
4-Chloroaniline
Concen: 39.15 ng/ul
RT: 10.68 min Scan# 1360
Delta R.T. -0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55





Tgt Ion	Ratio	Lower	Upper
127	100		
129	35.1	26.0	39.0








#29
Hexachlorobutadiene
Concen: 48.16 ng/ul
RT: 10.85 min Scan# 1389
Delta R.T. -0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55


Tgt Ion: 225 Resp: 75874
Ion Ratio Lower Upper
225 100
223 52.6 49.0 73.4
227 63.4 49.3 73.9



#30
Caprolactam
Concen: 29.38 ng/ul
RT: 11.44 min Scan# 1489
Delta R.T. 0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55


Tgt Ion: 113 Resp: 19925
Ion Ratio Lower Upper
113 100
55 72.2 66.1 99.1
56 75.5 69.0 103.6








#31
 4-Chloro-3-methylphenol
 Concen: 41.26 ng/ul
 RT: 11.81 min Scan# 1553
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion	Ratio	Lower	Upper
107	100		
144	19.4	14.6	22.0
142	59.9	48.2	72.2




Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E






#32
 2-Methylnaphthalene
 Concen: 42.11 ng/ul
 RT: 12.18 min Scan# 1616
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion	Ratio	Lower	Upper
142	100		
141	87.6	72.4	108.6







Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E








#34
 1,2,4,5-Tetrachlorobenzene
 Concen: 45.57 ng/ul
 RT: 12.55 min Scan# 1679
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55


Tgt Ion:	216	Resp:	114754
Ion	Ratio	Lower	Upper
216	100		
214	83.5	59.7	89.5
179	23.1	16.9	25.3
108	18.9	14.2	21.4



#35
 Hexachlorocyclopentadiene
 Concen: 54.50 ng/ul
 RT: 12.53 min Scan# 1675
 Delta R.T. -0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55


Tgt Ion:	237	Resp:	78466
Ion	Ratio	Lower	Upper
237	100		
235	62.1	53.0	79.4
272	14.6	13.0	19.4








#36
 2,4,6-Trichlorophenol
 Concen: 46.87 ng/ul
 RT: 12.87 min Scan# 1733
 Delta R.T. 0.08 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion:196 Resp: 71974
 Ion Ratio Lower Upper
 196 100
 198 88.6 81.6 122.4
 200 26.7 26.2 39.4




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E






#37
 2,4,5-Trichlorophenol
 Concen: 43.44 ng/ul
 RT: 12.87 min Scan# 1733
 Delta R.T. -0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion:196 Resp: 72514
 Ion Ratio Lower Upper
 196 100
 198 88.6 78.1 117.1
 200 26.7 24.8 37.2







Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E








#38
 1,1'-Biphenyl
 Concen: 41.48 ng/ul
 RT: 13.20 min Scan# 1789
 Delta R.T. -0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55


Tgt Ion:	154	Resp:	207701
Ion	Ratio	Lower	Upper
154	100		
153	39.4	28.7	43.1
76	12.2	8.2	12.4



#39
 2-Chloronaphthalene
 Concen: 40.96 ng/ul
 RT: 13.24 min Scan# 1796
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55





Tgt Ion:	162	Resp:	160205
Ion	Ratio	Lower	Upper
162	100		
164	34.4	25.8	38.6
127	46.8	33.3	49.9








#40
2-Nitroaniline
Concen: 40.34 ng/ul
RT: 13.45 min Scan# 1832
Delta R.T. 0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55


Tgt Ion: 65 Resp: 58973
Ion Ratio Lower Upper
65 100
92 69.7 54.7 82.1
138 79.6 61.9 92.9



#42
Dimethylphthalate
Concen: 42.51 ng/ul
RT: 13.83 min Scan# 1896
Delta R.T. -0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55





Tgt Ion: 163 Resp: 219622
Ion Ratio Lower Upper
163 100
194 6.9 5.7 8.5
164 8.9 7.8 11.6








#43
 2,6-Dinitrotoluene
 Concen: 42.68 ng/ul
 RT: 13.95 min Scan# 1917
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55


Tgt Ion	Ratio	Lower	Upper
165	100		
89	80.4	79.4	119.2
121	22.5	23.9	35.9



#45
 Acenaphthylene
 Concen: 41.46 ng/ul
 RT: 14.09 min Scan# 1941
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55





Tgt Ion	Ratio	Lower	Upper
152	100		
151	17.6	15.4	23.0
153	12.6	10.5	15.7








#46
 3-Nitroaniline
 Concen: 40.03 ng/ul
 RT: 14.28 min Scan# 1973
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55


Tgt Ion:138 Resp: 40022
 Ion Ratio Lower Upper
 138 100
 108 16.2 10.1 15.1#
 92 158.3 121.8 182.6



#47
 Acenaphthene
 Concen: 41.23 ng/ul
 RT: 14.43 min Scan# 1999
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55





Tgt Ion:153 Resp: 174840
 Ion Ratio Lower Upper
 153 100
 152 44.3 34.5 51.7
 154 94.0 73.0 109.4








#48
 2,4-Dinitrophenol
 Concen: 35.41 ng/ul
 RT: 14.49 min Scan# 2008
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55


Tgt Ion:184 Resp: 20252
 Ion Ratio Lower Upper
 184 100
 63 55.4 42.0 63.0
 154 61.0 39.6 59.4#



#50
 4-Nitrophenol
 Concen: 43.09 ng/ul
 RT: 14.60 min Scan# 2027
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion:109 Resp: 58606
 Ion Ratio Lower Upper
 109 100
 139 62.2 50.1 75.1
 65 81.5 65.4 98.2





#51

Dibenzenofuran

Concen: 41.78 ng/ul

RT: 14.77 min Scan# 2056

Delta R.T. -0.00 min

Lab File: BG015835.D


Acq: 31 Dec 2014 21:55

Tgt Ion:168 Resp: 266286

Ion Ratio Lower Upper




168 100

139 45.0 31.8 47.6



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E



#52

2,4-Dinitrotoluene

Concen: 43.77 ng/ul

RT: 14.74 min Scan# 2051

Delta R.T. -0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion:165 Resp: 70792

Ion Ratio Lower Upper

165 100

63 48.5 43.0 64.6



182 6.3 3.4 5.2#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BG0

Ion 182.00 (181.70 to 182.70): E





#53

2,3,4,6-Tetrachlorophenol

Concen: 46.87 ng/ul


RT: 15.00 min Scan# 2095

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55

Tgt Ion:	232	Resp:	82555
Ion	Ratio	Lower	Upper
232	100		
131	35.1	32.0	48.0
130	2.3	1.4	2.2#
166	22.2	17.7	26.5






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#54

Diethylphthalate

Concen: 43.77 ng/ul


RT: 15.20 min Scan# 2129

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55



Tgt Ion:	149	Resp:	241969
Ion	Ratio	Lower	Upper
149	100		
177	25.4	18.2	27.2
150	11.2	8.7	13.1




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E





Ion 150.00 (149.70 to 150.70): E








#56
 Fluorene
 Concen: 42.25 ng/ul
 RT: 15.42 min Scan# 2167
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55


Tgt Ion	Resp	Ratio	Lower	Upper
166	216886	100		
165		95.2	75.4	113.2
167		14.3	13.1	19.7



#57
 4-Chlorophenyl-phenylether
 Concen: 43.40 ng/ul
 RT: 15.41 min Scan# 2166
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55





Tgt Ion	Resp	Ratio	Lower	Upper
204	143772	100		
206		30.9	29.6	44.4
141		53.9	46.8	70.2








#58
4-Nitroaniline
Concen: 38.94 ng/ul
RT: 15.44 min Scan# 2171
Delta R.T. 0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55


Tgt Ion:138 Resp: 46804
Ion Ratio Lower Upper
138 100
92 66.2 57.5 86.3
108 167.9 134.6 202.0



#61
4,6-Dinitro-2-methylphenol
Concen: 44.84 ng/ul
RT: 15.50 min Scan# 2181
Delta R.T. 0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55

Tgt Ion:198 Resp: 46810
Ion Ratio Lower Upper
198 100
182 2.9 4.6 6.8#
77 29.3 26.3 39.5









#62

N-Nitrosodiphenylamine
 Concen: 42.41 ng/ul
 RT: 15.63 min Scan# 2203
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion:	169	Resp:	206513
Ion	Ratio	Lower	Upper
169	100		
168	59.8	49.9	74.9
167	32.6	26.7	40.1




Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E





#63


4-Bromophenyl-phenylether
 Concen: 44.80 ng/ul
 RT: 16.31 min Scan# 2318
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion:	248	Resp:	100951
Ion	Ratio	Lower	Upper
248	100		
250	98.0	78.4	117.6
141	70.7	60.5	90.7



Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E





#64

Hexachlorobenzene

Concen: 44.69 ng/ul

RT: 16.42 min Scan# 2337

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion:284 Resp: 109316

Ion Ratio Lower Upper

284 100

142 32.0 25.0 37.4




249 32.9 21.6 32.4#



Abundance Ion 284.00 (283.70 to 284.70): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#65

Atrazine

Concen: 43.56 ng/ul

RT: 16.58 min Scan# 2365

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion:200 Resp: 91303

Ion Ratio Lower Upper

200 100

173 20.0 18.2 27.4



215 59.1 41.0 61.4




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E


Ion 215.00 (214.70 to 215.70): E








#66
 Pentachlorophenol
 Concen: 48.85 ng/ul
 RT: 16.77 min Scan# 2396
 Delta R.T. -0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion	Ratio	Lower	Upper
266	100		
264	64.0	58.6	87.8
268	65.5	55.4	83.0




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 264.00 (263.70 to 264.70): E
 Ion 268.00 (267.70 to 268.70): E






#67
 Phenanthrene
 Concen: 41.78 ng/ul
 RT: 17.15 min Scan# 2462
 Delta R.T. -0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion	Ratio	Lower	Upper
178	100		
179	16.0	12.4	18.6
176	18.3	14.0	21.0




Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 176.00 (175.70 to 176.70): E








#69
 Anthracene
 Concen: 42.11 ng/ul
 RT: 17.25 min Scan# 2478
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion	Ratio	Lower	Upper
178	100		
179	15.1	12.0	18.0
176	20.2	15.0	22.6




Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 176.00 (175.70 to 176.70): E






#70
 Carbazole
 Concen: 43.12 ng/ul
 RT: 17.52 min Scan# 2524
 Delta R.T. 0.00 min
 Lab File: BG015835.D
 Acq: 31 Dec 2014 21:55

Tgt Ion	Ratio	Lower	Upper
167	100		
166	18.9	14.6	22.0
139	14.2	11.0	16.4



Abundance Ion 167.00 (166.70 to 167.70): E
 Ion 166.00 (165.70 to 166.70): E
 Ion 139.00 (138.70 to 139.70): E





#71

Di-n-butylphthalate

Concen: 44.92 ng/ul


RT: 18.08 min Scan# 2620

Delta R.T. 0.00 min




Lab File: BG015835.D

Acq: 31 Dec 2014 21:55

Tgt Ion:	149	Resp:	432677
Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.8	11.8
104	8.9	7.1	10.7



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 104.00 (103.70 to 104.70): E



#72

Fluoranthene

Concen: 46.40 ng/ul


RT: 19.17 min Scan# 2806

Delta R.T. 0.00 min



Lab File: BG015835.D


Acq: 31 Dec 2014 21:55

Tgt Ion:	202	Resp:	560013
Ion	Ratio	Lower	Upper
202	100		
101	6.2	4.7	7.1
100	4.7	4.7	7.1



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 100.00 (99.70 to 100.70): E





#75

Pyrene

Concen: 39.16 ng/ul





RT: 19.54 min Scan# 2868

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55

Tgt Ion:	202	Resp:	590516
Ion	Ratio	Lower	Upper
202	100		
101	6.9	5.7	8.5
100	5.1	4.6	7.0



#76

Butylbenzylphthalate

Concen: 39.76 ng/ul




RT: 20.44 min Scan# 3021


Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55

Tgt Ion:	149	Resp:	226503
Ion	Ratio	Lower	Upper
149	100		
91	75.9	68.6	102.8
206	33.0	23.7	35.5





#77

3,3'-Dichlorobenzidine

Concen: 43.84 ng/ul


RT: 21.21 min Scan# 3153

Delta R.T. -0.00 min




Lab File: BG015835.D

Acq: 31 Dec 2014 21:55

Tgt Ion:	252	Resp:	227039
Ion	Ratio	Lower	Upper
252	100		
254	65.9	53.1	79.7
126	7.0	4.3	6.5#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 254.00 (253.70 to 254.70): E
Ion 126.00 (125.70 to 126.70): E



#78

Benzo(a)anthracene

Concen: 41.01 ng/ul


RT: 21.28 min Scan# 3164

Delta R.T. -0.00 min



Lab File: BG015835.D


Acq: 31 Dec 2014 21:55

Tgt Ion:	228	Resp:	653342
Ion	Ratio	Lower	Upper
228	100		
229	20.4	15.8	23.8
226	27.1	21.9	32.9



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 229.00 (228.70 to 229.70): E
Ion 226.00 (225.70 to 226.70): E





#79

Bis(2-ethylhexyl)phthalate

Concen: 42.16 ng/ul

RT: 21.21 min Scan# 3152

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion:149 Resp: 330776

Ion Ratio Lower Upper

149 100

167 34.3 25.6 38.4




279 13.9 9.3 13.9



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#80

Chrysene

Concen: 41.30 ng/ul

RT: 21.33 min Scan# 3173

Delta R.T. -0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion:228 Resp: 616370

Ion Ratio Lower Upper

228 100

226 28.1 24.3 36.5



229 19.2 16.0 24.0




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#82

Di-n-octyl phthalate
Concen: 42.27 ng/ul
RT: 22.09 min Scan# 3302
Delta R.T. 0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55

Tgt Ion:149 Resp: 543281




Abundance Ion 149.00 (148.70 to 149.70): E





#83


Benzo(b)fluoranthene
Concen: 41.17 ng/ul
RT: 22.88 min Scan# 3436
Delta R.T. 0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55

Tgt Ion:252 Resp: 661847
Ion Ratio Lower Upper
252 100
253 23.3 17.7 26.5
125 4.7 3.4 5.0



Abundance Ion 252.00 (251.70 to 252.70): E





#84

Benzo(k)fluoranthene

Concen: 41.84 ng/ul

RT: 22.92 min Scan# 3444

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion:252 Resp: 630440

Ion Ratio Lower Upper

252 100

253 22.5 17.4 26.0




125 4.6 2.9 4.3#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#86

Benzo(a)pyrene

Concen: 40.24 ng/ul

RT: 23.46 min Scan# 3536

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55


Tgt Ion:252 Resp: 610235

Ion Ratio Lower Upper

252 100

253 22.2 17.0 25.4



125 4.3 4.2 6.4




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#87

Indeno(1,2,3-cd)pyrene
Concen: 41.53 ng/ul
RT: 25.87 min Scan# 3945
Delta R.T. 0.00 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55




Tgt Ion	Ratio	Lower	Upper
276	100		
138	10.5	7.0	10.6
277	24.6	22.0	33.0




#88

Dibenzo(a,h)anthracene
Concen: 40.28 ng/ul
RT: 25.88 min Scan# 3948
Delta R.T. 0.01 min
Lab File: BG015835.D
Acq: 31 Dec 2014 21:55

Tgt Ion	Ratio	Lower	Upper
278	100		
139	6.6	5.4	8.2
279	22.9	19.8	29.8





#89

Benzo(g,h,i)perylene

Concen: 41.82 ng/ul

RT: 26.57 min Scan# 4065

Delta R.T. 0.00 min

Lab File: BG015835.D

Acq: 31 Dec 2014 21:55

Tgt Ion:	276	Resp:	583912
Ion	Ratio	Lower	Upper
276	100		
138	8.0	6.9	10.3
277	24.1	19.4	29.2

