

Data Path : Z:\HPCHEM1\HPLC B\DATA\HB011013\
Data File : HB001038.D
Signal(s) : DAD1A.CH
Acq On : 10-Jan-2013, 22:10
Operator : OM
Sample : HSTDICV0500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
HPLC_B
ClientSampleId :
HSTDICV0500

Integration File: autoint1.e
Quant Time: Jan 31 05:43:29 2013
Quant Method : Z:\HPCHEM1\HPLC_B\METHOD\HB011013.M
Quant Title :
QLast Update : Fri Jan 11 12:05:05 2013
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 50 ul
Signal Phase : Synerqi HYDRO-RP
Signal Info : 250 X 4.60 mm 4 micron

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 S	4-Nitroaniline	500.000	465.759	6.8	91	-0.08
2 T	HMX	500.000	496.129	0.8	101	-0.03
3 T	RDX	500.000	482.511	3.5	97	-0.10
4 T	1,3,5-Trinitrobenzene	500.000	516.795	-3.4	103	-0.15
5 T	1,3-Dinitrobenzene	500.000	519.147	-3.8	104	-0.23
6 T	Tetryl	500.000	490.936	1.8	92	-0.33
7 T	Nitrobenzene	500.000	528.023	-5.6	105	-0.33
8 T	2,4,6-Trinitrotoluene	500.000	516.356	-3.3	102	-0.37
9 T	4-Amino-2,6-dinitrotoluene	500.000	496.594	0.7	97	-0.44
10 T	2-Amino-4,6-dinitrotoluene	500.000	469.681	6.1	102	-0.43
11 T	2,6-Dinitrotoluene	500.000	495.228	1.0	106	-0.63#
12 T	2,4-Dinitrotoluene	500.000	429.630	14.1	83	-0.53#
13 T	2-Nitrotoluene	500.000	458.988	8.2	75	-0.78#
14 T	4-Nitrotoluene	500.000	493.537	1.3	93	-0.81#
15 T	3-Nitrotoluene	500.000	476.687	4.7	98	-0.90#

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0