

Data Path : Z:\HPCHEM1\HPLC B\DATA\HB011013\
Data File : HB001039.D
Signal(s) : DAD1A.CH
Acq On : 10-Jan-2013, 22:54
Operator : OM
Sample : HSTDCCC0500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
HPLC_B
LabSampleId :
HSTDCCC0500

Integration File: autoint1.e
Quant Time: Jan 25 14:30:22 2013
Quant Method : Z:\HPCHEM1\HPLC_B\METHOD\HB011013.M
Quant Title :
QLast Update : Fri Jan 11 12:05:05 2013
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 50 ul
Signal Phase : Synerqi HYDRO-RP
Signal Info : 250 X 4.60 mm 4 micron

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 S	4-Nitroaniline	500.000	453.817	9.2	89	-0.10
2 T	HMX	500.000	483.587	3.3	99	-0.04
3 T	RDX	500.000	470.633	5.9	95	-0.12
4 T	1,3,5-Trinitrobenzene	500.000	501.272	-0.3	100	-0.18
5 T	1,3-Dinitrobenzene	500.000	506.584	-1.3	101	-0.29
6 T	Tetryl	500.000	531.913	-6.4	100	-0.42
7 T	Nitrobenzene	500.000	497.393	0.5	99	-0.40
8 T	2,4,6-Trinitrotoluene	500.000	488.657	2.3	96	-0.46
9 T	4-Amino-2,6-dinitrotoluene	500.000	513.451	-2.7	100	-0.56#
10 T	2-Amino-4,6-dinitrotoluene	500.000	458.985	8.2	100	-0.55#
11 T	2,6-Dinitrotoluene	500.000	477.617	4.5	102	-0.79#
12 T	2,4-Dinitrotoluene	500.000	480.550	3.9	93	-0.67#
13 T	2-Nitrotoluene	500.000	479.874	4.0	78	-0.97#
14 T	4-Nitrotoluene	500.000	512.148	-2.4	97	-1.03#
15 T	3-Nitrotoluene	500.000	494.297	1.1	102	-1.14#

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0