

Data File : Z:\HPCHEM1\HPLC_B\DATA\HB082212\HB000785.D Vial: 12
Acq On : 22-Aug-2012, 23:07 Operator: HNP
Sample : PB64357BL Inst : G1315A
Misc : Multiplr: 1.00
IntFile : autoint1.e
Quant Time: Aug 23 12:01 2012 Quant Results File: HB082212.RES

Quant Method : Z:\HPCHEM1\HPLC_B\METHOD\HB082212.M
Title :
Last Update : Thu Aug 23 11:41:24 2012
Response via : Initial Calibration
DataAcq Meth : NEWTEST.M

Volume Inj. : 10 ul
Signal Phase : Synergi HYDRO-RP
Signal Info : 250 X 4.60 mm 4 micron

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
1) S 1,2-Dinitrobenzene	10.27	240505	826.875 ug/ml
Spiked Amount 500.000		Recovery	= 165.38%


Target Compounds


Quantitation Report

Data File : Z:\HPCHEM1\HPLC_B\DATA\HB082212\HB000785.D Vial: 12
Acq On : 22-Aug-2012, 23:07 Operator: HNP
Sample : PB64357BL Inst : G1315A
Misc : Multiplr: 1.00
IntFile : autoint1.e
Quant Time: Aug 23 12:01 2012 Quant Results File: HB082212.RES

Quant Method : Z:\HPCHEM1\HPLC_B\METHOD\HB082212.M
Title :
Last Update : Thu Aug 23 11:41:24 2012
Response via : Multiple Level Calibration
DataAcq Meth : NEWTEST.M

Volume Inj. : 10 ul
Signal Phase : Synergi HYDRO-RP
Signal Info : 250 X 4.60 mm 4 micron





#1 1,2-Dinitrobenzene

R.T.: 10.275 min
Delta R.T.: 0.039 min
Response: 240505
Conc: 826.87 ug/ml