

Method Path : Z:\HPCHEM1\HPLC_B\METHOD\
 Method File : HB112712.M
 Title :
 Last Update : Wed Nov 28 11:17:49 2012
 Response Via : Initial Calibration

Calibration Files

5000 =HB000937.D 2000 =HB000938.D 1000 =HB000939.D
 500 =HB000940.D 100 =HB000941.D 20 =HB000943.D

	Compound	5000	2000	1000	500	100	20	Avg		%RSD
1) S	4-Nitroaniline	7.491	7.112	7.333	6.983	8.753		7.534	E2	9.41
2) T	HMX	8.562	7.732	7.846	8.513	9.723		8.475	E2	9.35
3) T	RDX	7.163	6.495	6.647	6.851	5.224	8.051	6.738	E2	13.73
4) T	1,3,5-Trinitr...	1.607	1.434	1.449	1.463	1.068	0.977	1.333	E3	18.75
5) T	1,3-Dinitrobe...	1.570	1.397	1.357	1.408	0.968	1.149	1.308	E3	16.39
6) T	Nitrobenzene	1.174	1.024	1.012	1.079	0.754	0.797	0.973	E3	16.85
7) T	Tetryl	1.100	0.973	0.950	1.087	0.682	0.739	0.922	E3	18.99
8) T	2,4,6-Trinitr...	1.357	1.179	1.167	1.255	1.102	0.905	1.161	E3	13.17
9) T	2-Amino-4,6-d...	2.048	1.508	1.471	1.533	1.464	2.099	1.687	E3	17.82
10) T	4-Amino-2,6-d...	0.900	1.157	1.145	1.277	1.394	0.831	1.117	E3	19.33
11) T	2,4-Dinitroto...	1.557	1.303	1.308	1.366	1.204	1.213	1.325	E3	9.75
12) T	2,6-Dinitroto...	8.979	8.984	8.644	9.803	8.969	4.209	8.264	E2	24.49
13) T	2-Nitrotoluene	8.832	8.365	7.875	9.321	8.029		8.485	E2	7.01
14) T	4-Nitrotoluene	1.196	1.127	1.107	1.322	0.751		1.101	E3	19.34
15) T	3-Nitrotoluene	1.125	1.078	1.010	1.209	0.732		1.031	E3	17.64

(#) = Out of Range