

Method Path : Z:\HPCHEM1\HPLC_B\METHOD\
 Method File : HB010913.M
 Title :
 Last Update : Wed Jan 09 14:12:24 2013
 Response Via : Initial Calibration

Calibration Files

5000 =HB001009.D 2000 =HB001015.D 1000 =HB001011.D
 500 =HB001027.D 100 =HB001013.D 20 =HB001014.D

	Compound	5000	2000	1000	500	100	20	Avg		%RSD
1) S	4-Nitroaniline	6.296	6.135	6.218	6.510	7.790	8.491	6.907	E2	14.32
2) T	HMX	0.731	0.759	0.766	0.785	1.018		0.812	E3	14.40
3) T	RDX	6.143	6.144	6.342	6.288	6.502		6.284	E2	2.39
4) T	1,3,5-Trinitr...	1.364	1.355	1.420	1.359	1.349	1.844	1.449	E3	13.50
5) T	1,3-Dinitrobe...	1.344	1.343	1.374	1.358	1.425	1.430	1.379	E3	2.84
6) T	Nitrobenzene	1.031	0.995	1.010	1.029	1.091	1.262	1.070	E3	9.31
7) T	Tetryl	0.959	0.921	0.939	1.025	1.095	1.197	1.023	E3	10.44
8) T	2,4,6-Trinitr...	1.023	1.165	1.176	1.190	1.180	1.249	1.164	E3	6.43
9) T	2-Amino-4,6-d...	1.658	1.360	1.408	1.487	1.194	1.568	1.446	E3	11.32
10) T	4-Amino-2,6-d...	0.986	1.039	1.130	1.266	0.934	1.211	1.094	E3	11.90
11) T	2,4-Dinitroto...	1.291	1.241	1.290	1.478	1.276	1.630	1.368	E3	11.21
12) T	2,6-Dinitroto...	0.850	0.852	0.867	1.081	0.901	1.011	0.927	E3	10.42
13) T	2-Nitrotoluene	7.908	7.910	8.244	8.801	6.584		7.889	E2	10.34
14) T	4-Nitrotoluene	1.080	1.018	1.089	1.173	1.224	1.603	1.198	E3	17.66
15) T	3-Nitrotoluene	1.039	1.006	1.021	1.065	0.865	1.223	1.036	E3	11.12

(#) = Out of Range