

Method Path : Z:\HPCHEM1\HPLC_L\METHODS\
 Method File : HL020613.M
 Title :
 Last Update : Thu Feb 07 11:25:51 2013
 Response Via : Initial Calibration

Calibration Files

5000 =HL005002.D 2000 =HL005003.D 1000 =HL005004.D
 500 =HL005005.D 100 =HL005006.D 20 =HL005007.D

	Compound	5000	2000	1000	500	100	20	Avg		%RSD
1) S	4-Nitroaniline	4.382	4.181	3.826	3.670	3.712	3.209	3.830	E3	10.78
2) T	HMX	4.313	4.447	4.150	4.175	4.189	3.744	4.170	E3	5.67
3) T	RDX	5.192	5.103	4.956	4.924	4.705	3.801	4.780	E3	10.62
4) T	1,3,5-Trinitr...	1.104	1.035	1.036	1.050	1.039	1.026	1.048	E4	2.70
5) T	1,3-Dinitrobe...	1.584	1.583	1.569	1.577	1.464	1.423	1.533	E4	4.63
6) T	Nitrobenzene	9.900	9.905	9.884	9.869	9.713	7.888	9.526	E3	8.46
7) T	Tetryl	6.342	6.431	6.602	6.533	6.651	6.614	6.529	E3	1.84
8) T	2,4,6-Trinitr...	1.179	1.100	1.096	1.119	1.106	1.060	1.110	E4	3.53
9) T	2-Amino-4,6-d...	1.226	1.081	1.056	1.068	1.080	1.078	1.098	E4	5.78
10) T	4-Amino-2,6-d...	5.018	6.404	6.696	6.784	6.905	6.301	6.351	E3	10.90
11) T	2,4-Dinitroto...	1.550	1.478	1.466	1.479	1.441	1.423	1.473	E4	2.97
12) T	2,6-Dinitroto...	7.290	7.848	7.894	7.839	7.408	6.650	7.488	E3	6.45
13) T	2-Nitrotoluene	6.638	6.568	6.584	6.554	6.590	8.180	6.852	E3	9.50
14) T	4-Nitrotoluene	5.665	5.457	5.540	5.596	5.825	5.247	5.555	E3	3.52
15) T	3-Nitrotoluene	7.492	7.211	6.948	6.939	6.297	5.356	6.707	E3	11.50

(#) = Out of Range