

Data Path : Z:\HPCHEM1\HPLC_L\Data\HL020613\
Data File : HL005029.D
Signal(s) : VWD1A.CH
Acq On : 06-Feb-2013, 23:57
Operator : QM
Sample : HSTDCCC0500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 08 07:48:10 2013
Quant Method : Z:\HPCHEM1\HPLC_L\METHODS\HL020613.M
Quant Title :
QLast Update : Thu Feb 07 11:25:51 2013
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 50 ul
Signal Phase : ZORBAX Extend-C18
Signal Info : 4.6 X 100mm 3.5-Micron 80A

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 S	4-Nitroaniline	500.000	431.878	13.6	90	0.00
2 T	HMX	500.000	473.447	5.3	95	0.00
3 T	RDX	500.000	519.074	-3.8	101	0.00
4 T	1,3,5-Trinitrobenzene	500.000	493.585	1.3	99	0.00
5 T	1,3-Dinitrobenzene	500.000	514.160	-2.8	100	0.00
6 T	Nitrobenzene	500.000	521.638	-4.3	101	0.00
7 T	Tetryl	500.000	443.916	11.2	89	0.04
8 T	2,4,6-Trinitrotoluene	500.000	502.506	-0.5	100	0.02
9 T	2-Amino-4,6-dinitrotoluene	500.000	485.951	2.8	100	0.04
10 T	4-Amino-2,6-dinitrotoluene	500.000	550.402	-10.1	103	0.04
11 T	2,4-Dinitrotoluene	500.000	501.804	-0.4	100	0.03
12 T	2,6-Dinitrotoluene	500.000	522.360	-4.5	100	0.03
13 T	2-Nitrotoluene	500.000	478.631	4.3	100	0.04
14 T	4-Nitrotoluene	500.000	484.363	3.1	96	0.04
15 T	3-Nitrotoluene	500.000	527.652	-5.5	102	0.04

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0