

Data File : H:\VOL1\HPLC\HPLC L\120405\HL002657.D Vial: 29
Acq On : 06-Apr-2012, 04:09 Operator: HLM
Sample : D1992-12 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e
Quant Time: Apr 6 10:08 2012 Quant Results File: HL032912.RES

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL032912.M (Chemstation Integrator)
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Fri Apr 06 10:10:10 2012
Response via : Initial Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

1) S 4-Nitroaniline	1.45	588563	154.533 ug/L
Spiked Amount 300.000	Range 37 - 149	Recovery =	51.51%

Target Compounds

Data File : H:\VOL1\HPLC\HPLC L\120405\HL002657.D Vial: 29
Acq On : 06-Apr-2012, 04:09 Operator: HLM
Sample : D1992-12 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e
Quant Time: Apr 6 10:08 2012 Quant Results File: HL032912.RES

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL032912.M (Chemstation Integrator)
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Fri Apr 06 10:10:10 2012
Response via : Multiple Level Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

