

Method Path : Z:\HPCHEM1\HPLC_L\methods\
 Method File : HL06071A.M
 Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
 Last Update : Fri Jun 08 13:23:23 2012
 Response Via : Initial Calibration

Calibration Files

CAL1 =HL002911.D CAL2 =HL003178.D CAL3 =HL003177.D
 CAL4 =HL003176.D CAL5 =HL003175.D CAL6 =HL003174.D

	Compound	CAL1	CAL2	CAL3	CAL4	CAL5	CAL6	Avg	%RSD
1) S	4-Nitroaniline							0.000	-1.00
2) T	HMX							0.000	-1.00
3) T	RDX							0.000	-1.00
4) T	1,3,5-Trinitr...							0.000	-1.00
5) T	Tetryl							0.000	-1.00
6) T	2,4,6-Trinitr...							0.000	-1.00
7) T	2-Amino-4,6-d...							0.000	-1.00
8) T	4-Amino-2,6-d...							0.000	-1.00
9) T	2,4-Dinitroto...							0.000	-1.00
10) T	2,6-Dinitroto...							0.000	-1.00
11) T	1,3-Dinitrobe...							0.000	-1.00
12) T	Nitrobenzene							0.000	-1.00
13) T	2-Nitrotoluene							0.000	-1.00
14) T	3-Nitrotoluene							0.000	-1.00
15) T	4-Nitrotoluene							0.000	-1.00
16) T	3,5-Dinitroan...							0.000	-1.00
17) T	Nitroglycerin	6.118	6.542	5.986	5.599	5.502		5.949 E1	7.05
18) T	PETN							0.000	-1.00

(#) = Out of Range