

Data File : Z:\HPCHEM1\HPLC_L\DATA\HL022713\HL005089.D Vial: 57
Acq On : 27-Feb-2013, 19:57 Operator: QM
Sample : HSTDICV0500 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : autoint1.e
Quant Time: Feb 28 15:18 2013 Quant Results File: HL022713.RES

Quant Method : Z:\HPCHEM1\HPLC_L\METHODS\HL022713.M
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Thu Feb 28 14:20:05 2013
Response via : Initial Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

Compound	R.T.	Response	Conc Units


System Monitoring Compounds			
1) S 4-Nitroaniline	1.48	2024637	464.716 ug/L
Spiked Amount 300.000	Range 37 - 149	Recovery	= 154.91%#
Target Compounds			
2) L1 3,5-Dinitroaniline	4.41	6168498	438.216 ug/L


Quantitation Report

Data File : Z:\HPCHEM1\HPLC_L\DATA\HL022713\HL005089.D Vial: 57
Acq On : 27-Feb-2013, 19:57 Operator: QM
Sample : HSTDICV0500 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : autoint1.e
Quant Time: Feb 28 15:18 2013 Quant Results File: HL022713.RES

Quant Method : Z:\HPCHEM1\HPLC_L\METHODS\HL022713.M
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Thu Feb 28 14:20:05 2013
Response via : Multiple Level Calibration
DataAcq Meth : 8330AGIP.M


Volume Inj. :
Signal Phase :
Signal Info :





#1 4-Nitroaniline

R.T.: 1.480 min
 Delta R.T.: 0.000 min
 Response: 2024637
 Conc: 464.72 ug/L



#2 3,5-Dinitroaniline

R.T.: 4.412 min
 Delta R.T.: 0.002 min
 Response: 6168498
 Conc: 438.22 ug/L