

Data File : Z:\HPCHEM1\HPLC_L\DATA\HL022713\HL005090.D Vial: 58
Acq On : 27-Feb-2013, 20:24 Operator: QM
Sample : PBW Inst : G1314A
Misc : Multiplr: 1.00
IntFile : autoint1.e
Quant Time: Feb 28 15:19 2013 Quant Results File: HL022713.RES

Quant Method : Z:\HPCHEM1\HPLC_L\METHODS\HL022713.M
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Thu Feb 28 14:20:05 2013
Response via : Initial Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
1) S 4-Nitroaniline	1.48	1178202	270.433 ug/L
Spiked Amount 300.000	Range 37 - 149	Recovery =	90.14%


Target Compounds


Quantitation Report

Data File : Z:\HPCHEM1\HPLC_L\DATA\HL022713\HL005090.D Vial: 58
Acq On : 27-Feb-2013, 20:24 Operator: QM
Sample : PBW Inst : G1314A
Misc : Multiplr: 1.00
IntFile : autoint1.e
Quant Time: Feb 28 15:19 2013 Quant Results File: HL022713.RES

Quant Method : Z:\HPCHEM1\HPLC_L\METHODS\HL022713.M
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Thu Feb 28 14:20:05 2013
Response via : Multiple Level Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :





#1 4-Nitroaniline

R.T.: 1.477 min
Delta R.T.: -0.003 min
Response: 1178202
Conc: 270.43 ug/L