

Method Path : S:\HPCHEM1\HPLC_L\methods\
Method File : HL022713.M
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Thu Feb 28 14:20:05 2013
Response Via : Initial Calibration

Calibration Files

5000	=HL005083.D	2000	=HL005084.D	1000	=HL005085.D
500	=HL005086.D	100	=HL005087.D	20	=HL005088.D

	Compound	5000	2000	1000	500	100	20	Avg		%RSD
1) S	4-Nitroaniline	4.405	4.413	4.350	4.347	5.078	3.547	4.357	E3	11.16
2) L1	3,5-Dinitroan...	1.436	1.454	1.433	1.425	1.390	1.309	1.408	E4	3.74
3) L2	PETN	9.456	9.289	9.297	9.946	8.885	7.878	9.125	E3	7.68

(#) = Out of Range